A-A-30164
May 28, 1992

# COMMERCIAL ITEM DESCRIPTION

## COVER, STERILITY MAINTENANCE

The General Services Administration has authorized the use of this commercial item description.

This commercial item description describes clear plastic sterility maintenance covers suitable for use as a dust cover for items that have been sterilized. Open ends are closed by use of either a heat seal or tape, or self-sealing adhesive flap closure.

Salient characteristics. The cover shall be constructed of clear plastic film and designed specifically to contain items after sterilization. The covers are to provide a barrier against moisture, dust, lint and airborne contaminants after closure. The cover shall be sealed, or closed, on three ends. One closed end shall have a tear line or other feature suitable for rapid opening without use of a scissors, or other instrument. A perforated tear line shall not be acceptable because this would compromise the barrier quality of the cover. The open end of the cover shall be readily sealable by use of one of the following methods.

Type A - Heat seal or tape
Type B - Self-sealing adhesive flap

The plastic film used in construction of the cover shall be 3 mils + or - 1/2 mil thick. Tolerances for the width measurement shall be + or - 1/4 inch (0.635 cm) and for the length measurement shall be + or - 1/2 inch (1.27 cm). The covers shall be supplied in the following sizes

Size 1. 6 inches wide x 10 inches long (18.24 x 25.40 cm).

Size 2. 8 inches wide x 12 inches long (20.32 x 30.48 cm).

Size 3. 10 inches wide x 15 inches long (25.40 x 38.10 cm).

Size 4. 12 inches wide x 20 inches long (30.48 x 50.80 cm).

Size 5. 16 inches wide x 20 inches long (40.64 x 50.80 cm).

Size 6. 16 inches wide x 22 inches long (40.64 x 55.88 cm).



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Size 7. 16 inches wide x 30 inches long (40.64 x 76.20 cm).

Size 8. 24 inches wide x 30 inches long (60.96 x 76.20 cm).

The cover shall be plainly marked with the manufacturers name and product number. Other acceptable markings include instructions for closing or opening the cover and statements about the use of the cover.

The cover shall be free of holes, tears, perforations, dirt, grease or foreign matter.

Workmanship. The sterility maintenance cover shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Case (CS) One case containing from 125 to 1000 covers, usually provided by the manufacturer as commercial practice, constitutes one unit. Unit shall be stated in the solicitation.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of zero (percent defective) and inspection level of S-1. Sampling for tests (dimensional), shall be in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-3. The unit or product for test purposes shall be one sterility maintenance cover.

Tests. Tests shall be conducted to determine compliance with this commercial item description requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to the first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereafter. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be the degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from the receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be that usually provided by the supplier as commercial practice, and as specified in the contract and/or order.

NOTE The following National Stock Numbers are covered by this document:

| | |
|---|---|
| 6530-01-328-4847; 6530-01-311-5216 | (Type A: Size 1) |
| 6530-01-328-8039, 6530-01-328-8040 | (Type B. Size 1) |
| 6530-01-328-4848; 6530-01-311-5217 | (Type A. Size 2) |
| 6530-01-328-4964; 6530-01-278-9411 | (Type B. Size 2) |
| 6530-01-328-4849, 6530-01-253-3806 | (Type A: Size 3) |
| 6530-01-294-2863, 6530-01-239-9795 | (Type B. Size 3) |
| 6530-01-328-4850; 6530-01-265-3379 | (Type A: Size 4) |
| 6530-01-294-2861, 6530-01-240-2113 | (Type B Size 4) |
| 6530-01-328-4851; 6530-01-328-4001 | (Type A Size 5) |
| 6530-01-328-4967 | (Type B. Size 5) |
| 6530-01-252-5566 | (Type A. Size 6) |
| 6530-01-294-2862, 6530-01-278-9412 | (Type B. Size 6) |
| 6530-01-195-4542 | (Type A Size 7) |
| 6530-01-328-4966, 6530-01-278-9413 | (Type B Size 7) |
| 6530-01-276-7997, 6530-01-328-4000 | (Type A. Size 8) |
| 6530-01-296-0855; 6530-01-328-8041 | (Type B. Size 8) |
| 6530-01-328-4965 | (Type B Size 8) |

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government and MIL-STD-105, Sampling Procedures and Tables for Inspection by Attributes, may be obtained from the Superintendent of Documents, U.S. Printing Office, Washington, DC 20420.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number and date of this commercial item description.

(b) Type, size and title



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

A-A-30164

MILITARY INTERESTS CIVIL AGENCY COORDINATING ACTIVITIES.

Military Coordinating Activity.

PHS
FDA-MPQAS

DoD-MB

PREPARING ACTIVITY.

VA-OSS VANAC (904)
Agent VA-OSS VACO (92A)

AGENT

DLA-DM

Custodians

Army-MD
Navy-MS
Air Force-03

Project No. 6530-2000

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.