A-A-30170
February 28, 1992

COMMERCIAL ITEM DESCRIPTION

CUSHION, WHEELCHAIR SEAT
(Rubber Air Cells)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description describes air inflatable wheelchair cushions constructed of rubber cells and designed for the long term use of the wheelchair patient.

Salient characteristics. Wheelchair cushions shall be constructed of soft, flexible and interconnected rubber cells which are air inflatable. This "dry floatation" shall assure evenly distributed body weight, with no peak pressures impacting on any part of the body. The cushion shall conform to the body shape. These cushions shall be compatible for use with a solid contour base commonly used to provide solid foundation to enhance positioning and stability in sling seat wheelchairs. These cushions shall be steam and ethylene oxide sterilizable.

An inflation pump, repair kit and operating instructions shall be provided with each cushion. In addition, a form-fitting cushion cover shall be provided with each cushion. The cover shall help protect the cushion surface, keep the cushion cells from protruding beyond the wheelchair seat and aid in slide transfer. The bottom of the cover design shall be made of a special non-skid material so the cushion does not slide or slip in the chair.

Single valve, single manifold style cushions shall be used to provide even pressure over the entire surface of the cushion. Dual valve, dual manifold style wheelchair cushions shall be used to achieve side-to-side or front-to-back distribution of pressure.

All measurements for cushions shown are approximate from edge to edge. Wheelchair cushions covered by this document shall be of the following sizes and styles:

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Size 1 - 15 X 15 X 2 inches (38 X 38 X 5 cm).

Size 2 - 15 X 15 X 4 inches (38 X 38 X 10 cm).

Size 3 - 15 X 16-3/4 X 2 inches (38 X 43 X 5 cm).

Size 4 - 15 X 16-3/4 X 4 inches (38 X 43 X 10 cm).

Size 5 - 16-3/4 X 16-3/4 X 2 inches (43 X 43 X 5 cm).

Size 6 - 16-3/4 X 16-3/4 X 4 inches (43 X 43 X 10 cm).

Size 7 - 16-3/4 X 18-1/2 X 2 inches (43 X 47 X 5 cm).

Style A - Low profile cushion with single valve, single manifold; 2 inch (5 cm) high rubber cells.

Style B - Low profile cushion with dual valve, dual manifold; 2 inch (5 cm) high rubber cells.

Style C - High profile cushion with single valve, single manifold; 4 inch (10 cm) high rubber cells.

Style D - High profile cushion with dual valve, dual manifold; 4 inch (10 cm) high rubber cells.

Workmanship. The wheelchair cushions shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). A single wheelchair cushion with cover, instructions, inflation pump and repair kit as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the government, The government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with this commercial item description requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for examination purposes shall be one wheelchair cushion as specified in this description.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Numbers are covered by this document:

Size 2 - Style C - NSN 6530-01-193-3602

Size 3 - Style A - NSN 6530-01-193-3603

Size 4 - Style C - NSN 6530-01-193-3601

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Size 4 - Style D - NSN 6530-01-346-4491

Size 6 - Style C - NSN 6530-01-346-4489

Size 7 - Style A - NSN 6530-01-346-4490

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government and MIL-STD-105, Sampling Procedures and Tables for Inspection by Attributes may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description
(b) Size and style.

MILITARY INTERESTS: CIVIL AGENCY COORDINATING ACTIVITIES:

Military Coordinating Activity: FDA-MPQAS
PHS

DoD-MB

PREPARING ACTIVITY:

VA-OSS VAMKC (904)
Agent VA-OSS VACO (92A)

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

Project No. 6530-2343

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.