INCH-POUND

A-A-51992B
19 March 1993
SUPERSEDING
A-A-51992A
21 September 1990

# COMMERCIAL ITEM DESCRIPTION

## HOLDER, SUTURE NEEDLE, Castroviejo

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers suture needle holders suitable for use during ophthalmic and plastic surgery.

Commercial Item Description Sheet. Suture needle holders shall be in accordance with all requirements specified herein and the applicable Commercial Item Description Sheet (slash sheet - See Notes).

Salient characteristics:

General. Suture needle holders shall be of the following styles and classes.

Style I - Delicate jaws

Class 1 - Straight, 5-1/2 inches
Class 2 - Curved, 5-1/2 inches

Style II - Heavy jaws

Class 1 - Straight, 6 inches

Material. The material shall be Class 4 corrosion resistant steel in accordance with ASTM F899.

Style, design and dimensions. Types and Classes shall be in accordance with applicable Commercial Item Description Specification Sheet and as specified herein (See Notes).

Hardness. Suture needle holders shall have a Rockwell Hardness C scale (HRC) in accordance with the requirements of ASTM F1325.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution in unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Finish. The finish of the instrument shall be mirror, satin, or mirror and satin. Should the holder have a mirror finish, the interior surfaces need not have a mirror finish.

Passivation. Suture needle holders shall be properly passivated.

Corrosion resistance. Suture needle holders shall show no signs of corrosion or plating of copper when tested for corrosion resistance in accordance with ASTM F1089.

Identification marking. Suture needle holders shall be legibly marked with the contractor's name or registered trademark and the word "stainless". Suture needle holders of foreign manufacture shall bear the name of the country of origin.

Performance. Suture needle holders shall meet the performance requirements of ASTM F1325.

Workmanship. Suture needle holders shall be in accordance with ASTM F1325 for workmanship, finish and appearance.

Unit. Each (EA). One suture needle holder of the style and class specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Suture needle holders shall be examined to determine compliance with all specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and as specified. Unit of product for examination purpose shall be one suture needle holder. Sampling shall be general inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. The unit of product for test purposes shall be one suture needle holder. Sampling shall be special inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion, tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities, labels and markings shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

Notes:

1. Contracting officer shall select the options permitted herein, and include the following information in procurement documents:

(a) Title, number and date of this Commerical Item Description.

(b) Title, number and date of the applicable Commerical Item Description slash sheet as follows:

| | |
|---|---|
| A-A-51992/1 | HOLDER, SUTURE NEEDLE, Castroviejo, 5-1/2 inches |
| A-A-51992/2 | HOLDER, SUTURE NEEDLE, Castroviejo, 6 inches |



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

2. Copies of Commerical Item Description A-A-51992B and Commerical Item Description Slash Sheets may be obtained by writing the Defense Printing Service Detachment Office; Bldg. 4D; DoD-SSP; 700 Robbins Avenue; Philadelphia, PA; 19222-5094

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

Project No. 6515-5353
Location: genhold\S17



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.