MIL-A-36163 (BUMED)
4 October 1963

MILITARY SPECIFICATION

AIRWAY, PHARYNGEAL, RUBBER

1. SCOPE

1.1 Scope. This specification covers rubber pharyngeal airway, suitable for use during anesthesia procedures.

1.2 Classification. Airways shall be of the following sizes as specified (see 6.1):

Size 0 - Infant.
Size 2 - Child.
Size 3 - Small adult.
Size 5 - Large adult.

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-665	Boxes, Paperboard, Metal Stayed.
PPP-B-676	Boxes, Set-up, Paperboard.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:09Z
Check the source to verify that this is the current version before use.

Military

MIL-L-10547 Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for shipment and Storage.

Federal

Fed.Test Method Std. No. 601 Rubber: Sampling and Testing.

DRAWING

Defense Medical Supply Center

20065 Airway Pharyngeal Rubber.

(Copies of specifications, standards and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. Material shall be rubber of the highest quality throughout, and suitable in all respects for the purpose intended.

3.2 Dimensions, style and design. The airway shall be a pharyngeal tube, nontraumatic and flexible, The airway shall be in accordance with unit A or unit B of Drawing 20065.

3.3 Toxicity. The airways shall be free from toxic material or any ingredient that may have a deleterious effect upon body tissues.

3.4 Accelerated aging test. Airways, after being subjected to accelerated aging test specified in 4.4.1, shall show no stiffening, softening, cracking, or other adverse change.

3.5 Identification marking. Airway shall be permanently and legibly marked with the name of manufacturer and date of manufacture (month and year).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:09Z
Check the source to verify that this is the current version before use.

3.6 Workmanship. Workmanship shall be first class throughout. Airways shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with Standard MIL-STD-105.

4.2.2 For accelerated aging test. Sufficient samples shall be selected from each lot to perform the test. Failure to comply with test requirements shall be cause for rejection of the lot represented.

4.2.3 Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The airway shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Accelerated aging test. The tube shall be subjected to the air heat test for seven days in accordance with method 7821 of Federal Test Method Standard No. 601.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:09Z
Check the source to verify that this is the currentversion before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One (1) airway, of the size specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table I shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table I. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	6 units	288 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit shall be packaged in a carton of appropriate size constructed in accordance with Specification PPP-B-566, PPP-B-665 or PPP-B-676 except that commercial colors shall be acceptable. Closure of carton shall be adequate to prevent spillage of contents under normal handling.

5.1.3.2 Intermediate package. Intermediate package shall be a carton of appropriate size constructed in accordance with Specification PPP-B-566 or PPP-B-676, except that commercial colors shall be acceptable, or PPP-B-636. Closure shall be adequate to prevent spillage of contents under normal handling.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with Specification PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table III; or PPP-B-636, class 2.

Source: https://assist.dla.mil -- Downloaded: 20260-21T08:09Z
Check the source to verify that this is the current version before use.

5.2.1.2 Case liner. Each level A box shall be lined with a waterproof case liner conforming to Specification MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto.

5.2.1.3 Strapping. Strapping, when required, shall be in accordance with applicable paragraphs of appendix thereto.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to Specification PPP-B-601, table I, using type III, class 1 plywood, or shall be designed for type 2 load, and constructed in accordance with Specification PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type I, class 1.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with Standard MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with Standard MIL-STD-129. In addition, the letter "L" shall precede the Federal stock number. One space shall be allowed between the letter "L" and the first digit of the stock number. The terms "Stock number" or "FSN" shall not be indicated. Service color markings are required only on direct shipments from the contractor's plant to overseas activities.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Size required (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:09Z
Check the source to verify that this is the current version before use.

6.2 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

6.3 Transportation description. Transportation description applicable to this item is:

Surgical Instruments NOI, in boxes
Carload minimum Weight: 10,000 pounds subject to Rule 34 of the Uniform Freight classification
Truckload minimum Weight: 10,000 pounds per National Motor Freight classification

Notice. When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, used, or sell any patented invention that may in any way be related thereto.

Custodians:

Army. - MD
Navy. - BuMed
Air Force. - AFCSG

Preparing activity:

Defense Supply Agency. - DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:09Z
Check the source to verify that this is the current version before use.