MIL-B-36225B
13 April 1976
SUPERSEDING
MIL-B-36225A(DSA-DM)
2 September 1971

MILITARY SPECIFICATION

BAND, PATIENT IDENTIFICATION

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers one type of plastic patient identification band. (see 6.1).

1.2 Classification. The identification bands shall be of the following classes and sizes, as specified:

1.2.1 Classes.

Class I - Continuous strip, supplied in roll form.
Class II - Supplied as individual bands, or in sheet form.

1.2.2 Sizes.

Size 1	- Mother-infant
Size 2	- Pediatric
Size 3	- Adult

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

QQ-S-766	- Steel Plates, Sheets, and Strip Corrosion-Resisting.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

FEDERAL

TT-I-563	- Ink, Writing
PPP-B-585	- Boxes, Wood, Wirebound.
PPP-B-601	- Boxes, Wood, Cleated-Plywood.
PPP-B-621	- Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	- Boxes, Shipping Fiberboard.

MILITARY

MIL-P-116	- Preservation - Packaging, Methods of
MIL-L-10547	- Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

MILITARY

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

24374 Band, Patient Identification Class I, continuous strip

(Copies of specifications, standards and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

TECHNICAL ASSOCIATION OF THE PULP AND PAPER INDUSTRY (TAPPI)

T400	Sampling of Paper
T410	Basis Weight of Paper and Paperboard.
T411	Thickness and Density of Paper.
T435	Hydrogen Ion Concentration (pH) of Paper Extracts.

(Application for copies should be addressed to the Technical Association of the Pulp and Paper Industry, One Dunwoody Park, Atlanta, GA 30341).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 First article. When specified in the procurement document, the contractor shall furnish sample unit(s) for first article inspection and approval (see 4.1.2.1 and 6.2).

3.2 Material.

3.2.1 Band. Class I band shall be a soft, pliable, natural color, plasticized, polyvinyl chloride, in tubular form. The vinyl film shall be dermally nontoxic, and shall contain no ingredient known to be a skin irritant or sensitizer. Class II band shall be a soft, pliable, plastic heat-sealed laminate. Material shall be plasticized, polyvinyl chloride, shall be dermally nontoxic, and contain no ingredient known to be a skin irritant or sensitizer.

3.2.2 Reinforcement strip. Class I reinforcement strip shall be a clear, polyethylene terephthalate film (see 6.4). Class II reinforcement strip shall be a biaxially-oriented polyvinyl chloride.

3.2.3 Insert card. Insert card shall be a white index bristol containing not less than 25 percent new rag or linen fibers; the remainder shall be free from unbleached or ground woodpulp. The paper shall comply with the following:

Thickness	0.0084 inch ± 0.0009
Acidity, pH value	4.8 min.
Weight (25 1/2 x 30 1/2, 1000 sheets)	.220 pounds, ± 5 percent.

3.2.3.1 Erasing quality. The erased card shall not show spreading of ink, when tested as specified in 4.4.1.

3.2.4 Clip seal device. Class I clip seal device shall be made of corrosion-resisting steel conforming to class 302, quarter-hard of QQ-S-766. A nickel-plated brass eyelet-rivet shall be used to secure the clip to the band. Class II clip seal device shall consist of two-piece locking snap molded of Delrin plastic, or equal.

3.3 Design and construction. The patient identification bands shall be supplied as class I or class II. The identification cards shall be supplied in pads. Design and construction for class I shall be in accordance with DPSC drawing 24374.

3.3.1 Band. Class I band shall be made in a flattened tubular form. The band shall be of uniform width, when flat, so that the insert card may be introduced smoothly and fit snugly when completely inserted. A reinforcement strip shall be secured by adhesive or heat-sealed to the inside of the

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

tubular band for its entire length. A clip seal shall be riveted to each band and shall provide secure closure without the use of tools and which cannot be opened by hand without destroying the fastening device. Class II band shall have a thickness of 0.022 ± 0.002 inch, except the mother-infant size shall be 0.024 ± 0.002. The plastic laminate shall contain a reinforcement strip, 0.0025 ± 0.0005 inch thick. The locking snap shall have a diameter of 3/8 ± 1/64 inch, and a thickness of 0.257 ± 0.010 inch when closed. The locking snap shall be attached 1/4 ± 1/64 inch from the end of the band, and shall permit secure closure by hand which cannot be opened by hand without destroying the fastening device. The side of the laminate near the patient's wrist shall contain a window so that the insert card can be introduced smoothly and fit snugly. All bands shall have soft, rounded edges. The band or window material of all bands shall have sufficient transparency to allow reading of standard typewriter printing at normal reading distance. When tested as specified in 4.4.4, there shall be no damage to the band after applying a weight of 20 pounds.

3.3.1.1 Mother-infant band (size 1). The class I mother-infant band (size 1) shall be provided in a continuous strip. The class I, size 1 band shall be furnished as a roll of not less than 200 sets, average, per roll; each set having one mother band and two infant bands. Die-cuts for band separation shall be made as shown on the drawing. The class II mother-infant band shall be provided in a set of two infant bands and one mother band. The bands in a set shall be attached together, but each band shall be easily separated from the set for use. The two bands for the infant shall be 6-5/8 ± 1/8 inches long from end to end, with the first adjustment hole 1/8 ± 1/16 inch from the end. Band shall have 14 evenly spaced holes for adjustment of band length to the infant. The infant band shall be 1/2 + 1/16, -0 inch wide in the area of band adjustment, and 3/4 + 1/16, -0 inch wide in the window area. The mother band shall be 1/2 + 1/16, -0 inch wide in the area of band adjustment, and 1 + 1/16, -0 inch wide in the window area. The mother band shall be 9-3/8 ± 1/8 inches long from end to end with the first adjustment hole 1/2 ± 1/8 inch from the end and shall have 11 evenly spaced holes for adjustment of band length to the mother. Each unit package shall contain 200 sets, that is, 400 infant and 200 mother bands.

3.3.1.2 Pediatric (size 2) and adult (size 3) bands. The class I pediatric and adult bands shall be provided in a continuous strip with indentations or die-cuts to allow easy separation. The adult and pediatric sizes shall be furnished in rolls with not less than 400 bands average on a roll. The class II adult size band shall be furnished in sheets of ten that are readily separated for use. The length of the adult band shall be 11-11/16 ± 1/8 inches from end to end, and shall have 12 evenly spaced holes for adjustment of band length to the patient. The first adjustment hole shall be 2-5/8 ± 1/8 inch from the end of the band. The width shall be 9/16 ± 1/16, -0 inch. The class II pediatric band shall be 8-1/8 ± 1/8 inches long, from the slot for insertion of the end tab to the opposite end; and shall have

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

14 holes for adjustment of the band length to the patient. The first adjustment hole shall be 1/8 ± 1/16 inch from the end of the band. The class II pediatric band shall be 1/2 + 1/16, -0 inch wide in the area of band adjustment and 3/4 + 1/16, -0 inch wide in the window area. Not less than 400 bands average of class II, size 1 or 2, shall be furnished in each unit package.

3.3.2 Insert cards. One or more pads, made up of sheets of paper perforated to provide individual insert cards, shall be provided. Each insert card shall consist of a lined writing space and a separable stub. The junction of the writing space and the separable stub shall be perforated to permit easy separation. The perforations separating one insert card from another and the perforations separating the two parts of an individual card shall allow one card to be separated from the pad without separating the two parts of the card. The insert cards shall be capable of being easily separated along the perforations without causing distortion to the card. The detached strip shall be free of protuberances along the edges making the card difficult to insert into the band. The inside cover of each pad of insert cards shall contain instructions for the use of the insert cards, and for the attachment of the bands. The card for class I, size 1 band (mother-infant) shall have a cut-out to allow visibility of the imprinted serial numbers (see 3.4.1). Cards for class I bands shall be constructed in accordance with DPSC drawing 24374. The following minimum quantity (average) of insert cards shall be supplied with each unit package (see 5.1.1).

Size	Quantity
1	720
2	480
3	480

The insert card for the class II bands shall also be furnished in one or more pads. Adult card shall be 5-1/16 inches long total, having a 1-1/16 inch long insertion tab, and shall be 27/64 ± 1/64 inch wide. The pediatric size card shall be 3 inches long having 15/16 inch long insertion tab, and shall be 5/8 ± 1/64 inch wide. The infant size card shall be 2-1/16 inches long and 5/8 ± 1/64 inch wide. The mother size card shall be 2-5/8 inches long and 7/8 ± 1/64 inch wide. Suitable insertion tabs shall be provided for the mother-infant size cards. Tolerance for card length shall be ± 1/8 inch.

3.3.3 Pen. One indelible type ballpoint pen whose ink, after drying, does not bleed in water shall be included in each unit package of bands (see 6.5). Test shall be performed as specified in 4.4.5.

3.4 Identification marking.

3.4.1 Mother-infant band. Each section of the class I mother-infant band shall be imprinted on the inside of the tube with the word "mother" or "baby" as applicable, and an identical serial number, for each set, located as shown on the drawing. The manufacturer's name or registered trademark

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

shall be imprinted inside the tube in the "mother" section of the band only. For the class II band, the manufacturer's name or registered trademark shall be printed at the end of the insert card which is placed near the locking snap. Each band of the mother-infant set shall be hot stamped, on the side away from the patient with the set's serial number. All markings shall be permanent and legible. Serial numbers shall be in sequence. The numbers shall be as indicated in the contract or purchase order.

3.4.2 Pediatric and adult band. Each class I pediatric and adult band shall have the manufacturer's name or registered trademark permanently and legibly imprinted inside the tube near the clip seal. For the class II pediatric and adult band, the required information shall be printed at the end of the insert card which is placed near the locking snap.

3.5 Instructions. Comprehensive instructions for use shall be furnished with each unit package.

3.6 Workmanship. Workmanship shall be first class throughout. Exposed edges of the clip seals shall be smooth and rounded. Identification bands and insert cards shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification is defined as both examination (such as visual and auditory investigation, without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

4.1.2.1 Classification of inspection. The inspection requirements specified herein are classified as follows:

a. First article inspection.
b. Quality conformance inspection.

4.1.2.1.1 First article inspection. First article inspection shall be performed on the bands when a first article sample is required (see 3.1). This inspection shall include examination of 4.3 and tests of 4.4.

4.1.2.1.2 Quality conformance inspection. Quality conformance inspection shall include the examination of 4.3 and tests of 4.4.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturers of the paper, plastic, and metals comprising the bands, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and table I.

TABLE I. Sampling for examination

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	S-2	1.0
Minor defects	S-2	4.0
For dimensional examination	S-1	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II. Unit of inspection for test 4.4.3 shall be one unit package; unit of inspection for tests 4.4.1 and 4.4.4 shall be one band. Acceptance number shall be zero (0) regardless of lot size.

TABLE II. Sampling for tests

Component	Characteristic	Requirement	Test procedure	Inspection level
Paper	Acidity	3.2.3	T435	1/

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

TABLE II. Sampling for tests (continued)

Component	Characteristic	Requirement	Test procedure	Inspection level
Paper	Thickness	3.2.3	T411	1/
	Weight	3.2.3	T410	1/
	Erasing	3.2.3.1	4.4.1	S-1
Pen	Water resistance	3.3.3	4.4.5	S-1
Clip seal (Class I)	Chemical and physical properties	3.2.4	4.4.2	4.4.2
Band	Quantity per unit package	3.3.1.1 3.3.1.2 3.3.2	4.4.3	S-2
	Strength	3.3.1	4.4.4	S-1

1/ Not less than 5 percent of mill rolls in a delivery of paper shall be sampled. Sampling shall be in accordance with TAPPI method T-400.

4.3 Examination. The bands shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the classification of defects shown in table III. Examination shall not be restricted to the classified possible defects listed below.

TABLE III. Classification of defects

Categories	Defects
Major	
101	Band not as specified.
102	Band does not contain a reinforcement strip.
103	Reinforcement strip not secured in place.
104	Band does not have seal device.
105	Seal does not close securely, or can be opened by hand.
106	Band does not separate easily from roll or sheet.
107	Ball point pen not supplied, not as specified.
108	Insert pad not supplied, not as specified.
109	Insert cards do not fit easily into band, not as specified.
110	Size 1 bands not serialized as specified.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of defects (continued)

Categories	Defects
Minor	
201	Insert cards do not separate easily.
202	Detached strip of insert card is not free of protuberances.
203	Ballpoint pen does not write legibly or is damaged.
204	Marking on band or card not present, not legible, not permanent.

4.3.2 Dimensional examination. The identification bands shall be examined for defects in dimensions. Any dimension not within the tolerance specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with all requirements contained in this specification.

4.4.1 Erasing quality. Characters shall be written on two paper cards of each selected package with a medium-point pen using ink in conformance with TT-I-563. The characters shall be erased, and then reapplied. There shall be no feathering of the inked characters upon application to the erased cards.

4.4.2 Chemical and physical properties, (class I). The steel shall be tested in accordance with the applicable procedures of QQ-S-766.

4.4.3 Quantity. The quantity of bands and insert cards in the selected packages from each lot shall be averaged to determine conformance to 3.3.1.1, 3.3.1.2 and 3.3.2.

4.4.4 Strength. The band shall be fastened, as if in normal use, from a stationary support. A weight of twenty pounds shall be applied to the band for a period of 10 minutes. After removal of the weight, the band and locking device shall be examined for any sign of damage.

4.4.5 Pen. Characters shall be written on the insert card and allowed to dry for 1 hour. The characters shall be rubbed lightly 5 times with a clean cloth dipped in cold tap water. Visual examination shall reveal no smearing or bleeding of the characters.

4.5 Packaging inspection. The inspection of the preservation-packaging and interior package marking shall be in accordance with the group A and B quality conformance inspection requirements of MIL-P-116. The inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

5. PACKAGING

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Unit of issue. One package containing the number of components as specified in 3.3.1.1, 3.3.1.2 or 3.3.3 for the class and size specified, constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Unit package. Each unit of issue shall be packaged in accordance with MIL-P-116, method III. Box shall be of appropriate size and design. Closure shall be adequate to prevent accidental opening.

5.1.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified. (see 6.2).

5.2.1 Level A. Six (6) units of issue of the class and size specified shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather resistant, grade V3c. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof in accordance with the appendix of PPP-B-636.

5.2.2 Level B. Six (6) units of issue of the class and size specified shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Unit package and exterior container. Each unit package and exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The item covered by this specification is intended to assure the positive identification of patients in medical facilities.

6.2 Ordering data. Procurement documents should specify the following:

a. Title, number and date of this specification.
b. Selection of applicable levels of packaging and packing (See 5.1 and 5.2).
c. Size and class required.
d. When a first article is required (see 3.1, 4.1.2.1 and 6.3).
e. Serial numbers for size 1 bands (see 3.4.1).

6.3 First article. When a first article is required, it shall be tested and approved under the appropriate provisions of 7-104.55 of the Armed Services Procurement Regulation. The contract or purchase order specifies the quantity, procedure for delivery and other specific instructions for inspection and approval of the first article.

6.4 Polyethylene terephthalate film specified herein is commonly known as "Mylar".

6.5 Pen shall be Fisher Pen Co.'s R-34, Hollister, Inc. 6801, or equal.

6.6 This specification covers the following items listed in the Federal Supply Catalog:

NSN	Item Identification
6530-00-982-5118	Band, Patient Identification, Pediatric Size, Plastic, 400s
6530-00-104-7631	Band, Patient Identification, Adult Size, Plastic, 400s
6530-00-104-7639	Band, Patient Identification, Mother-Infant Size, Plastic, 200s

6.7 Changes from previous issue. Asterisks are not used in this revision to identify changes with respect to the previous issue, due to the extensiveness of the changes.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6530-1337

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-B-36225B - BAND, PATIENT IDENTIFICATION

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

GPO 1974—603 680 6126 2-1

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.