MIL-F-36176A
3 October 1968
SUPERSEDING
MIL-F-36176(DM)
17 December 1963

MILITARY SPECIFICATION

FORCEPS, DENTAL ARTICULATING PAPER

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers dental forceps for use in carrying and holding dental articulating paper for occlusal registration.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
GG-P-106	Paper, Articulating, Dental.
MMM-A-250	Adhesive, Water-Resistant, (For Sealing Fiberboard Boxes).
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, (for Carton Sealing).

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:10Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

Military

MIL-P-116 Preservation, Methods of.

STANDARDS

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.

MIL-STD-129 Marking for Shipment and Storage.

DRAWING

DEFENSE PERSONNEL SUPPORT CENTER

20684 Forceps, Articulating Paper, Dental, 6 inches

(Copies of Specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 General requirements. The forceps shall be in accordance with the applicable requirements of GG-I-526 and as specified herein.

3.2 Material. Forceps shall be fabricated of class 4, type 420 martensitic stainless steel of GG-I-526, except that the carbon content shall be 0.10 percent minimum.

3.3 Design and construction. The forceps shall be cross-locking, self-closing spring forceps. Forceps shall be suitable for use in carrying and holding dental articulating paper (conforming to type I, class A or B of the latest issue in effect of GG-P-106) between the teeth while registering for occlusal prematurities. The halves of the forceps shall be joined by riveting or by securely spot welding as shown in Defense Personnel Support Center (DPSC) Drawing 20684. When pressure is released on the spring handle, all jaw (blade) serrations shall mesh properly as shown in DPSC Drawing 20684. Forceps shall conform to DPSC Drawing 20684 for style, design, and dimensions.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:10Z
Check the source to verify that this is the current version before use.

3.4 Hardness. After proper stress relieving the forceps shall have a Rockwell hardness of C40 to C45.

3.5 Finish. The forceps shall have a mirror finish except for the jaw(blade) serrations which shall have a satin to mirror finish. All edges of the forceps shall be smooth.

3.6 Deflection. The weight necessary to open the jaw (blade) tips a distance of 1/32 inch shall be 5 to 7 pounds when tested as specified in 4.3.1.

3.7 Corrosion resistance. The copper sulfate and boil tests as specified in GG-I-526 shall apply to the entire forceps.

3.8 Identification marking. Forceps shall be permanently and legibly marked in accordance with GG-I-526.

3.9 Workmanship. Workmanship shall be first class throughout. Forceps shall be free from defects such as burrs or sharp edges which detract from its appearance or may impair its serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in tables I and II (see 4.2.1) and table III (see 4.2.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:10Z
Check the source to verify that this is the current version before use.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQL) shall be as indicated in table I. Unit of product for examination purposes shall be one forceps.

Table I. Examination Criteria

Type defect	Inspection level	AQL Percent defective
Major A	II	1.0
Major B	S-1	1.5
Minor A	II	2.5
Minor B	S-1	4.0

4.2.2 Examination. The forceps shall be examined for defects, including, but not limited to, those listed in table II.

Table II. Classification of Defects

Categories	Defects
Major A:	
101	Forceps not a self-locking, self-closing, spring forceps.
102	Halves of forceps not securely joined by riveting or spot welding.
103	Jaw (blade) serrations do not mesh properly when pressure released on handle.
104	Finish of jaw (blade) not as specified.
Major B:	
151	Jaw (blade) dimensions not as specified.
152	Number of serrations on jaw (blade) not as specified.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:10Z
Check the source to verify that this is the current version before use.

Table II. (Contd.)

Minor A:	
201	Finish of handle not as specified.
202	Burrs, sharp edges present on forceps.
203	Instrument not permanently and legibly marked with the name or trademark of contractor.
204	Packaging, packing, and marking not as specified.
Minor B:	
251	Handle dimensions not as specified.

4.2.3 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table III except that the acceptance number shall be zero for all sample sizes. Unit of product for all tests except composition analysis shall be one forceps. Where feasible, the same sample unit shall be used for the determination of two or more different test characteristics.

Table III. Test Criteria

Characteristic	Reqmt.	Test procedure	Inspection level
Composition	3.2	4.3	*
Hardness	3.4	4.3	S-2
Deflection	3.6	4.3.1	S-2
Copper sulfate test	3.7 & GG-I-526	4.3	S-2
Boil test	3.7 & GG-I-526	4.3	S-2

*A sufficient amount from each lot of metal used in the construction of the forceps shall be selected for composition analysis.

4.3 Tests. Tests shall be conducted to determine compliance with specification requirements. Test procedures shall be as specified in GG-I-526 and as specified herein.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:10Z
Check the source to verify that this is the current version before use.

4.3.1 Deflection test. The forceps shall be placed on its side with the handle resting on a level surface and the jaws and cross-locking sections free to move. A downward force sufficient to open the jaw tips a distance of 1/32 inch shall be applied at 3 inches from the proximal (joined) end.

5. PREPARATION FOR DELIVERY.

5.1 Packaging. Packaging shall be level A or C, as specified (see 6.1).

5.1.1 Unit of issue. One forceps, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging quantities

Unit package	Intermediate package	Exterior container
1 unit	12 units	576 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Cleaning and drying. Each unit shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116.

5.1.3.2 Preservative. Not required.

5.1.3.3 Intimate wrap. Each unit shall be wrapped in neutral tissue or other suitable anticorrosive materials.

5.1.3.4 Unit package. Each unit shall be packaged in accordance with MIL-P-116, method IA-8.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:10Z
Check the source to verify that this is the current version before use.

5.1.3.5 Intermediate package. Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class weather-resistant. Boxes shall be closed by taping the top and bottom flaps with 2 inch wide tape conforming to PPP-T-76. Closing of bottom flaps with adhesive conforming to MMM-A-250 or metal stitches is permitted.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B, or C, as specified (see 6.1).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table III, using type I, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant. Closure and strapping shall be in accordance with appendix of applicable box specification.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic. Closure shall be in accordance with appendix of applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:10Z
Check the source to verify that this is the current version before use.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.

(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification does not include all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet Military requirements.

Custodians:
Army. - MD
Navy. - BuMed
Air Force. - AFMSC

Review activities:
Army. - MD
Navy. - BuMed
Air Force. - AFMSG

Preparing activity:
Defense Supply Agency - DM

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

CATEGORY N.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:10Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or serve to amend contractual requirements .

SPECIFICATION

MIL-F-36176A

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 65 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:10Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL CODE: ATS
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

☆ U. S. GOVERNMENT PRINTING OFFICE: 1968-341-513/A-2791

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:10Z
Check the source to verify that this is the current version before use.