MIL-L-36224(DM)
22 June 1964

MILITARY SPECIFICATION

LAP, OPHTHALMIC, CYLINDRICAL

1. SCOPE

1.1 This specification covers cylindrical ophthalmic laps designed with a convex surface, suitable for grinding ophthalmic lenses.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Set-Up, Paperboard.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, (for Carton Sealing).

Military

MIL-P-116	Preservation, Methods of.
MIL-B-10377	Box, Wood, Cleated, Veneer, Paper Overlaid.
MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

: FSC 6540 :

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

STANDARDS

Federal

Fed. Test Method Std. No. 151 Metals; Test Methods.

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

DRAWING

Military Medical Supply Agency

20226 Lap, Ophthalmic, Cylindrical.

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. Material shall be cast iron of appropriate quality, suitabl in all respects for the purpose intended.

3.2 Style, design, and construction. Each lap shall be a finished ophthalmic cylinder lap, designed with a convex surface to grind concave, shall have a flatback base, and shall be in accordance with the drawing. Laps shall be of the measurements in diopter base curve and in diopter cross curve, as specified (see 6.1), and as listed below for the appropriate stock number and item identification.

Stock No. and Item Identification	*(See Note)	Diopter Base Curve	Diopter Cross Curve
6540-655-8609		3.00	3.25
6540-655-8610		3.00	3.50

*Note: Item identification appearing on unit carton, intermediate package, and exterior container shall be written as follows: "Lap, Ophthalmic, Cylindrical: ____ Diopter Base Curve, _____ Diopter Cross Curve." The applicable information shall be inserted for the appropriate stock number.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

Stock No. and Item Identification	*(See Note) Diopter Base Curve	Diopter Cross Curve
6540-655-8611	3.00	3.75
6540-655-8612	3.00	4.00
6540-655-8613	3.00	4.25
6540-655-8614	3.00	4.50
6540-655-8615	3.00	4.75
6540-655-8616	3.00	5.00
6540-655-8617	3.00	5.25
6540-655-8618	3.00	5.50
6540-655-8619	3.00	5.75
6540-655-8620	3.00	6.00
6540-655-8621	3.00	6.25
6540-655-8622	3.00	6.50
6540-655-8623	3.00	6.75
6540-655-8624	3.00	7.00
6540-655-8625	3.00	7.25
6540-655-8627	3.00	7.50
6540-655-8629	3.00	7.75
6540-655-8630	3.00	8.00
6540-655-8631	3.00	8.25
6540-655-8634	3.00	8.50
6540-655-8635	3.00	8.75
6540-655-8636	3.00	9.00
6540-655-8637	3.25	3.50
6540-655-8638	3.25	3.75
6540-655-8639	3.25	4.00
6540-655-8640	3.25	4.25
6540-655-8641	3.25	4.50
6540-655-8642	3.25	4.75
6540-655-8643	3.25	5.00
6540-655-8644	3.25	5.25
6540-655-8645	3.25	5.50
6540-655-8646	3.25	5.75

*Note: Item identification appearing on unit carton, intermediate package, and exterior container shall be written as follows: "Lap, Ophthalmic, Cylindrical:_____ Diopter Base Curve,_____ Diopter Cross Curve." The applicable information shall be inserted for the appropriate stock number.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:14Z
Check the source to verify that this is the current version before use.

Stock No. and Item Identification	*(See Note) Diopter Base Curve	Diopter Cross Curve
6540-655-8647	3.25	6.00
6540-655-8648	3.25	6.25
6540-655-8649	3.25	6.50
6540-655-8650	3.25	6.75
6540-655-8651	3.25	7.00
6540-655-8652	3.25	7.25
6540-655-8653	3.25	7.50
6540-655-8654	3.25	7.75
6540-655-8656	3.25	8.00
6540-655-8657	3.25	8.25
6540-655-8658	3.25	8.50
6540-655-8659	3.25	8.75
6540-655-8660	3.25	9.00
6540-655-8661	3.25	9.25
6540-655-8662	3.50	3.75
6540-655-8663	3.50	4.00
6540-655-8664	3.50	4.25
6540-655-8666	3.50	4.50
6540-655-8667	3.50	4.75
6540-655-8668	3.50	5.00
6540-655-8669	3.50	5.25
6540-655-8670	3.50	5.50
6540-655-8671	3.50	5.75
6540-655-8672	3.50	6.00
6540-655-8674	3.50	6.25
6540-655-8675	3.50	6.50
6540-655-8676	3.50	6.75
6540-655-8677	3.50	7.00
6540-655-8678	3.50	7.25
6540-655-8679	3.50	7.50
6540-655-8680	3.50	7.75

*Note: Item identification appearing on unit carton, intermediate package, and exterior container shall be written as follows: "Lap, Ophthalmic. Cylindrical:_____ Diopter Base Curve,_____ Diopter Cross Curve." The applicable information shall be inserted for the appropriate stock number.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:14Z
Check the source to verify that this is the currentversion before use.

Stock No. and Item Identification	*(See Note) Diopter Base Curve	Diopter Cross Curve
6540-655-8681	3.50	8.00
6540-655-8682	3.50	8.25
6540-655-8683	3.50	8.50
6540-655-8684	3.50	8.75
6540-655-8685	3.50	9.00
6540-655-8686	3.50	9.25
6540-655-8688	3.50	9.50
6540-655-8689	3.75	4.00
6540-655-8690	3.75	4.25
6540-655-8691	3.75	4.50
6540-655-8692	3.75	4.75
6540-655-8693	3.75	5.00
6540-655-8694	3.75	5.25
6540-655-8695	3.75	5.50
6540-655-8696	3.75	5.75
6540-655-8697	3.75	6.00
6540-655-8698	3.75	6.25
6540-655-8699	3.75	6.50
6540-655-8700	3.75	6.75
6540-655-8701	3.75	7.00
6540-655-8702	3.75	7.25
6540-655-8703	3.75	7.50
6540-655-8704	3.75	7.75
6540-655-8705	3.75	8.00
6540-655-8706	3.75	8.25
6540-655-8707	3.75	8.50
6540-655-8708	3.75	8.75
6540-655-8709	3.75	9.00
6540-655-8710	3.75	9.25
6540-655-8711	3.75	9.50
6540-655-8712	3.75	9.75
6540-655-8713	4.00	4.25

*Note: Item id ntification appearing on unit carton, intermediate package, and exterior container shall be written as follows: "Lap, Ophthalmic, Cylindrical:_____ Diopter Base Curve,_____ Diopter Cross Curve." The applicable information shall be inserted for the appropriate stock number.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:14Z
Check the source to verify that this is the current version before use.

Stock No. and Item Identification	*(See Note) Diopter Base Curve	Diopter Cross Curve
6540-655-8714	4.00	4.50
6540-655-8715	4.00	4.75
6540-655-8716	4.00	5.00
6540-655-8717	4.00	5.25
6540-655-8718	4.00	5.50
6540-655-8719	4.00	5.75
6540-655-8720	4.00	6.00
6540-655-8721	4.00	6.25
6540-655-8722	4.00	6.50
6540-655-8723	4.00	6.75
6540-655-8724	4.00	7.00
6540-655-8725	4.00	7.25
6540-655-8726	4.00	7.50
6540-655-8727	4.00	7.75
6540-655-8728	4.00	8.00
6540-655-8729	4.00	8.25
6540-655-8730	4.00	8.50
6540-655-8731	4.00	8.75
6540-655-8732	4.00	9.00
6540-655-8733	4.00	9.25
6540-655-8734	4.00	9.50
6540-655-8735	4.00	9.75
6540-655-8736	4.00	10.00
6540-655-8737	4.25	4.50
6540-655-8739	4.25	4.75
6540-655-8740	4.25	5.00
6540-655-8741	4.25	5.25
6540-655-8742	4.25	5.50
6540-655-8743	4.25	5.75
6540-655-8744	4.25	6.00
6540-655-8745	4.25	6.25
6540-655-8746	4.25	6.50

*Note: Item identification appearing on unit carton, intermediate package, and exterior container shall be written as follows: "Lap, Ophthalmic, Cylindrical:_____ Diopter Base Curve,_____ Diopter Cross Curve." The applicable information shall be inserted for the appropriate stock number.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:14Z
Check the source to verify that this is the current version before use.

Stock No. and Item Identification	*(See Note) Diopter Base Curve	Diopter Cross Curve
6540-655-8747	4.25	6.75
6540-655-8748	4.25	7.00
6540-655-8749	4.25	7.25
6540-655-8750	4.25	7.50
6540-655-8751	4.25	7.75
6540-655-8752	4.25	8.00
6540-655-8753	4.25	8.25
6540-655-8754	4.25	8.50
6540-655-8755	4.25	8.75
6540-655-8756	4.25	9.00
6540-655-8757	4.25	9.25
6540-655-8758	4.25	9.50
6540-655-8759	4.25	9.75
6540-655-8760	4.25	10.00
6540-655-8761	4.25	10.25
6540-655-8762	4.50	4.75
6540-655-8763	4.50	5.00
6540-655-8764	4.50	5.25
6540-655-8765	4.50	5.50
6540-655-8766	4.50	5.75
6540-655-8767	4.50	6.00
6540-655-8768	4.50	6.25
6540-655-8769	4.50	6.50
6540-655-8770	4.50	6.75
6540-655-8771	4.50	7.00
6540-655-8772	4.50	7.25
6540-655-8773	4.50	7.50
6540-655-8774	4.50	7.75
6540-655-8775	4.50	8.00
6540-655-8776	4.50	8.25
6540-655-8777	4.50	8.50

*Note: Item identification appearing on unit carton, intermediate package, and exterior container shall be written as follows: "Lap, Ophthalmic, Cylindrical:_____ Diopter Base Curve,_____ Diopter Cross Curve." The applicable information shall be inserted for the appropriate stock number.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

Stock No. and Item Identification	*(See Note) Diopter Base Curve	Diopter Cross Curve
6540-655-8778	4.50	8.75
6540-655-8779	4.50	9.00
6540-655-8780	4.50	9.25
6540-655-8781	4.50	9.50
6540-655-8782	4.50	9.75
6540-655-8783	4.50	10.00
6540-655-8784	4.50	10.25
6540-655-8785	4.50	10.50
6540-655-8786	4.75	5.00
6540-655-8787	4.75	5.25
6540-655-8788	4.75	5.50
6540-655-8789	4.75	5.75
6540-655-8790	4.75	6.00
6540-655-8791	4.75	6.25
6540-655-8792	4.75	6.50
6540-655-8793	4.75	6.75
6540-655-8794	4.75	7.00
6540-655-8795	4.75	7.25
6540-655-8796	4.75	7.50
6540-655-8797	4.75	7.75
6540-655-8798	4.75	8.00
6540-655-8799	4.75	8.25
6540-655-8800	4.75	8.50
6540-655-8801	4.75	8.75
6540-655-8802	4.75	9.00
6540-655-8803	4.75	9.25
6540-655-8804	4.75	9.50
6540-655-8805	4.75	9.75
6540-655-8806	4.75	10.00
6540-655-8807	4.75	10.25
6540-655-8808	4.75	10.50
6540-655-8809	4.75	10.75
6540-655-8810	5.00	5.25

*Note: Item identification appearing on unit carton, intermediate package, and exterior container shall be written as follows: "Lap, Ophthalmic, Cylindrical:_____ Diopter Base Curve,_____ Diopter Cross Curve." The applicable information shall be inserted for the appropriate stock number.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

Stock No. and Item Identification *(See Note)	Diopter Base Curve	Diopter Cross Curve
6540-655-8811	5.00	5.50
6540-655-8812	5.00	5.75
6540-655-8813	5.00	6.00
6540-655-8814	5.00	6.25
6540-655-8815	5.00	6.50
6540-655-8816	5.00	6.75
6540-655-8817	5.00	7.00
6540-655-8818	5.00	7.25
6540-655-8819	5.00	7.50
6540-655-8820	5.00	7.75
6540-655-8821	5.00	8.00
6540-655-8822	5.00	8.25
6540-655-8823	5.00	8.50
6540-655-8824	5.00	8.75
6540-655-8825	5.00	9.00
6540-655-8826	5.00	9.25
6540-655-8827	5.00	9.50
6540-655-8828	5.00	9.75
6540-655-8829	5.00	10.00
6540-655-8830	5.00	10.25
6540-655-8831	5.00	10.50
6540-655-8832	5.00	10.75
6540-655-8833	5.00	11.00
6540-655-8834	5.25	5.50
6540-655-8835	5.25	5.75
6540-655-8836	5.25	6.00
6540-655-8837	5.25	6.25
6540-655-8838	5.25	6.50
6540-655-8839	5.25	6.75
6540-655-8840	5.25	7.00
6540-655-8841	5.25	7.25

*Note: Item identification appearing on unit carton, intermediate package, and exterior container shall be written as follows: "Lap, Ophthalmic, Cylindrical:_____ Diopter Base Curve,_____ Diopter Cross Curve." The applicable information shall be inserted for the appropriate stock number.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:14Z
Check the source to verify that this is the current version before use.

Stock No. and Item Identification	*(See Note) Diopter Base Curve	Diopter Cross Curve
6540-655-8842	5.25	7.50
6540-655-8843	5.25	7.75
6540-655-8844	5.25	8.00
6540-655-8845	5.25	8.25
6540-655-8846	5.25	8.50
6540-655-8847	5.25	8.75
6540-655-8848	5.25	9.00
6540-655-8849	5.25	9.25
6540-655-8850	5.25	9.50
6540-655-8851	5.25	9.75
6540-655-8852	5.25	10.00
6540-655-8853	5.25	10.25
6540-655-8854	5.25	10.50
6540-655-8855	5.25	10.75
6540-655-8856	5.25	11.00
6540-655-8857	5.25	11.25
6540-655-8858	5.50	5.75
6540-655-8859	5.50	6.00
6540-655-8860	5.50	6.25
6540-655-8861	5.50	6.50
6540-655-8862	5.50	6.75
6540-655-8863	5.50	7.00
6540-655-8864	5.50	7.25
6540-655-8865	5.50	7.50
6540-655-8866	5.50	7.75
6540-655-8867	5.50	8.00
6540-655-8868	5.50	8.25
6540-655-8869	5.50	8.50
6540-655-8870	5.50	8.75
6540-655-8871	5.50	9.00
6540-655-8872	5.50	9.25

*Note: Item identification appearing on unit carton, intermediate package, and exterior container shall be written as follows: "Lap, Ophthalmic, Cylindrical:_____ Diopter Base Curve,_____ Diopter Cross Curve." The applicable information shall be inserted for the appropriate stock number.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:14Z
Check the source to verify that this is the currentversion before use.

Stock No. and Item Identification *(See Note)	Diopter Base Curve	Diopter Cross Curve
6540-655-8873	5.50	9.50
6540-655-8874	5.50	9.75
6540-655-8875	5.50	10.00
6540-655-8876	5.50	10.25
6540-655-8877	5.50	10.50
6540-655-8878	5.50	10.75
6540-655-8879	5.50	11.00
6540-655-8880	5.50	11.25
6540-655-8881	5.50	11.50
6540-655-8882	5.75	6.00
6540-655-8883	5.75	6.25
6540-655-8884	5.75	6.50
6540-655-8885	5.75	6.75
6540-655-8886	5.75	7.00
6540-655-8887	5.75	7.25
6540-655-8888	5.75	7.50
6540-655-8889	5.75	7.75
6540-655-8890	5.75	8.00
6540-655-8891	5.75	8.25
6540-655-8892	5.75	8.50
6540-655-8893	5.75	8.75
6540-655-8894	5.75	9.00
6540-655-8895	5.75	9.25
6540-655-8896	5.75	9.50
6540-655-8897	5.75	9.75
6540-655-8898	5.75	10.00
6540-655-8899	5.75	10.25
6540-655-8900	5.75	10.50
6540-655-8901	5.75	10.75
6540-655-8902	5.75	11.00

*Note: Item identification appearing on unit carton, intermediate package, and exterior container shall be written as follows: "Lap, Ophthalmic, Cylindrical:_____ Diopter Base Curve,_____ Diopter Cross Curve." The applicable information shall be inserted for the appropriate stock number.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:14Z
Check the source to verify that this is the current version before use.

Stock No. and Item Identification	*(See Note) Diopter Base Curve	Diopter Cross Curve
6540-655-8903	5.75	11.25
6540-655-8904	5.75	11.50
6540-655-8905	5.75	11.75
6540-655-8906	6.00	6.25
6540-655-8907	6.00	6.50
6540-655-8908	6.00	6.75
6540-655-8909	6.00	7.00
6540-655-8910	6.00	7.25
6540-655-8911	6.00	7.50
6540-655-8912	6.00	7.75
6540-655-8913	6.00	8.00
6540-655-8914	6.00	8.25
6540-655-8915	6.00	8.50
6540-655-8916	6.00	8.75
6540-655-8917	6.00	9.00
6540-655-8918	6.00	9.25
6540-655-8919	6.00	9.50
6540-655-8920	6.00	9.75
6540-655-8921	6.00	10.00
6540-655-8922	6.00	10.25
6540-655-8923	6.00	10.50
6540-655-8924	6.00	10.75
6540-655-8925	6.00	11.00
6540-655-8926	6.00	11.25
6540-655-8927	6.00	11.50
6540-655-8928	6.00	11.75
6540-655-8929	6.00	12.00
6540-655-8930	6.25	6.50
6540-655-8931	6.25	6.75
6540-655-8932	6.25	7.00
6540-655-8933	6.25	7.25

*Note: Item identification appearing on unit carton, intermediate package, and exterior container shall be written as follows: "Lap, Ophthalmic, Cylindrical:_____ Diopter Base Curve,_____ Diopter Cross Curve." The applicable information shall be inserted for the appropriate stock number.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

Stock No. and Item Identification	*(See Note) Diopter Base Curve	Diopter Cross Curve
6540-655-8934	6.25	7.50
6540-655-8935	6.25	7.75
6540-655-8936	6.25	8.00
6540-655-8937	6.25	8.25
6540-655-8938	6.25	8.50
6540-655-8939	6.25	8.75
6540-655-8940	6.25	9.00
6540-655-8942	6.25	9.25
6540-655-8944	6.25	9.50
6540-655-8945	6.25	9.75
6540-655-8946	6.25	10.00
6540-655-8947	6.25	10.25
6540-655-8949	6.25	10.50
6540-655-8950	6.25	10.75
6540-655-8951	6.25	11.00
6540-655-8952	6.25	11.25
6540-655-8953	6.25	11.50
6540-655-8954	6.25	11.75
6540-655-8955	6.25	12.00
6540-655-8956	6.25	12.25
6540-655-8957	6.50	6.75
6540-655-8958	6.50	7.00
6540-655-8959	6.50	7.25
6540-655-8960	6.50	7.50
6540-655-8961	6.50	7.75
6540-655-8962	6.50	8.00
6540-655-8963	6.50	8.25
6540-655-8964	6.50	8.50
6540-655-8965	6.50	8.75
6540-655-8966	6.50	9.00
6540-655-8967	6.50	9.25

*Note: Item identification appearing on unit carton, intermediate package, and exterior container shall be written as follows: "Lap, Ophthalmic, Cylindrical:____ Diopter Base Curve,____ Diopter Cross Curve." The applicable information shall be inserted for the appropriate stock number.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:14Z
Check the source to verify that this is the current version before use.

Stock No. and Item Identification	*(See Note)	Diopter Base Curve	Diopter Cross Curve
6540-655-8968		6.50	9.50
6540-655-8969		6.50	9.75
6540-655-8970		6.50	10.00
6540-655-8972		6.50	10.25
6540-655-8973		6.50	10.50
6540-655-8974		6.50	10.75
6540-655-8975		6.50	11.00
6540-655-8976		6.50	11.25
6540-655-8977		6.50	11.50
6540-655-8978		6.50	11.75
6540-655-8979		6.50	12.00
6540-655-8980		6.50	12.25
6540-655-8981		6.50	12.50
6540-655-8982		6.75	7.00
6540-655-8983		6.75	7.25
6540-655-8984		6.75	7.50
6540-655-8985		6.75	7.75
6540-655-8986		6.75	8.00
6540-655-8987		6.75	8.25
6540-655-8988		6.75	8.50
6540-655-8989		6.75	8.75
6540-655-8990		6.75	9.00
6540-655-8991		6.75	9.25
6540-655-8992		6.75	9.50
6540-655-8993		6.75	9.75
6540-655-8994		6.75	10.00
6540-655-8995		6.75	10.25
6540-655-8996		6.75	10.50
6540-655-8997		6.75	10.75
6540-655-8998		6.75	11.00
6540-655-8999		6.75	11.25
6540-655-9000		6.75	11.50

*Note: Item identification appearing on unit carton, intermediate package, and exterior container shall be written as follows: "Lap, Ophthalmic, Cylindrical:_____ Diopter Base Curve,_____ Diopter Cross Curve." The applicable information shall be inserted for the appropriate stock number.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

Stock No. and Item Identification	*(See Note) Diopter Base Curve	Diopter Cross Curve
6540-655-9001	6.75	11.75
6540-655-9002	6.75	12.00
6540-655-9003	6.75	12.25
6540-655-9004	6.75	12.50
6540-655-9005	6.75	12.75
6540-655-9006	7.00	7.25
6540-655-9007	7.00	7.50
6540-655-9008	7.00	7.75
6540-655-9009	7.00	8.00
6540-655-9010	7.00	8.25
6540-655-9011	7.00	8.50
6540-655-9012	7.00	8.75
6540-655-9013	7.00	9.00
6540-655-9014	7.00	9.25
6540-655-9015	7.00	9.50
6540-655-9016	7.00	9.75
6540-655-9017	7.00	10.00
6540-655-9018	7.00	10.25
6540-655-9019	7.00	10.50
6540-655-9020	7.00	10.75
6540-655-9021	7.00	11.00
6540-655-9022	7.00	11.25
6540-655-9023	7.00	11.50
6540-655-9024	7.00	11.75
6540-655-9025	7.00	12.00
6540-655-9026	7.00	12.25
6540-655-9027	7.00	12.50
6540-655-9028	7.00	12.75
6540-655-9029	7.00	13.00
6540-655-9030	7.25	7.50
6540-655-9031	7.25	7.75

*Note: Item identification appearing on unit carton, intermediate package, and exterior container shall be written as follows: "Lap, Ophthalmic, Cylindrical,_____ Diopter Base Curve,_____ Diopter Cross Curve." The applicable information shall be inserted for the appropriate stock number.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:14Z
Check the source to verify that this is the currentversion before use.

Stock No. and Item Identification	*(See Note) Diopter Base Curve	Diopter Cross Curve
6540-655-9032	7.25	8.00
6540-655-9033	7.25	8.25
6540-655-9034	7.25	8.50
6540-655-9035	7.25	8.75
6540-655-9036	7.25	9.00
6540-655-9037	7.25	9.25
6540-655-9038	7.25	9.50
6540-655-9039	7.25	9.75
6540-655-9040	7.25	10.00
6540-655-9041	7.25	10.25
6540-655-9042	7.25	10.50
6540-655-9043	7.25	10.75
6540-655-9044	7.25	11.00
6540-655-9045	7.25	11.25
6540-655-9046	7.25	11.50
6540-655-9047	7.25	11.75
6540-655-9048	7.25	12.00
6540-655-9049	7.25	12.25
6540-655-9050	7.25	12.50
6540-655-9051	7.25	12.75
6540-655-9052	7.25	13.00
6540-655-9053	7.25	13.25
6540-655-9054	7.50	7.75
6540-655-9055	7.50	8.00
6540-655-9056	7.50	8.25
6540-655-9057	7.50	8.50
6540-655-9058	7.50	8.75
6540-655-9059	7.50	9.00
6540-655-9060	7.50	9.25
6540-655-9061	7.50	9.50
6540-655-9062	7.50	9.75

*Note: Item identification appearing on unit carton, intermediate package, and exterior container shall be written as follows: "Lap, Ophthalmic, Cylindrical:_____ Diopter Base Curve,_____ Diopter Cross Curve." The applicable information shall be inserted for the appropriate stock number.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

Stock No. and Item Identification	*(See Note) Diopter Base Curve	Diopter Cross Curve
6540-655-9063	7.50	10.00
6540-655-9064	7.50	10.25
6540-655-9065	7.50	10.50
6540-655-9066	7.50	10.75
6540-655-9067	7.50	11.00
6540-655-9068	7.50	11.25
6540-655-9069	7.50	11.50
6540-655-9070	7.50	11.75
6540-655-9071	7.50	12.00
6540-655-9072	7.50	12.25
6540-655-9073	7.50	12.50
6540-655-9074	7.50	12.75
6540-655-9075	7.50	13.00
6540-655-9076	7.50	13.25
6540-655-9077	7.50	13.50
6540-655-9078	7.75	8.00
6540-655-9079	7.75	8.25
6540-655-9080	7.75	8.50
6540-655-9081	7.75	8.75
6540-655-9082	7.75	9.00
6540-655-9083	7.75	9.25
6540-655-9084	7.75	9.50
6540-655-9085	7.75	9.75
6540-655-9086	7.75	10.00
6540-655-9087	7.75	10.25
6540-655-9088	7.75	10.50
6540-655-9089	7.75	10.75
6540-655-9090	7.75	11.00
6540-655-9091	7.75	11.25
6540-655-9092	7.75	11.50
6540-655-9093	7.75	11.75

*Note: Item identification appearing on unit carton, intermediate package, and exterior container shall be written as follows: "Lap, Ophthalmic, Cylindrical:_____ Diopter Base Curve,_____ Diopter Cross Curve." The applicable information shall be inserted for the appropriate stock number.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:14Z
Check the source to verify that this is the current version before use.

Stock No. and Item Identification	*(See Note) Diopter Base Curve	Diopter Cross Curve
6540-655-9094	7.75	12.00
6540-655-9095	7.75	12.25
6540-655-9096	7.75	12.50
6540-655-9097	7.75	12.75
6540-655-9098	7.75	13.00
6540-655-9099	7.75	13.25
6540-655-9100	7.75	13.50
6540-655-9101	7.75	13.75
6540-655-9102	8.00	8.25
6540-655-9103	8.00	8.50
6540-655-9104	8.00	8.75
6540-655-9105	8.00	9.00
6540-655-9106	8.00	9.25
6540-655-9107	8.00	9.50
6540-655-9108	8.00	9.75
6540-655-9109	8.00	10.00
6540-655-9110	8.00	10.25
6540-655-9111	8.00	10.50
6540-655-9112	8.00	10.75
6540-655-9113	8.00	11.00
6540-655-9114	8.00	11.25
6540-655-9115	8.00	11.50
6540-655-9116	8.00	11.75
6540-655-9117	8.00	12.00
6540-655-9118	8.00	12.25
6540-655-9119	8.00	12.50
6540-655-9120	8.00	12.75
6540-655-9121	8.00	13.00
6540-655-9122	8.00	13.25
6540-655-9123	8.00	13.50
6540-655-9124	8.00	13.75

*Note: Item identification appearing on unit carton, intermediate package, and exterior container shall be written as follows: "Lap, Ophthalmic, Cylindrical:_____ Diopter Base Curve,_____ Diopter Cross Curve." The applicable information shall be inserted for the appropriate stock number.

Source: https://assist.dla.mil -- Downloaded: 20260-21T08:14Z
Check the source to verify that this is the current version before use.

Stock No. and Item Identification	*(See Note) Diopter Base Curve	Diopter Cross Curve
6540-655-9125	8.00	14.00
6540-655-9126	8.25	8.50
6540-655-9127	8.25	8.75
6540-655-9128	8.25	9.00
6540-655-9129	8.25	9.25
6540-655-9130	8.25	9.50
6540-655-9131	8.25	9.75
6540-655-9132	8.25	10.00
6540-655-9133	8.25	10.25
6540-655-9134	8.25	10.50
6540-655-9135	8.25	10.75
6540-655-9136	8.25	11.00
6540-655-9137	8.25	11.25
6540-655-9138	8.25	11.50
6540-655-9139	8.25	11.75
6540-655-9140	8.25	12.00
6540-655-9141	8.25	12.25
6540-655-9142	8.25	12.50
6540-655-9143	8.25	12.75
6540-655-9144	8.25	13.00
6540-655-9145	8.25	13.25
6540-655-9146	8.25	13.50
6540-655-9147	8.25	13.75
6540-655-9148	8.25	14.00
6540-655-9149	8.25	14.25
6540-655-9150	8.50	8.75
6540-655-9151	8.50	9.00
6540-655-9152	8.50	9.25
6540-655-9153	8.50	9.50
6540-655-9154	8.50	9.75
6540-655-9155	8.50	10.00
6540-655-9156	8.50	10.25

*Note: Item identification appearing on unit carton, intermediate package, and exterior container shall be written as follows: "Lap, Ophthalmic, Cylindrical:_____ Diopter Base Curve,_____ Diopter Cross Curve." The applicable information shall be inserted for the appropriate stock number.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:14Z
Check the source to verify that this is the current version before use.

Stock No. and Item Identification	*(See Note) Diopter Base Curve	Diopter Cross Curve
6540-655-9157	8.50	10.50
6540-655-9158	8.50	10.75
6540-655-9159	8.50	11.00
6540-655-9160	8.50	11.25
6540-655-9161	8.50	11.50
6540-655-9162	8.50	11.75
6540-655-9163	8.50	12.00
6540-655-9164	8.50	12.25
6540-655-9165	8.50	12.50
6540-655-9166	8.50	12.75
6540-655-9167	8.50	13.00
6540-655-9168	8.50	13.25
6540-655-9169	8.50	13.50
6540-655-9170	8.50	13.75
6540-655-9171	8.50	14.00
6540-655-9172	8.50	14.25
6540-655-9173	8.50	14.50
6540-655-9174	8.75	9.00
6540-655-9175	8.75	9.25
6540-655-9176	8.75	9.50
6540-655-9177	8.75	9.75
6540-655-9178	8.75	10.00
6540-655-9179	8.75	10.25
6540-655-9180	8.75	10.50
6540-655-9181	8.75	10.75
6540-655-9182	8.75	11.00
6540-655-9183	8.75	11.25
6540-655-9184	8.75	11.50
6540-655-9185	8.75	11.75
6540-655-9186	8.75	12.00
6540-655-9187	8.75	12.25
6540-655-9188	8.75	12.50

*Note: Item identification appearing on unit carton, intermediate package, and exterior container shall be written as follows: "Lap, Ophthalmic, Cylindrical:_____ Diopter Base Curve,_____ Diopter Cross Curve.". The applicable information shall be inserted for the appropriate stock number.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:14Z
Check the source to verify that this is the current version before use.

Stock No. and Item Identification	*(See Note) Diopter Base Curve	Diopter Cross Curve
6540-655-9189	8.75	12.75
6540-655-9190	8.75	13.00
6540-655-9191	8.75	13.25
6540-655-9192	8.75	13.50
6540-655-9193	8.75	13.75
6540-655-9194	8.75	14.00
6540-655-9195	8.75	14.25
6540-655-9196	8.75	14.50
6540-655-9197	8.75	14.75
6540-655-9198	9.00	9.25
6540-655-9199	9.00	9.50
6540-655-9200	9.00	9.75
6540-655-9201	9.00	10.00
6540-655-9202	9.00	10.25
6540-655-9203	9.00	10.50
6540-655-9204	9.00	10.75
6540-655-9205	9.00	11.00
6540-655-9206	9.00	11.25
6540-655-9207	9.00	11.50
6540-655-9208	9.00	11.75
6540-655-9209	9.00	12.00
6540-655-9210	9.00	12.25
6540-655-9211	9.00	12.50
6540-655-9212	9.00	12.75
6540-655-9213	9.00	13.00
6540-655-9214	9.00	13.25
6540-655-9215	9.00	13.50
6540-655-9216	9.00	13.75
6540-655-9217	9.00	14.00
6540-655-9218	9.00	14.25
6540-655-9219	9.00	14.50
6540-655-9220	9.00	14.75
6540-655-9221	9.00	15.00

*Note: Item identification appearing on unit carton, intermediate package, and exterior container shall be written as follows: "Lap, Ophthalmic, Cylindrical:_____ Diopter Base Curve,_____ Diopter Cross Curve." The applicable information shall be inserted for the appropriate stock number.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:14Z
Check the source to verify that this is the current version before use.

3.3 Preproduction sample. Unless otherwise specified, a sample preproduction lap shall be submitted for approval to the Defense Medical Supply Center, to determine conformance with the requirements of the specification (see 6.1).

3.4 Marking for identification. Each lap shall be permanently and legibly marked with the appropriate diopter base curve and diopter cross curve as listed in 3.2. In addition, each lap shall be permanently and legibly marked with the name or registered trademark of the manufacturer.

3.5 Workmanship. Workmanship shall be first class throughout. All burs and sharp edges shall be removed, including any burs which might result from application of marking for identification. All surfaces shall be free from nicks, gouges, cracks, holes, toolmarks, etc. Cylindrical ophthalmic laps shall be free from defects which detract from their appearance or may impair their serviceability. Laps shall be suitable in all respects for the purpose intended.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.1.3 Preproduction sample approval. Preproduction sample, when required, shall be inspected for compliance with the requirements of this specification (see 3.3).

4.2 Sampling. Sampling shall be conducted in accordance with Standard MIL-STD-105.

	Inspection level	AQL (percent defective)	Minimum sample size
Dimensional examination	S-1	*	3

* Acceptance number shall be zero (0) for all sample sizes.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

4.2.1 For chemical testing of metals. Four (4) ounces of each lot of each metal used in the fabrication of the laps shall be selected for chemical analysis as specified in 4.4.1.

4.3 Examination. The laps shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Dimensional examination. The laps shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Chemical analysis of the metals. The samples selected in accordance with 4.2.3 shall be analyzed in conformance with methods 111 or 112 of Federal Test Method Standard No. 151. In case of dispute, method 111 shall be used for referee analysis.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One (1) lap, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table I shall be packaged in each unit and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table I - Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	Not required	36 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:14Z
Check the source to verify that this is the current version before use.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit shall be packaged in accordance with Specification MIL-P-116, method I, using P-1 preservative, and shall be further packaged in a carton of appropriate size constructed in accordance with Specification PPP-B-566 or PPP-B-676, except that commercial colors shall be acceptable. Each carton shall be adequately secured to prevent accidental opening.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with Specification PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table I; PPP-B-636, type I, class 2; or MIL-B-10377, table II.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to Specification MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix of Specification PPP-B-636. In addition, all seams, including manufacturer's joint, shall be closed by tape conforming to the requirements of Specification PPP-T-76. The tape shall be not less than 2 inches wide. Three strips shall be applied to both top and bottom so that all seams and joints shall be covered their full length. Tape applied over the lengthwise seams shall extend a minimum of 2 inches onto each of the end panels. Tape applied over the end edge seams shall extend a minimum of 2 inches along the side and top, and side and bottom panels.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with applicable box specification.

NOTE: Strapping will not be required for shipment forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to Specification PPP-B-601, table I, using type III, class 1 plywood, or shall be designed for type 2 load, and constructed in accordance with Specification PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; PPP-B-636, type I, class 1; or MIL-B-10377, table I.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Unit packages. Each unit package shall be marked in accordance with Standard MIL-STD-129, except that when labels are utilized, waterproofing shall not be required. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with Standard MIL-STD-129. In addition, the letter "L" shall precede the Federal stock number. One space shall be allowed between the letter "L" and the first digit of the stock number. The terms "Stock number" or "FSN" shall not be indicated.

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Required measurements in diopter base and in diopter cross curve (see 3.2).
(c) When preproduction sample is not required (see 3.3).
(d) Selection of applicable levels of packing (see 5.2).

6.2 Bausch and Lomb Optical Co.'s Catalog No. 21-12-42 and Shuron Optical Co.'s Catalog No. 48-A lap are considered to meet the specification requirements.

6.3 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

Custodians:
Army. - MD
Navy. - BuMed
Air Force. - AFMSG

Review activities:
Army. - MD
Navy. - BuMed
Air Force. - AFMSG

Preparing activity:
Defense Supply Agency. - DM

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

FORM APPROVED BUDGET BUREAU NO. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use of the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity.

SPECIFICATION: MIL-L-36224(DX), DATED 22 June 1964 - LAP, OPHTHALMIC, CYLINDRICAL

ORGANIZATION

CITY AND STATE

CONTRACT NO.

QUANTITY OF ITEMS PROCURED

DOLLAR AMOUNT
$

MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?
   ☐ YES ☐ NO IF "YES" IN WHAT WAY?

4. REMARKS (ATTACH ANY PERTINENT DATA WHICH MAY BE OF USE IN IMPROVING THIS SPECIFICATION. IF THERE ARE ADDITIONAL PAPERS, ATTACH TO FORM AND PLACE BOTH IN AN ENVELOPE ADDRESSED PREPARING ACTIVITY)

SUBMITTED BY (PRINTED OR TYPED NAME AND ACTIVITY)

DATE

) FORM 1426

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE SUPPLY AGENCY
FENSE MEDICAL SUPPLY CENTER
RD AVENUE AND 29TH STREET
BROOKLYN, N.Y., 11232

OFFICIAL BUSINESS

POSTAGE FEES PAID
DEFENSE SUPPLY AGENCY

Defense Supply Agency
Defense Medical Supply Center
ATTN: Directorate of Technical Operations (Code T-M2)
3rd Avenue and 29th Street
Brooklyn, N.Y., 11232

FOLD

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:14Z
Check the source to verify that this is the current version before use.

DRAWING