MIL-M-36221A
4 May 1987
Superseding
MIL-P-36221(DM)
12 June 1964

MILITARY SPECIFICATION

MOUTHPIECE, SALIVA EJECTOR, WITH REFLECTOR BLADES, DENTAL

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a saliva ejector mouthpiece with reflector blades and an adjustable chin clamp, for use during dental operative procedures.

2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specifications and standards. The following specifications and standards form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of these documents shall be those listed in the issue of the Department of Defense Index of Specifications and Standards (DoDISS) and supplement thereto, cited in the solicitation.

SPECIFICATIONS

Federal

PPP-B-566 - Boxes, Folding, Paperboard.
PPP-B-636 - Boxes, Shipping, Fiberboard.
PPP-B-676 - Boxes, Set-up, Paperboard.

AMSC N/A FSC 6520

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 S. 20th St., Phila., Pa. 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

Military

MIL-P-116 Preservation, Methods of.

STANDARDS

Federal

FED-STD-151 - Metal; Test Methods.

Military

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.
MIL-STD-147 - Palletized Unit Loads.
MIL-STD-794 - Parts and Equipment, Procedures for Packaging of.

DRAWING

Defense Personnel Support Center (DPSC)

20733 - Mouthpiece, Saliva Ejector, Dental, With Reflector Blades.

(Copies of specifications, standards and drawing required by contractors in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting activity.)

2.2 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification shall take precedence. Nothing in this specification, however shall supersede applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

3.1 Material. The saliva ejector tube, shall be fabricated of brass suitable for the purpose intended and shall be chromium over nickel plated. The saliva ejector spring, fillister head spring screw, chin clamp assembly (clamp, plate and knurled head thumb screw) and the three reflector blades shall be fabricated of chromium-nickel austenitic corrosion resistant steel.

3.2 Style, design and dimensions. The saliva ejector mouthpiece shall consist of a saliva ejector tube, saliva ejector spring and fillister head screw, chin clamp assembly (clamp, plate and knurled head thumb screw), and three sizes of interchangeable reflector blades. The saliva ejector mouthpiece, when attached to the saliva ejector tube assembly of the dental operating unit, shall be suitab

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

for removing fluids from the mouth freely, without any obstructions. The reflector blades when assembled into the saliva ejector tube, with adjustable chin clamp, shall serve as a mouth mirror, light reflector, and tongue guard. The reflector blades shall be suitable for securely mounting in the intra-oral portion of the saliva ejector tube in a manner to permit ready interchange of reflector blades without the use of any tools. The saliva ejector mouthpiece and reflector blades, size 1, 2 and 3 shall conform to Defense Personnel Support Center(DPSC) Drawing 20733 for style, design, and dimensions.

3.2.1 Drawings. The drawing forming a part of this specification is not a production drawing. The drawing establishes dimensional limitations and controls design and configuration. The supplier shall be responsible for preparing his own shop drawings. The supplier shall select dimensions and tolerances within the limitations prescribed on the drawings to provide a product with proper design, configuration, fit and function. No deviations from the dimensional limitations prescribed on the drawing is permissible without prior approval of the contracting officer.

3.3 Finish. The mouthpiece components shall be smoothly finished. All edges shall be broken to prevent any injury to the soft tissues of the oral cavity. The finish of the reflector blades shall be of such quality and durability so as to provide a surface of optical quality to serve as a mirror during operative procedures. The saliva ejector mouthpiece set shall withstand repeated use without loss of finish.

3.4 Corrosion-resistance. Mouthpiece sets shall show no sign of corrosion when subjected to the boil test specified in 4.3.2.

3.5 Instructions. Instructions for use shall accompany each mouthpiece set.

3.6 Identification marking. The chin clamp of each mouthpiece set of domestic manufacture shall be marked with the name or registered trademark of the manufacturer. The chin clamp of each mouthpiece set of foreign manufacture shall be marked with the name of the country of origin, together with the name or registered trademark of the manufacturer. All markings shall be legible and permanent.

3.7 Workmanship. The mouthpiece shall be free from defects which may detract from its appearance or impair its serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

Except as otherwise specified in the contract or order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification, where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Certificates of quality. Certificates of quality, supplied by the manufacturer of the components or materials, may be furnished in lieu of actual performance of such testing by the contractor, provided that lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component or material, the date of testing, the test method, individual test results, the specification requirements, the lot number, the lot size and sample size.

4.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties). Sampling for inspection shall be performed in accordance with MIL-STD-105, except where otherwise indicated hereinafter.

4.2.1 Examination. Mouthpieces shall be examined for defects including but not limited to, those listed in table I.

TABLE I. Classification of defects.

Categories	Defects
Major:	
101	Style and design not as specified.
102	Any component missing.
103	Finish not as specified.
104	Dimensions of intra-oral portion of saliva ejector tube and reflector blades not as specified.
105	Reflector blades do not mount securely in intra-oral portion of saliva ejector tube.
106	Reflector blades not interchangeable without the use of tools.
107	Chin clamp not easily adjustable.
108	Saliva ejector tube not free of obstructions.
109	Finish of the reflector blades does not provide a surface of optical quality to serve as a mirror.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

TABLE I. Classification of defects. (Continued)

Categories	Defects
Minor:	
201	Mouthpiece not furnished clean.
202	Dimensions (other than 104) not as specified.
203	Identification markings on chin clamp not as specified.
204	Instructions absent, incomplete, illegible, or incorrect.

The inspection levels and acceptable quality level (AQL) shall be as listed in table II. The unit of product for examination shall be one mouthpiece.

TABLE II. Examination criteria.

Type Defect	Inspection level	AQL (Percent Defective)
Major	S-2	1.5
Minor	S-2	4.0

4.2.2 Testing. Mouthpieces shall be tested as specified in table III and 4.3. Inspection levels and acceptable quality level (AQL) shall be as listed in table III. The unit of product for testing shall be one mouthpiece.

TABLE III. Test criteria.

Characteristic	Component	Requirement	Test procedure	Insp. level	AQL (percent defective)
Boil test	Saliva ejector mouthpiece set	3.4	4.3.2	S-2	1.5

4.2.3 Inspection of packaging. The sampling and inspection of the preservation and interior package marking shall be in accordance with groups A and B quality conformance inspection requirements of MIL-P-116. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.3 Tests. Tests shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

4.3.1 Chemical analysis of the metals. In the absence of certificates of quality for the raw material, chemical analysis shall be conducted in accordance with method 111 or 112 of FED-STD-151. In case of dispute, method 111 shall be used for referee analysis.

4.3.2 Boil test. Before testing, the saliva ejector mouthpiece set shall be scrubbed with soap and water, rinsed thoroughly in distilled water, and dried. After cleaning, the mouthpiece set shall be boiled in distilled water in a glass beaker for 30 minutes and allowed to remain submerged for an additional 24 hours.

5. PACKAGING

5.1 Preservation. Preservation shall be level A or C, as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Cleaning and drying. Components comprising each set shall be cleaned and dried in accordance with MIL-P-116.

5.1.1.2 Preservative applications. Preservatives shall not be used.

5.1.1.3 Unit package. Each set shall be packaged in accordance with MIL-P-116, method IC-1.

5.1.1.4 Intermediate package. Twelve unit packages shall be packaged in a box conforming to PPP-B-566, PPP-B-636, Class domestic or PPP-B-676. Closure shall be as specified in the applicable box specification.

5.1.2 Level C. The level C preservation for mouthpiece set shall conform to the MIL-STD-794 requirements for this level.

5.2 Packing. Packing shall be level A, B or C, as specified (see 6.1).

5.2.1 Level A. Twelve intermediate packages (144 units) shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, overseas type, or PPP-B-621, class 2. Closure, and strapping shall be as specified in the applicable box specification.

5.2.1.1 Case Liner. Each level A wood box shal be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the liner specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

5.2.3 Level C. The level C packing for mouthpiece set shall conform to the MIL-STD-794 requirements for this level.

5.2.2 Level B. Twelve intermediate packages (144 units) shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, domestic type, PPP-B-621, class 1 or PPP-B-636, class domestic. Closure shall be as specified in the applicable box specification.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.2.5 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination equal 40 cubic feet or more. Loads shall be unitized on pallets conforming to MIL-STD-147. Quantities less than 40 cubic feet need not be unitized. Pallet loads including the pallet shall not exceed 43 inches in height, 48 inches in width, and 40 inches in length, and shall conform to Type IV or Type V.

5.2.5.1 Levels A and B. Mouthpiece set packed as specified in 5.2.1 and 5.2.2 shall be unitized on pallets conforming to MIL-STD-147.

5.2.5.2 Level C. Mouthpiece set packed as specified in 5.2.3 shall be unitized as specified in 5.2.5.

5.3 Marking.

5.3.1 Levels A, B, and C. Each unit pack, intermediate package, exterior container and unitized load shall be marked as specified in MIL-STD-129.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.2.3.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Ordering data.

6.1.1 Acquistion requirements. Acquisition documents should specify the following:

(a) Title, number and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 National Stock Number (NSN) coverage. This specification covers the following item appearing in the Federal Supply Catalog:

6.3 Subject term (key word) listing.

Mouthpiece
Saliva Ejector
Reflector Blades
Chin Clamps

6520-00-074-4947 MOUTHPIECE, SALIVA EJECTOR, DENTAL

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air-Force - 03

Review Activities:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

Project No. 6520-1864

Location: NCR DISK #1

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
MIL-M-3622/14	

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C.
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personal Support Center
Vernell Sanders (DSTC)
2800 So. 20th ST.
Phila, PA 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.