MIL-N-36220A
Amendment 1
10 October 1974

MILITARY SPECIFICATION

NEEDLE, HYPODERMIC, MALLEABLE, LEMMON, SPINAL PUNCTURE

This Amendment forms a part of Military Specification MIL-N-36220A, dated 25 May 1967, and is approved for use by all Departments and Agencies of the Department of Defense.

Page 1:

2.1 Delete the following under "SPECIFICATIONS, Federal", "PPP-T-76" and "M-A-250". Add the following under "SPECIFICATIONS, Federal": "PPP-B-566 Boxes, Folding, Paperboard, PPP-B-676 Boxes, Set-Up." Add the following under "STANDARDS, Military", "MIL-L-10547 Liners, Case, and Sheet Overwrap; Waterproof or Waterproof, Flexible."

Page 2:

2.2 Other publications. Delete entirely and substitute:

"2.2 Other publications. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

AMERICAN NATIONAL STANDARDS INSTITUTE (ANSI)
ANSI Z70.1 American Standard Dimensions of Glass and Metal Luer Tapers for Medical Applications.

(Application for copies should be addressed to the American National Standards Institute, 1430 Broadway, New York, New York 10018)."

Page 3:

3.2.2 Hub. Delete second sentence and substitute: "Hub and hub cavity shall be in accordance with Figure 3 of ANSI Standard Z70.1."

Page 4:

3.7 Workmanship. In line 8, delete "under 7", and substitute "under 10".

Delete 4.1, 4.1.1 and 4.1.2 and substitute:

"4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

"4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

"4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

"4.1.3 Certificates of quality. Certificates of quality supplied by the manufacturer of the nickel-silver alloy, brass alloy and corrosion-resistant steel alloy may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, the individual test results, and the specification requirements."

Page 5:

4.2.2, 4.2.3, and 4.2.4, delete entirely and substitute:

"4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table II. The unit of product for test purposes shall be one needle. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

Table II. Sampling Criteria for Tests

Component	Characteristic	Requirement	Test Procedure	Inspection Level	AQL
Needle Cannula and stylet wire	Boil test	3.1	GG-I-526	S-2	1.5
dle hub and t t kn b	Plating test	GG-I-526	QQ-N-290, or QQ-N-290 and QQ-C-320	S-2	1.5
uncture	Dimensions	3.2	---	S-2	2.5
	Diameter and depth of Female Luer Lock hub cavity	3.2	4.4.1	S-2	1.5

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:14Z
Check the source to verify that this is the current version before use.

Table II. Sampling Criteria for Tests (Con't)

Component	Characteristic	Requirement	Test Procedure	Inspection Level	AQL
Spinal puncture needle	Freedom from foreign matter	3.3	4.4.2	S-4	1.5
	Security of assembly	3.4	4.4.3	S-4	1.5
	Fluid tightness	3.5	4.4.4	S-4	1.5

"4.2.3 Raw materials. A sufficient quantity of each lot of each metal used in the construction of the needle shall be selected for chemical analysis."

Pages 6 and 7:

4.3 Examination. Delete first word of third sentence and substitute "Examination."
In Table III, for category 108, delete "7X" and substitute "10X". Delete categories 113, 118, 151, 152 and 202. Delete category 153 under Major E and substitute: "151 Cannula does not possess slot through wall at distal end."

Page 7:

4.4.1 Delete entirely and substitute:

"4.4.1 Diameter and depth of female Luer Lock hub cavity. The hub cavity of the female Luer lock connector shall be tested with the male Luer taper plug gage in accordance with Figure 5 of ANSI Standard Z70.1 (this gage shall be as shown in the ANSI Standard, except that it shall be constructed with a suitable fluid path to permit performance of the fluid tightness test specified in 4.4.4). The female Luer cavity shall be placed onto this gage with a straight motion, without turning or twisting the plug gage into the cavity. Once the female Luer cavity is placed onto the gage, no additional force shall be exerted to force the plug gage into the Luer cavity. The fit of the Luer cavity onto this gage shall be tested by means of the feeler gage in accordance with Figures 5 and 6 of ANSI Standard Z70.1."

4.4.3 Freedom from foreign matter test. In lines 2 and 5, delete "7X" and substitute "10X."

Page 8:

4.4 Fluid tightness test. Delete entirely and substitute:

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:14Z
Check the source to verify that this is the current version before use.

"4.4.4 Fluid tightness test. Samples of the needle shall be subjected to an internal water pressure of 45 pounds per square inch gage(psig). The plug gage specified in 4.4.1 shall be inserted into the female Luer Lock hub cavity with a straight motion, without turning or twisting the plug gage into the cavity with a force not to exceed 5 pounds. The feeler gage shall be placed onto the plug gage to assure that the "GO" position is not violated. The lumen of the cannula shall be sealed in a leakproof manner at the distal end. Water shall then be internally admitted to the testing arrangement, with the needle in a horizontal position, in a leakproof manner until the entire lumen and cavity space is filled (air bubbles eliminated) with internal pressure of 45 psig for 10 seconds. The entire assembly shall be noted for leakage.

5.1.3.4 Unit package. Delete method 1A-15 and subsitute "method 1C-2."

Page 9:

5.1.3.5 Intermediate package. Delete entirely and substitute:

"5.1.3.5 Intermediate package. Intermediate package shall be a box of appropriate size constructed in accordance with PPP-B-566. PPP-B-676 or PPP-B-636, class domestic, except that commercial colors are acceptable. Closure shall be adequate to prevent spilling of contents under normal handling."

5.2 Packing. Delete entirely and substitute:

"5.2 Packing.

"5.2.1 Level A.

"5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, V3c. Special requirements as specified in table II of PPP-B-636 shall apply for fiberboard boxes.

"5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

"5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with PPP-B-601 or PPP-B-636.

Source: https://assist.dla.mil -- Downloaded: 20260-21T08:14Z
Check the source to verify that this is the current version before use.