MILITARY SPECIFICATION

SYRINGES (AND NEEDLES), HYPODERMIC, DISPOSABLE

This amendment forms a part of Military Specification MIL-S-36157E, dated 19 November 1975, and is approved for use by all Departments and Agencies of the Department of Defense.

PAGE 1

1.2 Classification. Before "Size 1." insert "Size 1/2. - 1/2 cubic centimeter."

PAGE 4

3.2.1 Barrel. Line 5, after "0.060" insert "± .010".
Line 10, delete "female" and substitute "male".
Line 14, delete "0.155" and substitute "0.154".
Line 15, after "inch." insert "The maximum diameter at the distal end of the Luer tip shall be 0.160 inch."
Line 16, delete "0.170 to 0.175 inch" and substitute "0.169 inch reference."
Line 19, delete "extent" and substitute "extend".

PAGE 5

Table I. Barrel dimensions and graduations. Above Size 1, insert the following:

"1/2 (1/2cc)	2-1/2 to 3-1/2	0.15 to 0.28	0.020	0.015 to 0.080	0.01cc and 1U	0.05cc and 5U	0.05cc and 5U"

For Size 5: delete "29/32 ± 1/32" and substitute "0.856 to 0.938".

PAGE 6

3.2.1.1 Graduations and other markings. Line 12, after "readable." insert "On size 1/2 insulin syringes, the barrel shall bear the word "UNITS" adjacent to "50" at the proximal end of the scale."

Add, at end of paragraph: "On size 3 dental irrigating syringes, the graduations may go to 6cc."

3.2.2 Plunger. After line 16, insert "properly treated with the silicone lubricant specified in 3.1.4. An excess".

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:08Z
Check the source to verify that this is the current version before use.

Add the following new paragraph:

"3.2.2.1 Plunger tip resiliency. The rubber plunger tip when tested as specified in 4.4.2.4 shall not crack, crumble or become fissured. Furthermore, after a one-minute recovery time, it shall again be circular in appearance. An obviously oval shape indicating a lack of adequate recovery is cause for rejection."

PAGE 7

3.2.3 Needle (type II syringe only). Add, at end of paragraph: "Also, for permanently attached needles, the deviation of the cannula from the central longitudinal axis may be up to 6 degrees 15 minutes."

3.3 Accuracy. Add, at end of paragraph: "For size 1/2 syringe, at 0.2cc or more, accuracy shall be ± 5%; and at less than 0.2cc, accuracy shall be within 0.007cc. The dead space in syringes graduated in units shall be no greater than one unit."

PAGE 8

Add the following new paragraph:

"3.4.4 Plunger fit. The plunger shall make a snug fit in the barrel to provide leak-tight operation. At no part of the travel of the plunger shall the plunger fit be so loose as to permit movement of the plunger by its own weight when tested as specified in 4.4.2.5."

PAGE 10

Table II. Sampling for examination. Delete footnote 1/.

PAGE 11

Table III. Sampling criteria for tests. After "Dimensions" add line to read as follows: "Component: Plunger tip; Characteristic: Resiliency; Requirement: 3.2.2.1; Test procedure: 4.4.2.4; Inspection level: S-3; AQL: 0.25."

For Characteristic "Pressure test" delete "2.5" and substitute "1.0".

For Characteristic "Suction test" delete "S-3" and "1.5" and substitute "S-2" and "1.0" respectively.

After "Suction test" add line to read as follows: "Component: Syringe; Characteristic: Plunger fit; Requirement: 3.4.4; Test procedure: 4.4.2.5; Inspection level: S-3; AQL: 0.25."

Delete footnote 1/.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:08Z
Check the source to verify that this is the current version before use.

PAGE 13

4.4.1 Accuracy test. Last sentence, after "syringes" insert ", except for syringes graduated in units,".

4.4.2.1 Pressure test. Line 1, after "with" insert "removable".
Line 2, delete colon and substitute period, following which insert: "For syringes supplied without needles, the syringe shall be fitted to a standard ANSI Z70.1 Luer hub, except that it shall be drilled to allow passage of fluid and shall be connected to the pressure test system so that the pressure within the syringe can be determined."
Line 5, delete "3-1/2" and substitute "5".
Line 8, after "locks." insert "For syringes with Luer locks, a needle with a hub complying with ANSI Z70.1 shall be screwed into the Luer lock and applied to the Luer tip by means of the needle protector."
Line 8, after "needle" insert ", if present,".

PAGE 14

Add the following new paragraphs:

"4.4.2.4 Plunger tip resiliency. The unmounted rubber plunger tip shall be free of visible lubricant. If necessary wipe dry with a cloth. The outside diameter shall be measured to the nearest 1/16 inch. The plunger tip shall then be compressed laterally to 7/16 to 1/2 of the outside diameter for 15 seconds. This shall be done in a vise the jaws of which have smooth faces. The compressed plunger tip shall be observed for cracks, crumbling or fissures. The plunger tip is removed from the vise and allowed to recover for 60 seconds after which it is examined for circularity."

"4.4.2.5 Plunger fit. The syringe shall be held vertically with the Luer tip exposed. The plunger shall be worked back and forth twice, for the full length of the barrel. The plunger shall be withdrawn from zero, in increments of 1/10 the capacity, and observed for 10 seconds at each level. The Luer tip can be up or down in this test. Movement of the plunger by its own weight is cause for rejection. The same test shall be used for syringes with permanently attached needles except that 20 seconds observation time shall be used."

PAGE 18

"4.4.9 Examination for preparation for delivery. Line 1, after "made" insert "of intermediate and shipping containers".

"5.1.1 Unit of issue. Line 1, after "containing" insert "thirty,".

Table VI. Packaging quantities. Add line 6 to read as follows: "Type: II, Size: 1/2; Unit package: 1 unit; Intermediate package: 3 units; Exterior container: 24 units".

Add line 7 to read as follows: "Type: II, Size: 3; Unit package: 1 unit; Intermediate package: Not required; Exterior container: 5 units".

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:08Z
Check the source to verify that this is the current version before use.

PAGE 19

5.1.3.2.1 Type I or II, Size 1 thru 6. In paragraph heading, delete "Size 1 thru 6" and substitute "Size 1/2 thru 6".

PAGE 20

5.3.2 Unit package. Delete the seventh sentence and caution legend.

PAGE 22

6.4 Add the following at end of paragraph:

"6515-01-032-1458	SYRINGE AND NEEDLE, HYPODERMIC, Disposable, 100U Insulin, 26 or 27 Gage, 1/2 Inch Needle, 1/2 ml, 30s.	II	1/2
6520-01-036-0723	SYRINGE, IRRIGATING, DENTAL, Disposable, with 23 Gage x 1-1/4 Inch Cannula, 6 ml, 50s."	II	3

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
DLA-DM

Project No. 6515-2123

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:08Z
Check the source to verify that this is the current version before use.