MIL-S-36197(DM)
5 May 1964

MILITARY SPECIFICATION

SPLINT SUPPORT AND FOOTREST, LEG

1. SCOPE

1.1 This specification covers a lightweight splint support and footrest which shall provide proper support to the foot of a patient when the foot is encased in footwear.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

QQ-A-200/8	Aluminum Alloy Bar, Rod, Shapes, and Tube, Extruded, 6061 and 6062.
QQ-A-250/8	Aluminum Alloy, 5052, Plate and Sheet.
TT-E-527	Enamel, Alkyd, Lustreless.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-T-76	Tape, Pressure-Sensitive, Adhesive Paper, (for Carton Sealing).

Military

MIL-C-5541	Chemical Films for Aluminum and Aluminum Alloys.
MIL-P-8585	Primer Coating, Zinc Chromate, Low-Moisture-Sensitivity.
MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Federal

Fed. Test Method Std. No. 151	Metals; Test Methods.
Fed. Std. No. 595	Colors.

: FSC 6515 :

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:12Z
Check the source to verify that this is the current version before use.

DRAWING

Defense Medical Supply Center

20412 Splint Support and Footrest, Leg.

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material.

3.1.1 Support and coupling. Aluminum alloy shall be 6061-T4 conforming to Specification QQ-A-200/8.

3.1.2 Clamp. Aluminum alloy shall be 5052-H32 conforming to Specification QQ-A-250/8.

3.2 Construction. The splint support and footrest shall be constructed as shown in the drawing.

3.3 Welding. All welds shall be continuously even, smooth, and free from flaws. Welds shall be ground smooth to eliminate traces of weld marks and blended into adjacent surfaces.

3.4 Finish. The splint support and footrest shall be protected against corrosion in accordance with the requirements of Specification MIL-C-5541. All outside and accessible inside surfaces shall receive a zinc-chromate primer, conforming to color Y of Specification MIL-P-8585, which shall be followed by two coats of lustreless olive drab enamel conforming to Specification TT-E-527 and to color X34087 of Federal Standard No. 595 (see 6.2). All enamel coats shall be sprayed on to accomplish a complete and uniform coverage of the coated parts, and shall be baked to give a smooth, hard, and tough finish, which shall be free from runs and which shall resist cracking and chipping.

3.5 Identification marking. Each splint support and footrest shall be stamped with the name or registered trademark of the manufacturer.

3.6 Workmanship. Workmanship shall be first class throughout. There shall be no rough edges. The splint support and footrest shall be free from defects which detract from its appearance or may impair its serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination.. Sampling for examination shall be conducted in accordance with Standard MIL-STD-105.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with Standard MIL-STD-105.

	Inspection level	AQL (percent defective)	Minimum sample size
Component material testing	S-1	*	3

*Acceptance number shall be zero (0) for all sample sizes.

4.2.3 For chemical analysis of metals. Four ounces of each lot of each metal used in the construction of the splint support and footrest shall be selected for chemical analysis as specified in 4.4.1.

4.2.4 Component materials. Sufficient samples of the component materials shall be selected from each lot of material used in the manufacture of the splint support and footrest to perform tests as required by the applicable subsidiary specifications and by tests specified herein. When sample size for component materials covered by subsidiary specifications is not specified, the sample size for each test shall be as indicated in 4.2.2. The unit of product for testing shall be the quantity of material necessary to perform all required tests one time each, taking into consideration the fact that, where possible, the same material may be used to perform two or more tests.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:12Z
Check the source to verify that this is the current version before use.

4.3 Examination. The splint support and footrest shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Chemical analysis of the metals. The samples selected in accordance with 4.2.3 shall be analyzed in conformance with methods 111 or 112 of Federal Test Method Standard No. 151. In case of dispute, method 111 shall be used for reference analysis.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One (1) splint support and footrest, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table I shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in procurement document.

Table I. - Packaging quantities

Unit package	Intermediate package	Exterior container
1 unit	12 units	48 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit shall be wrapped in kraft paper of suitable weight. Wrapping shall be adequately secured to prevent accidental opening.

5.1.3.2 Intermediate package. Intermediate package shall be a carton of appropriate size constructed in accordance with Specification PPP-B-636. Each carton shall be adequately secured to prevent accidental opening.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with Specification PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type I, class 2.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to Specification MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix of Specification PPP-B-636. In addition, all seams, including manufacturer's joint, shall be closed by tape conforming to the requirements of Specification PPP-T-76. The tape shall be not less than 2 inches wide. Three strips shall be applied to both top and bottom so that all seams and joints shall be covered their full length. Tape applied over the lengthwise seams shall extend a minimum of 2 inches onto each of the end panels. Tape applied over the end edge seams shall extend a minimum of 2 inches along the side and top, and side and bottom panels.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to Specification PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load, and constructed in accordance with Specification PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type I, class 1.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with Standard MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

5.3.2 Exterior container. Exterior container shall be marked in accordance with Standard MIL-STD-129. In addition, the letter "L" shall precede the Federal stock number. One space shall be allowed between the letter "L" and the first digit of the stock number. The terms "Stock number" or "FSN" shall not be indicated.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packing (see 5.2).

6.2 Availability of color standards. The color standards of the color referenced in 3.4 are available from:

Commanding General
Aberdeen Proving Ground, Maryland
ATTN: Coating and Chemical Laboratory

6.3 This specification does not include all types, classes, grades, si , etc., of the commodity indicated by the title of this specification, or whic. are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

Custodians:
Army. - MD
Navy. - BuMed
Air Force. - AFMSC

Review activities:
Army. - MD
Navy. - BuMed
Air Force. - AFMSC

Preparing activity:
Defense Supply Agency. - DM

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:12Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

FORM APPROVED BUDGET BUREAU NO. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use of the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity.

SPECIFICATION	MIL SPEC MIL-S-36197(DM), Dated 5 May 1964 SPLINT SUPPORT AND FOOTREST, LEG
ORGANIZATION	CITY AND STATE
CONTRACT NO.	QUANTITY OF ITEMS PROCURED / DOLLAR AMOUNT $

MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?
   ☐ YES ☐ NO IF "YES" IN WHAT WAY?

4. REMARKS (ATTACH ANY PERTINENT DATA WHICH MAY BE OF USE IN IMPROVING THIS SPECIFICATION. IF THERE ARE ADDITIONAL PAPERS, ATTACH TO FORM AND PLACE BOTH IN AN ENVELOPE ADDRESSED TO PREPARING ACTIVITY)

SUBMITTED BY (PRINTED OR TYPED NAME AND ACTIVITY)	DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE SUPPLY AGENCY
DEFENSE MEDICAL SUPPLY CENTER
3RD AVENUE AND 29TH STREET
BROOKLYN, N.Y., 11232

OFFICIAL BUSINESS

POSTAGE FEES P
DEFENSE SUPPLY A

Defense Supply Agency
Defense Medical Supply Center
ATTN: Directorate of Technical Operations (Code T-M2)
3rd Avenue and 29th Street
Brooklyn, N.Y., 11232

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.