MIL-T-36192A
22 July 1970
SUPERSEDING
MIL-T-36192(DM)
15 April 1964

MILITARY SPECIFICATION

TUBE, CONNECTING, SYRINGE AND NEEDLE, DISPOSABLE (WITH TWO INJECTION SITES)

This specification is mandatory for use by all the Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers a sterile, disposable, plastic, intravenous extension set, with two injection sites, suitable for use in infusions and transfusions, and for syringe or drip administration of sodium pentothal.

2. APPLICABLE DOCUMENTS

* 2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

MMM-A-250	Adhesive, Water-Resistant, (for Sealing Fiberboard Boxes).
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Set-up.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, Water-Resistant (for Carton Sealing).

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

* 2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply:

AMERICAN NATIONAL STANDARDS INSTITUTE, INCORPORATED

ANSI Z70.1 American National Standard
Dimensions of Glass and Metal Luer Tapers For
Medical Applications.

(Application for copies should be addressed to the American National Standards Institute, 1430 Broadway, New York, N.Y. 10018).

UNITED STATES PHARMACOPEIAL CONVENTION, INCORPORATED

Pharmacopeia of the United States (U.S.P.)

(Application for copies should be addressed to the Mack Publishing Company, Easton, Pa. 18042).

3. REQUIREMENTS.

3.1 Material.

3.1.1 Rubber and plastic components. All rubber and plastic components shall be of high quality and suitable in all respects for the intended purpose. All rubber and plastic components shall be sterilizable, free from pyrogens, compatible, inert, and safe (nontoxic) when tested as specified in section 4. Those rubber and plastic components coming in contact with solutions being administered, and end protectors, shall contain not more than 5 parts per million (p.p.m.) of lead when tested as specified in 4.4.6.2.

3.1.1.1 Rubber. Rubber parts shall be fabricated from latex.

3.1.1.2 Plastic tubing. Plastic tubing shall be clear and transparent. The tubing shall be flexible and pliable, but sufficiently rigid to prevent collapse during normal use. The tubing shall be free from all liquid lubricating substances which oxidize, exude, or sweat-out to the surface of the tubing.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

3.1.1.3 Other material. All other materials hereinafter specified shall be of high quality and suitable in all respects for the purpose intended.

* 3.2 Style, design and construction. The unit shall be a disposable, plastic, intravenous extension set suitable for use in infusions and transfusions, and for syringe and drip administration of sodium pentothal. The sets shall be sterile, pyrogen free, safe (nontoxic), and inert (see 3.8, 3.9 and 3.10). The set shall consist of plastic tubing fitted with a female Luer adapter at one end, a male Luer adapter at the opposite end, two injection sites, located from 2 to 4-1/2 inches from each end adapter, a screw clamp positioned between the female adapter and nearest injection site, and another screw clamp between the two injection sites. The set shall measure 30 plus or minus 3 inches in overall length, and the tubing shall have an outside diameter of 1/8 to 3/16 inch. Injection sites shall be fabricated in such a manner that injection is not through the main tubing but will be accomplished by use of a Branch "T" or "Y" with a puncturable, self-sealing rubber diaphragm on the branch (Branch "Y"'s shall be assembled with the rubber diaphragm extending towards the female adapter). The injection sites shall be sealed but shall allow for the insertion of a hypodermic needle. The assembly of the set shall be structurally sound and shall not leak air or fluid during use. Open ends of the set shall be tightly covered with readily removable plastic covers, or end protectors, which shall maintain the sterility and freedom from pyrogens of the fluid path after sterilization. At the contractor's option, a rubber molded part shall be firmly secured between the male Luer adapter and the plastic tubing. The molded rubber part shall be constructed with an enlarged center section having a shoulder to permit addition of medication by easy insertion of a hypodermic needle. The molded rubber part shall be self-sealing when the needle is withdrawn.

3.2.1 Injection sites. The injection site shall consist of a rigid, transparent "Y" or "T" branch, with a puncturable, self-sealing, rubber diaphragm. Two ends of the branch shall be securely attached to the plastic tubing, and the rubber diaphragm shall be firmly secured to the third branch so as to be readily accessible for making injections while the set is in use.

3.2.2 Adapters. The adapters shall be fabricated of a suitable plastic, and so constructed, to be attached at one end to the plastic tubing (or rubber molded part) in a leakproof manner.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

* 3.2.2.1 Male adapter. The adapter shall be provided with a standard Luer taper in accordance with American Standards Association (ANSI) Standard Z70.1, and shall make a leakproof connection with a standard Luer hub cavity constructed in accordance with figure 3 of ANSI Standard Z70.1 for a hypodermic needle. This tapered end of the adapter shall be tested for dimensions as specified in 4.4.1. When so tested the adapters shall fit the test gage within the "GO" - "NO GO" limits of the gage as shown in figure 7 of ANSI Standard Z70.1. Adapters with Luer tapered ends which do not fit the test gage within the specified limits shall be considered not acceptable. In addition, when the test gage is fitted onto the Luer tapered end, as specified in 4.4.1, there shall be no movement of the gage on the tapered end, when this gage is pressed lightly from side to side in all directions.

* 3.2.2.2 Female adapter. The female adapter shall make a secure water-tight union with male Luer slip connectors. The female adapter shall be tested for depth and diameter as specified in 4.4.2. The adapter shall fit the plug gage within the "GO" - "NO GO" limits of the feeler gage as shown on figures 5 and 6 of ANSI Standard. Adapters which do not fit the plug gage within these limits shall be considered not acceptable. In addition, when the adapter is fitted onto the plug gage as specified, there shall be no movement of the adapter when the adapter is pressed lightly from side to side in all directions.

* 3.2.3 Screw clamp. Each apparatus shall be provided with two suitable screw clamps, assembled on the set as specified in 3.2, which, by compression of the tubing, will regulate flow of fluids through the tubing, and which will also stop the flow of fluids (i.e., complete shutoff).

3.3 Compatibility. The materials of which the sets are manufactured shall not react with, or be acted upon by, the liquids with which they will be in contact during normal use and operation, nor shall there be any materials, incorporated in the set which will cause any harmful or deleterious effect when the set is used for the purpose intended.

3.4 Leakage and separation of set. The sets shall be tested for leakage and separation of set as specified in 4.4.3. Sets which show leakage or separation at any part of the set shall be rejected.

3.5 Clear passage. Sets shall be tested for clear passage as specified in 4.4.4. Sets which are sealed or constricted shall be rejected.

3.6 Security of assembly. The adapters or branches shall not separate or loosen from the plastic tubing (or rubber molded part) when tested as specified in 4.4.5.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

3.7 Manufacturer's record. The manufacturer shall place a code or lot number on the paper label, or on the printed carton, which can be verified against the manufacturer's records of all tests made on nontoxic, sterile, pyrogen-free sets. These records shall be made available for inspection by any duly authorized Government representative. No shipment shall be made until all tests have been completed, verified, and recorded in the permanent records of the manufacturer.

* 3.8 Safety (non-toxicity) and lead content.

* 3.8.1 Safety (non-toxicity). When tested as specified in 4.4.6.1, the sets shall comply with safety test requirements of the Pharmacopeia of the United States (U.S.P.) for Transfusion and Infusion Assemblies.

3.8.2 Lead content. When tested as specified in 4.4.6.2, all plastic and rubber components coming in contact with the medication being administered, and end protectors, shall contain not more than 5 parts per million (p.p.m.) of lead.

3.9 Freedom from pyrogens. When tested as specified in 4.4.7, sets shall comply with the pyrogen test requirements of the U.S.P., for Transfusion and Infusion Assemblies.

3.10 Sterility. Sets shall comply with the sterility test requirements of the U.S.P., when tested as specified in 4.4.8.

* 3.11 Certificates. For destination inspection only, certificates of sterility, freedom from pyrogens, and safety, as specified below, shall be submitted by the contractor to Directorate of Medical Materiel, Defense Personnel Support Center, Attn: ATQ, 2800 S. 20th Street, Philadelphia, Pa. 19101, and one copy of each shall be forwarded with each shipment to the consignee.

3.11.1 Certificates of safety (toxicity). A certificate of safety, listing all lot or control numbers, shall state that samples from each lot or control have been tested for safety as specified in 3.8.1 and 4.4.6, and were found to comply with such requirements. The certificate shall reference the method used in determining safety.

3.11.2 Certificates of freedom from pyrogens. A certificate of freedom from pyrogens, listing all lot or control numbers, shall state that samples from each lot or control have been tested for pyrogens as specified in 3.9 and 4.4.7, and were found to comply with such requirements. The certificate shall reference the method used in determining freedom from pyrogens.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

3.11.3 Certificate of sterility. A certificate of sterility, listing all sterility lot or control numbers, shall state that samples from each sterility lot have been tested for sterility, as required in 3.10 and 4.4.8, and have been found to be sterile. The certificate shall reference the method used in determining sterility.

3.12 Instructions. Instructions for proper use of the sets shall be provided with each unit package.

3.13 Workmanship. Workmanship shall be first class throughout. The sets shall be free from defects which detract from their appearance or impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.2 Sampling.

* 4.2.1 For examination. Sampling for examination shall be conducted in accordance with Standard MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in Table I. The unit of product for examination purposes shall be one syringe and needle connecting tube complete with injection sites.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

Table I. Sampling for examination

Type of Defect	Inspection level	AQL (Percent defective*
Major	I	1.0
Minor	I	2.0

* If acceptance number is not exceeded, lot shall be acceptable. If rejection number is equaled or exceeded, inspection at level II may be performed to determine final acceptance or rejection.

* 4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in Table II.

Table II. Sampling for Tests

Component	Characteristic	Requirement	Test procedure	Insp. level	AQL (% defective
Male adapter	Fit	3.2.2.1	4.4.1	S-2	*
Female adapter	Fit	3.2.2.2	4.4.2	S-2	*
Screw clamp	Control	3.2.3	4.4.9	S-1	*
Set	Leakage & separation of set	3.4	4.4.3	S-2	*
Set	Flow	3.5	4.4.4	S-2	*
Set	Security of assembly	3.6	4.4.5	S-2	*
Set	Safety (non-toxicity)	3.8.1	4.4.6.1	S-1	*
Set	Lead	3.8.2	4.4.6.2	S-1	*
Set	Freedom from pyrogens	3.9	4.4.7	S-2	*
Set	Sterility	3.10	4.4.8	S-3	*

* Acceptance number shall be zero for all sample sizes.

4.2.3 Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

* 4.3 Examination. The sets shall be examined to determine compliance with all requirements contained in this specification. Examination shall be conducted in accordance with MIL-STD-105 and the following classification of defects. Examination shall include but shall not be limited to the defects listed in table III.

Table III. Classification of Defects

Categories	Defects
Major	
101	Luer adapters not uniform in tapers.
102	Luer adapters not free of burrs or other defects which would affect fluid tightness.
103	Cracks in fittings.
104	Flow regulator cannot be operated with one hand because of tightness of components.
105	Flow regulator is not effective because it is of improper size or design and allows tubing to slip to side of screw.
106	Set not free of foreign matter.
Minor	
201	Tubing not free of kinks.
202	Fittings not free of burrs on interior.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

* 4.4.1 Taper of male Luer adapter. The male Luer adapter shall be tested with the ring gage shown on figure 7 of ANSI Standard. This gage shall be slipped onto the tip with very light pressure. Tips that do not fit the gage within these limits shall be considered not acceptable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

* 4.4.2 Diameter and depth of cavity of female Luer adapter. The cavity of the female Luer adapter shall be tested on the male Luer taper gage in accordance with plug gage as shown on figure 5 of ANSI Standard (except that it may be constructed with a suitable fluid path). The adapter shall be slipped onto this gage with very light pressure. The fit of the adapter on this gage shall be tested by means of the feeler gage as shown on figure 5 of ANSI Standard.

* 4.4.3 Leakage and separation of set. Samples of finished sets, after sterilization, shall be subjected to an internal hydraulic pressure of 45 pounds per square inch gage (p.s.i.g.). This test shall be performed with a gage as shown on figure 8 of ANSI Standard Z70.1 (and having a suitable fluid path throughout the length of the gage) fitted to the Luer-tapered-tip of the male adapter and gage as shown on figure 5 of ANSI Standard Z70.1 (and having a suitable fluid path throughout the length of the gage) fitted to the female adapter. This pressure shall be maintained for 30 minutes. The set shall be observed for leakage.

4.4.4 Clear passage test. Samples, after sterilization, shall be subjected to an internal water pressure of 3 pounds p.s.i. and when subjected to this pressure, shall be examined for free passage of the testing fluid.

* 4.4.5 Security of assembly test. The tubing of the set shall be severed at the midpoint of the distance between the two branches. The security of assembly of each adapter and branch shall then be tested by gradually applying a four pound weight to the severed end of the tubing for 15 minutes while the adapter is held in a clamping device so that the tubing hangs vertically downward.

* 4.4.6 Safety (non-toxicity) and lead content tests.

* 4.4.6.1 Safety (non-toxicity). The finished sets shall be tested for safety in accordance with the U.S.P., for Infusion and Transfusion Assemblies.

4.4.6.2 Test for lead. Ten (10) Gms. of rubber or plastic shall be taken, ignited to char, and then further ignited in a muffler at 700° C. to a white ash. The ash shall be boiled in 15 milliliters (ml.) of 10 percent nitric acid, filtered into a volumetric flask and made up to 100 ml. An aliquot portion, equivalent to one gram of sample, shall be taken and placed in a separator. The dithizone test for lead, as given in the U.S.P., shall be applied using a control of 5 p.p.m. or 12 p.p.m. of lead solution as given in the U.S.P. test. If the sample solution is darker than the limit, smaller aliquots shall be used until the color of the control is matched by the sample solution and the concentration of lead in p.p.m. is determined.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

4.4.7 Freedom from pyrogens test. The pyrogen test shall be performed as described in the U.S.P., for Transfusion and Infusion Assemblies.

* 4.4.8 Sterility test. The sterility test shall be performed as described in the U.S.P. under Procedural Details for Sterility Tests of Class B, Type II, Inoculated Products, and for Parenteral Devices. As an alternate, testing for sterility shall be in accordance with the method specified in the U.S.P. Procedural Details for Sterility Tests, Class B, Type II, Inoculated Carrier, and in Parenteral Devices, except that the inoculated carrier shall contain sterilization test strips carrying spores of Bacillus subtilis var. globigii and Bacillus stearothermophilus, Type B of MIL-S-36586.

* 4.4.9 Screw clamp. The sample shall be connected to a regular administration set with 54 inch tubing set up with water from a bottle set at 5 feet. The effectiveness of the screw clamp shall be observed by varying flow rates from fully opened to fully closed.

* 5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.1).

5.1.1 Unit of issue. One package (20 tubes), as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	12 units	24 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

5.1.3 Level A.

5.1.3.1 Immediate container. Each tube shall be coiled in such a manner as to prevent kinking and packaged in suitable, waterproof immediate container which shall retain the sterility until manually opened. As an alternate, the immediate container may be supplied in a transparent, waterproof overwrap.

5.1.3.2 Unit package. Each unit (20 immediate containers) shall be packaged in a box of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676, except that commercial colors will be acceptable, or PPP-B-636, type CF, class domestic. Closure shall be adequate to prevent accidental opening.

5.1.3.3 Intermediate package. Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class weather-resistant. Boxes shall be closed by taping the top and bottom flaps with 2 inch wide tape conforming to PPP-T-76. Closing of bottom flaps with adhesive conforming to MMM-A-250 is permitted. Metal stitches are not permitted.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be Level A, B, or C as specified (see 6.2).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Immediate container. In addition to any commercial markings, each immediate container shall be marked with item identification, the word "STERILE", sterility lot control number, requirements of the Federal Food, Drug and Cosmetics Act, and the legend:

"FLUID PATH AND AREAS UNDER PROTECTIVE COVERINGS OF SET ARE STERILE AND NON-PYROGENIC UNLESS PACKAGE IS DAMAGED OR OPENED" or similar statement.

5.3.2 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. In addition, marking shall include contract or purchase order number, name of contractor, sterility lot control number and the method of sterilization (G-Gas, R-Radiation, S-Steam, D-Dry Heat Sterilization) shall be shown following the level of packaging and date of packaging. Example "A-2/70-G". Final sterilization shall be on or after indicated packaging date.

5.3.3 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. Sterility lot (control) number shall be shown. In addition, the method of sterilization marking as specified in 5.3.2 shall be shown.

NOTE: Each lot number shall be traceable to a specific sterilizer charge.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.

(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

* 6.2 This specification is applicable to the following Federal Supply Catalog item:

6515-959-8116 TUBE, CONNECTING, SYRINGE and NEEDLE

6.3 Supersession data. This specification includes requirements of Purchase Description No. 5, dated 14 January 1969.

6.4 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and Contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the previous issue.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1065

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
MAR 68

EDITION AUG 66, WILL BE USED UNTIL EXHAUSTED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or serve to amend contractual requirements.

SPECIFICATION

MIL-T-36192A FSN:

ORGANIZATION

CITY AND STATE | CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)* | DATE

DD FORM 1 JAN 66 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.