MIL-T-36202B
18 August 1977
SUPERSEDING
MIL-T-36202A
9 December 1969

MILITARY SPECIFICATION

TABLE, SURGICAL INSTRUMENT AND SURGICAL DRESSING

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers corrosion-resisting steel, surgical instrument and surgical dressing tables, suitable for use in military medical facilities.

1.2 Classification. The surgical instrument and surgical dressing tables shall be of the following types, as specified (see 6.2):

Type I - 20 by 16 by 34-1/2 inches with one drawer.
Type II - 48 by 24 by 40-1/2 inches with two drawers.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

FF-C-77 - Casters, Rigid and Swivel (Institutional Duty).
PPP-B-636 - Boxes, Shipping, Fiberboard.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, Code ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

MILITARY

MIL-P-116 - Preservation-packaging, Methods Of.

STANDARDS

FEDERAL

FED-STD-151 - Metals, Test Methods

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

20532 - Table, Surgical Instrument and Surgical Dressing, Corrosion Resisting Steel.

(Copies of specifications, standards, drawings, and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

AMERICAN IRON AND STEEL INSTITUTE (AISI)

AISI Steel Products Manual - Stainless and Heat Resistant Steels.

(Application for copies should be addressed to the American Iron and Steel Institute, 1000 16th St N.W., Washington, DC 20036).

NATIONAL FIRE PROTECTION ASSOCIATION (NFPA)

No. 56A Standard for the Use of Inhalation Anesthetics

(Application for copies should be addressed to the National Fire Protection Association, 470 Atlantic Avenue, Boston, MA 02210.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 First article. When specified in the procurement document, the contractor shall furnish sample unit(s) for first article inspection and approval (see 4.1.2.1 and 6.2).

3.2 Materials.

3.2.1 Corrosion-resisting steel. Corrosion-resisting steel shall comply with the AISI Type 300 series as indicated in the AISI steel products manual when tested as specified in 4.4.1.

3.3 Construction. Construction shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 20532. Unless otherwise specified the material used in the manufacture of the table shall be corrosion-resisting steel. The table shall consist of four tubular uprights or legs fitted with casters, a top working surface with guardrail, a drawer, or drawers, located directly beneath the top and enclosed in a housing, and a lower shelf. The table shall be of the straight-line, sanitary, flush construction, all parts closely fitted, welded throughout, durable and sound in every detail. The entire structure shall be properly reinforced and designed to minimize vibration when in motion. Gages specified herein are minimum; unless otherwise specified, commercial tolerances shall be allowed. The drawer housing flanges need not be hemmed providing they are machined such that no sharp edges are present.

3.3.1 Top. The top shall be constructed of corrosion-resisting steel with all edges flanged as shown on the drawing. The corners of the top shall be rounded to fit the contour of the housing. The top for the type II table shall be reinforced with longitudinal or lateral channels, as shown in DPSC Drawing No. 20532. The top shall be fitted tightly over the drawer housing and shall be held in position by the guardrail screws.

3.3.2 Drawer housing. On the type I table, the top edge of the fascia strip shall be flanged inward 1-1/4 ± 1/4 inches and the bottom edge shall be flanged inward a minimum of 3/4 inch. On the type II table, the top, bottom, and side edges of each drawer opening shall be flanged inward 1-1/4 ± 1/4 inches.

3.3.3 Drawer. The drawer front shall be filled with sound-deadening material and shall close against rubber bumpers. The drawer shall be provided with a pull handle made of corrosion-resisting steel, or anodized aluminum. When the drawer is closed there shall not be more than 1/16 inch clearance at any point between drawer front and enclosing frame, no more than 1/16 inch vertical movement of the drawer, and no more than 1/16 inch lateral movement of the drawer. The drawer face shall close flush with the front of the drawer housing. The clearance between the sides of the drawer front and the enclosing frame shall not differ by more than 1/32 inch. Also, on type II tables, the clearance between the top and the bottom of the drawer front and the enclosing frame shall not differ by more than 1/32 inch.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

3.3.4 Drawer slides. There shall be two pairs of corrosion-resisting steel slides, formed as shown on the drawing, associated with each drawer. One pair of slides, the housing slides, shall be securely welded to the back and to the front turned-in flanges of the housing The other pair of slides shall be securely welded to the side of the drawer body. At the front of each housing slide and the back of each drawer body slide, there shall be a ball bearing roller with a nylon facing. The drawer slides shall include a drawer stop to prevent inadvertent removal of the drawer. The drawer slides shall function to enable opening and closing of the drawer without binding.

3.3.5 Guardrail. The guardrail shall be formed from corrosion-resisting steel rod and shall be continuous around the back and sides of the tabletop. The rail shall be supported by four vertical legs, two of which shall be formed integrally with the rail by bending the ends, the other two being separate pieces of rod continuously welded to the guardrail at appropriate locations. The guardrail shall be removable and shall be secured to the tabletop flange and the drawer housing by corrosion-resisting steel bolts or screws, nuts and lock washers. The screw or bolt head shall be located on the outside of the table. Each guardrail shall be mounted with either all screws or all bolts.

3.3.6 Casters Each upright of the table shall be fitted with a caster that is electrically conductive. The table, with the casters in place, shall meet the requirements for Reduction of Electrostatic Hazards - Furniture as specified in NFPA No 56A, when tested as specified in 4.4.2. Casters shall be zinc or cadmium plated. Each table shall have four casters of the same size.

3.3.6.1 Type I. The four casters for the type I table shall conform to type II, Class A, 3 inch, style b, mounting A or T of FF-C-77.

3.3.6.2 Type II. The four casters for the type II table shall conform to the following requirements of FF-C-77:

a. Two casters - Type I, class A or B, 3 inch, style a, mounted in a solid stem, fitted to the inside of the tubular upright and bolted in place.
b Two casters - Type II, class A or B, 3 inch, style a, mounting A or T, with brakes.

3.4 Finish. All normally visible surfaces of the table, including interior as well as exterior surfaces of drawer, shall be given a No. 4 commercial finish. All other surfaces shall be cleaned and finished in accordance with best commercial practice.

3.5 Identification marking. Each unit shall be marked permanently and legibly with the following:

Contract or purchase order number.
Name or registered trademark of the manufacturer.
National Stock Number.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

3.6 Service data. The contractor shall furnish, with each unit, two copies each of a technical manual which contains complete instructions for installation, operation, maintenance, cleaning and lists of components parts.

3.7 Workmanship. Workmanship shall be first class throughout. Tables shall be free from defects which detract from their appearance or may impair their serviceability. All burrs and sharp edges shall be removed.

3.7.1 Welding. Welding shall be durable, sound in every detail and in accordance with the latest practices recommended by the American Welding Society. All welds shall be produced by qualified welding operators with satisfactory welding equipment using satisfactory filler metal or electrode where applicable to provide the best possible weld. Surfaces to be welded shall be free from grease, oil, oxides or any other foreign matter. Any cleaning method which will remove these substances and leave the surface clean is satisfactory. All weld joints shall blend into the adjacent metal in gradual, smooth curves. Beads shall be smooth and free of slag, excessive undercut, or excessive splatter. Sufficient metal shall be added to provide a suitable fillet or reinforcement unless otherwise specified. Welds shall not be oxidized. All welds shall be continuous with no visible voids, pin holes, or cracks.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2.1 Classification of inspection. The inspection requirements specified herein are classified as follows:

a. First article inspection.
b. Quality conformance inspection.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

4.1.2.1.1 First article inspection. First article inspection shall be performed on tables when a first article sample is required (see 3.1). This inspection shall include examination of 4.3 and tests of 4.4.

4.1.2.1.2 Quality conformance inspection. Quality conformance inspection shall include the examination of 4.3 and tests of 4.4.

4.1.3 Certificate of quality. Certificates of quality, supplied by the manufacturer of the materials may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for sampling purposes shall be one table.

TABLE I. Sampling for examination.

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Total defects (Major and Minor combined)	II	4.0
For dimensional examination	S-1	2.5

4.2.2 For tests. Sampling for the electrical conductivity test shall be conducted on 100% of the lot and the acceptance number shall be zero. The unit of product for sampling purposes shall be one table.

4.2.3 Raw materials. A sufficient quantity of each lot of corrosion-resisting steel used in the manufacture of the tables shall be selected for chemical analysis as specified in 4.4.1.

4.3 Examination. The tables shall be examined to determine compliance with all requirements contained in this specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table II. Examination shall not be restricted to the classified possible defects listed below.

TABLE II. Classification of defects.

Categories	Defects
Major:	
101	Components not properly aligned and fitted.
102	Metal parts not free of cracks, tears, ruptures, and deformed surfaces (check hemmed edges).
103	Any edge not flanged or hemmed as specified.
104	Any weld cracked, broken, or not sound.
105	Any joint between shelf corner and upright not closely fitted.
106	Caster cracked, broken, or not securely attached.
107	Any component missing.
108	Front of drawer not flush with front surface of drawer housing.
109	Drawer does not slide freely.
110	Drawer not provided with stops to prevent inadvertent removal.
111	Any screw stripped, missing, or loose.
112	Swivel caster does not swivel (where required).
113	Caster does not lock (where required).
Minor:	
201	Any exposed or accessible edge, corner, or surface not free of burrs, sharpness, and sharp projections.
202	Finish not uniform, not free of coarse grind marks and scratched areas.
203	Any weld not continuous, not cleaned, not free of pitting, splatter, and discoloration.
204	Any exposed weld not ground smooth to blend with adjacent surfaces.
205	Unit not clean.
206	Identification marking missing, incomplete, illegible, or not permanently attached.
207	Bumpers not provided to prevent drawer front from metal-to-metal contact with the housing.
208	Sound-deadening material missing from drawer front.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

4.3.2 Dimensional examination. The tables shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Chemical composition. Four ounces from each lot of material shall be tested in accordance with Fed. Test Method Std. No. 151.

4.4.2 Electrical conductivity. The completely assembled unit shall be tested using the procedure and the equipment described in NFPA 56A. The test shall be conducted with the metal plate successively under each caster.

4.5 Packaging inspection. The inspection of the packing and marking for shipment and storage shall be in accordance with quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

5. PACKAGING

5.1 Preservation-packaging. Preservation-packaging shall be level A or C, as specified, (See 6.2).

5.1.1 Unit of issue. One table of the type specified constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Cleaning. Tables shall be cleaned in accordance with MIL-P-116, process C-1.

5.1.2.2 Drying. Tables shall be dried in accordance with MIL-P-116.

5.1.2.3 Preservative application. Preservatives shall not be used.

5.1.2.4 Unit preparation. Each table shall be completely assembled with the exception of the drawer handles, which shall be attached on the inside of the drawers for shipping purposes. The drawers shall be secured with tape to prevent accidental opening. Points of contact of the table with interior faces of the container shall be cushioned with fiberboard pads or other suitable material. All cushioning material shall be secured with tape or by other suitable means.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

5.1.3 Level C. Each table of the type specified shall be packaged in a manner that will afford adequate protection against corrosion and physical damage during shipment from the supply source to the first receiving activity.

5.2 Packing. Packing shall be level A, B or C, as specified (see 6.2)

5.2.1 Level A. One table of the type specified shall be packed in an exterior container designed for a type 3 load and conforming to PPP-B-636, type CF, class weather-resistant, variety DW. Closure, strapping and waterproofing shall be as specified in the appendix of the box specification.

5.2.2 Level B. One table of the type specified shall be packaged in an exterior container designed for a type 3 load and conforming to PPP-B-636, type CF, class domestic, variety DW. Closure shall be as specified in the appendix of the box specification.

5.2.3 Level C. The packaged tables shall be packed in shipping containers in a manner that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall confrom to the applicable carrier rules and regulations.

5.3 Marking. Each exterior or shipping container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The tables covered by this specification are intended for use in military medical facilities.

6.2 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents.

a. Title, number and date of this specification.
b. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
c. Type required (see 1.2)
d. When a first article is required (see 3.1, 4.1.2.1 and 6.3).

6.3 First article. When a first article is required, it shall be tested under the appropriate provisions of the Armed Services Procurement Regulations. The contract or purchase order specifies the quantity, procedure for delivery and other specific instructions for inspection and approval of the first article.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

6.4 This specification covers the following items appearing in the Federal Supply Catalog.

National Stock No.	ITEM IDENTIFICATION	TYPE
6530-00-710-0220	TABLE, SURGICAL INSTRUMENT AND SURGICAL DRESSING, Corrosion-Resisting Steel, 20 by 16 by 34-1/2 inches.	I
6530-00-710-0305	TABLE, SURGICAL INSTRUMENT AND SURGICAL DRESSING, Corrosion-Resisting Steel, 48 by 24 by 40-1/2 inches.	II

6.5 Changes from previous issue. Asterisks are not used in this revision to identify changes with respect to the previous issue, due to the extensiveness of the changes

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activity:
DOD - MB

Preparing activity:
DLA-DM

Project No. 6530-1409

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No 22 R255

INSTRUCTIONS The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay or will otherwise enhance use of the document DoD contractors government activities or manufacturers / vendors who are prospective suppliers of the product are invited to submit comments to the government Fold on lines on reverse side staple in corner and send to preparing activity Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements Attach any pertinent data which may be of use in improving this document If there are additional papers, attach to form and place both in an envelope addressed to preparing activity

DOCUMENT IDENTIFIER AND TITLE
MIL-T-36202B TABLE, SURGICAL INSTRUMENT AND SURGICAL DRESSING

NAME OF ORGANIZATION AND ADDRESS	CONTRACT NUMBER
	MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1 HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A GIVE PARAGRAPH NUMBER AND WORDING

B RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2 COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3 IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes" in what way?)*

4 REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*	TELEPHONE NO
	DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

GPO 1974— 603 680/8126 2 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD 304

U.S. MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN DIRECTORATE OF MEDICAL MATERIEL CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:12Z
Check the source to verify that this is the current version before use.