MILITARY SPECIFICATION

FORCEPS, HEMOSTATIC CLIP, APPLYING, SMITHWICK

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a hemostatic clip applying forceps of the Smithwick design.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Boxes, Shipping Fiberboard.
PPP-B-676	Boxes, Set-up.

Military

MIL-L-116	Preservation, Methods of.
MIL-L-10547	Liners, Case and Sheet, Overwrap; Water-Vaporproof, or Waterproof, Flexible.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:27Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

20710	Forceps, Hemostatic Clip, Applying, Smithwick

(Copies of specifications, standards and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 General requirements. The hemostatic clip forceps shall be in accordance with the applicable requirements of GG-I-526 and the following option exercises under that specification:

3.2 Material. The hemostatic clip forceps shall be fabricated from Class 4, Type 420 material.

3.3 Hardness. After proper stress relieving, the Rockwell hardness of the hemostatic clip forceps shall be C40 to C45.

3.4 Style, design and dimensions. The hemostatic forceps shall be of the Smithwick design. Shall be suitable for use with National Stock Numbers 6515-00-358-6000 Roch, Hemostatic Clip, Mc Kenzie and 6515-00-377-7380 Clip, Hemostatic, Mc Kenzie, Silver, 100s. The forceps shall be in strict conformance with Defense Personnel Support Center Drawing 20710 except shall have an adjustable limiting screw.

3.5 Finish. The entire surface of the hemostatic clip forceps shall have a mirror, or mirror and satin finish.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:27Z
Check the source to verify that this is the current version before use.

3.6 Locks. Locks shall be of the box lock type. Locks shall be tested for stiffness as specified in GG-I-526.

3.7 Identification marking. The hemostatic clip forceps shall be marked with the name or registered trademark of the contractor. Forceps of other than U.S. manufacture shall also be marked stainless.

3.8 Workmanship. Workmanship shall be first class throughout. Instruments shall be free from defects which detract from their appearance or impair their serviceability, proper functioning and intended use. Instruments shall be furnished free of crevices, cracks, pits, burrs, scale and corrosion.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection shall be in accordance with GG-I-526.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the corrosion-resistant steel alloys may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:27Z
Check the source to verify that this is the current version before use.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and as indicated in Table I. The unit of product for examination purposes shall be one forceps.

Table I. Sampling for examination

Type defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table II. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics. The unit of product for test purposes shall be one forceps.

Table II. Sampling for tests

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL(percent defective)
Forceps	Hardness	3.3	GG-I-526	S-2	1.5
	Dimensions	3.4	--------	S-2	2.5
	Lock stiffness test	3.6	GG-I-526	S-2	2.5
	Boil test	GG-I-526	GG-I-526	S-2	1.0
	Copper sulfate test	GG-I-526	GG-I-526	S-2	1.0

4.2.3 A sufficient quantity of each lot of metal used in the manufacture of the forceps shall be selected for composition analysis in accordance with GG-I-526.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:27Z
Check the source to verify that this is the current version before use.

4.3 Examination. The forceps shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in Tables I and III. Examination, however, shall not be restricted to defects listed in Table III. Any forceps in the sample containing one or more defects shall be rejected and if the number of defective forceps in any sample exceeds the acceptance number for that sample, the lot represented by that sample shall be rejected.

Table III. Classification of defects

Categories	Defects
Critical	None defined.
Major	
101	Style and design not as specified.
102	Forceps not free from cracks, pits, burrs, sharp edges, rough areas, tool marks, nicks, dents, overlaps, crevices, scratches or blemishes.
103	Pin for box lock not flush with surface and not free from crevices.
104	Jaws misaligned.
Minor	
201	Forceps not identified as specified in 3.7
202	Finish not as specified.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.5 Examination of preparation for delivery. An examination shall be made to determine whether the packaging, packing and marking comply with the requirements of Section 5. Defects shall be scored as specified in Table IV. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 (defects per hundred units).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:27Z
Check the source to verify that this is the current version before use.

Table IV. Examination of preparation for delivery

Examination	Defect
Containers	Not as specified.
Quantity	Quantity in interior packages and shipping containers not as specified.
Packaging and packing materials	Any component missing or damaged.
Workmanship	Inadequate application of components such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, incorrect, illegible, improper size, location, sequence, or method of application.

4.5.1. Examination of closure and reinforcement for containers. When shipping containers are required to comply with PPP-B-636, examination for defects in closure and reinforcement shall be in accordance with the appendix to that specification.

5. PREPARATION FOR DELIVERY

5.1 Preservation and Packaging. Preservation and packaging shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Cleaning, drying and preservative. Each forceps shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116. Preservative shall not be applied.

5.1.1.2 Unit package. Each forceps shall be wrapped in neutral tissue or other suitable anticorrosive material. Each forceps shall be further packaged in accordance with MIL-P-116, method IC-1.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:27Z
Check the source to verify that this is the current version before use.

5.1.1.3 Intermediate package. Twelve (12) forceps shall be packaged in a box of appropriate size and design constructed in accordance with PPP-B-676 or PPP-B-636, type CF, class domestic. Closure shall be adequate to prevent accidental opening.

5.1.2 Level C. Forceps shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. One hundred and forty four (144) forceps shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3c. Closure shall be as specified in the appendix of the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof in accordance with the appendix of PPP-B-636.

5.2.2 Level B. One hundred and forty four (144) forceps shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers at the lowest rate to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior and exterior containers. Each unit, intermediate and exterior container shall be marked in accordance with MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:27Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Intended use. The hemostatic clip forceps of the Smithwick design are used in neurosurgical procedures.

6.2 Ordering data. Purchasers should exercise any desired options offered herein, and the procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Selection of applicable levels of packaging and packing. (see 5.1 and 5.2).
(c) National Stock Number.

6.3 National stock no. coverage. This specification covers the following item appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6515-00-065-3182	FORCEPS, HEMOSTATIC CLIP, APPLYING, SMITHWICK

It does not include all types or sizes as indicated by the title nor all types which are commercially available.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1980
is

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:27Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-F-37332(DSA-DM) FORCEPS, HEMOSTATIC CLIP, APPLYING, SMITHWICK

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:27Z
Check the source to verify that this is the current version before use.