MILITARY SPECIFICATION

TEST KIT, BILIRUBIN DETERMINATION

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers Test Kit, Bilirubin Determination (See 6.2).

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

FEDERAL

PPP-C-186 - Containers, Packaging and Packing for Drugs, Chemicals, and Pharmaceuticals (see 6.3)

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

(Copies of specifications and standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Hq, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-ATT, 2800 S. 20th St., Phila. PA 19101, by using the self-addressed Standardization Document Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6505

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:27Z
Check the source to verify that this is the current version before use.

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

U. S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE, FOOD AND DRUG ADMINISTRATION

Federal Food, Drug, and Cosmetic Act and Regulations Promulgated Thereunder.

(Application for copies should be addressed to the Food and Drug Administration, Washington, DC 20204.)

3. REQUIREMENTS

3.1 Material. Shall be tablets containing Sulfosalicylic Acid, Sodium Bicarbonate, Boric Acid, and 0.40 to 0.65 mg of p-Nitrobenzenediazonium p-Toluenesulfonate.

3.1.1 Shall be suitable for determining urine bilirubin.

3.1.2 The test kit shall consist of 100 tablets and 100 absorption mats which are composed of suitable material for use in the bilirubin test, in accordance with commercial practice.

3.1.3 The tablets shall be tested as specified in 4.3.

3.2 Expiration dating period. The expiration dating period shall be not less than 60 months.

3.2.1 Delivery. Not more than six (6) months of the expiration dating period shall have elapsed at the time of delivery to the Government.

3.3 Workmanship. The material and its containers shall be free from defects which may detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:27Z
Check the source to verify that this is the current version before use.

4.1.1 Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and, (i) as to any expiration dated item, for the expiration dating period specified by the contractor for such item, or such longer period as may be required by regulation of any federal agency; or (ii) as to non-dated items, for not less than 3 years after delivery of the items to the Government.

4.1.2 No company supplying any ingredient(s) to the contractor will be considered an acceptable facility for the performance of any inspection requirements specified herein.

4.2 Lot. For purpose of this specification, a lot, batch, or control is that single uniform quantity of tablets produced from one formulation, subjected to the same manufacturing operation, and placed in the immediate containers.

4.3 Tests. The tablets shall be tested by appropriate accurate, and reproducible method(s) as used for commercial release.

4.4 Examination of preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of Section 5. Defects shall be scored as specified in table I. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 defects per hundred units.

TABLE I. Examination of preparation for delivery 1/

Examination	Defect
Containers	Not as specified.
Quantity	Quantity in interior packages and shipping containers not as specified.
Packaging and packing materials	Any component missing or damaged.
Workmanship	Inadequate application of components such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, incorrect, illegible, improper size, location, sequence, or method of application.

1/ The above examination is only applicable to the unit, intermediate and exterior containers.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:27Z
Check the source to verify that this is the current version before use.

5. PACKAGING

5.1 Packaging requirements. The requirements for package shall be in accordance with all applicable requirements of PPP-C-186 (see 6.3), as specified herein.

5.2 Packaging. Packaging shall be level A or C, as specified (see 6.1).

5.2.1 Unit of issue. One test kit containing 100 tablets and 100 absorption mats, as specified, constitutes one unit of issue.

5.2.2 Level A.

5.2.2.1 Immediate containers. Immediate container shall be a commercially available immediate container considered acceptable by the Food and Drug Administration (FDA) for the product contained therein and shall be of such compatibility that neither the contents nor the container are altered in any way by each other.

5.2.2.2 Unit package. Each unit shall be packaged as specified in 5.2.5 of PPP-C-186.

5.2.2.3 Intermediate package. Forty-eight (48) units of issue shall be packaged as specified in 5.2.7.1 of PPP-C-186.

5.2.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.3 Packing. Packing shall be level A, B or C, as specified (see 6.1).

5.3.1 Level A. Two hundred eighty-eight (288) units of issue shall be packed as specified in 5.4.4 of PPP-C-186.

5.3.2 Level B. Two hundred eighty-eight (288) units of issue shall be packed as specified in 5.4.3.1.1 of PPP-C-186.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:27Z
Check the source to verify that this is the current version before use.

5.3.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.4 Marking.

5.4.1 Labeling. Labeling shall be in accordance with the requirements of the Federal Food, Drug and Cosmetic Act, and shall include the information required below.

5.4.1.1 Immediate container. Each immediate container label shall bear the following information:

(a) label information in accordance with commercial practice.
(b) the expiration date.

5.4.1.2 Unit package. Each unit package label shall bear the following information:

(a) label information in accordance with commercial practice.
(b) the national stock number.

NOTES: (1) The National stock number shall not be preceded by any prefix.
(2) Strip labeling may be used.
(3) Label adhesion test shall not apply.

5.4.2 Intermediate package. Each intermediate package shall be marked as specified in 5.5.3 of PPP-C-186. Type I shelf life markings shall be shown as specified in MIL-STD-129.

5.4.3 Exterior container. Exterior container shall be marked as specified in 5.5.4 of PPP-C-186. Type I shelf life marking shall be shown as specified in MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number and date of this specification.
(b) the National Stock Number.
(c) Selection of applicable levels of packaging and packing (See 5.2 and 5.3).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:27Z
Check the source to verify that this is the current version before use.

6.2 This specification covers the following item(s):

National Stock No.	Item Identification
6505-01-008-6C10	TEST KIT, BILIRUBIN DETERMINATION

6.3 All references to PPP-C-186 refer to Interim Federal Specification PPP-C-00186B which is mandatory for use until superseded by a revision.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing acitivity:
Defense Supply Agency - DM

Project No. 6505-1727

☆ U. S. GOVERNMENT PRINTING OFFICE: 1976-703-020/329

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:27Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-T-37329A — TEST KIT, BILIRUBIN DETERMINATION

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

GPO 1974— 603 680/6126 2-1

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:27Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304



OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:27Z
Check the source to verify that this is the current version before use.