
FED. SUP CLASS
6640



MS36451-1 MS36451-2 MS36451-3 MS36451-4 MS36451-7
MS36451-5 MS36451-8
MS36451-6

PART NUMBER	DIMENSIONAL DATA (MM)	
	LENGTH	DIAMETER
MS36451-1	100	17
MS36451-2	100	17
MS36451-3	100	17
MS36451-4	150	16
MS36451-5	150	16
MS36451-6	150	16
MS36451-7	150	16
MS36451-8	150	16

MATERIAL: Tube shall be of clear, colorless plastic. Friction type caps shall be of rigid plastic. Applicators may be of wood or plastic, tipped with surgical cotton or polyester-fiber.

DESIGN: Applicator shall consist of a wood or plastic stick tipped with either cotton or plastic fibers and mounted in a cylindrical closed bottom tube with a flexible friction type or screw type cap. All screw capped swab applicators shall be coded as follows:

Red Cap - denotes cotton tip
Green Cap - denotes polyester fiber tip

MS36451-1 shall have a friction cap that has two sockets to accommodate two cotton-tipped wood applicators.

MS36451-2 shall have a friction cap that supports one cotton-tipped applicator.

MS36451-3 shall have a friction cap that supports a cotton-tipped wood applicator shielded to assist insertion and reduce contamination of swab during insertion. Shield shall be molded of plastic.

MS36451-4 shall have a screw cap that supports a single wood cotton tipped applicator.

(B) ENTIRE STANDARD REVISED

USER ACTIVITIES: ARMY - ME, DM
NAVY - MS

ARMY - ME, MD
NAVY - MS
AIR FORCE - 03

PREPARING ACTIVITIES:

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Suggestions for all new engineering and design applications and for repetitive use shall be made from this document.

APPROVED 14 DECEMBER 1967 REVISED (B) 10 FEB 1982

FSC 6640	INTERNATIONAL INTEREST	TITLE: TUBE, BIOLOGICAL CULTURE SAMPLING	MILITARY STANDARD MS36451
Instrumentation Specification: NONE	SUPERSEDES:		SHEET 1 OF 2

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-0598
REVISION DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:40Z
Check the source to verify that this is the current version before use.

MS36451-5 shall have a screw cap that supports a single wood or plastic polyester cotton-tipped applicator.

MS36451-6 shall have a screw cap that supports a cotton-tipped flexible plastic applicator.

MS36451-7 shall have a screw cap that supports two cotton-tipped wood applicators.

MS36451-8 shall have a screw cap that supports two cotton-tipped plastic applicators.

NOTES:

1. Applicator assemblies shall be sterile and wrapped in plastic film strip packs with individually sealed compartments.
2. Applicator assembly shall be discarded after a single use.
3. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

PREPARING ACTIVITY:	INTERNATIONAL INTERCHANGE	TITLE: TUBE, BIOLOGICAL CULTURE SWABBING	MILITARY STANDARD MS36451
			SHEET 2 OF 2

Source: https://assist.dla.mil -- Downloaded: 202610-23T17:40Z
Check the source to verify that this is the current version before use.