A-A-0051837(DM)
20 February 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

DRAPE, SURGICAL

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This interim commercial item description covers one size of a sterile disposable, nonwoven fabric, split sheet, surgical drape with a centered opening at tail or foot end. Aperture shall be reinforced with an absorbent but impervious laminate panel. The surgical drape shall be suitable for general surgery.

Salient characteristics:

The surgical drape shall be sterile, packaged in a sealed pack containing one disposable split drape and one outer wrap suitable for use in surgical operating rooms of medical treatment facilities. The drape and outerwrap shall be made of a nonwoven fabric.

Physical characteristics of the drape shall be as follows:

Drape size:	Minimum - 118 in. (300 cm) x 86 in. (218 cm)
Aperture size:	
1) Length:	Minimum - 44 in. (112 cm)
2) Width:	Minimum - 1.5 in. (3.7 cm) Maximum - 4 in. (10.16 cm)
Panel:	Minimum - 35 in. (89 cm) x 17 in. (43 cm)
Panel border (inner end):	Minimum - 13 in. (33 cm)

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Tensile strength:

1)	Machine direction:	Min. avg. - 10.0 lbs. (4.5 kg)/ in. (2.54 cm)
2)	Cross direction:	Min. avg. - 8.0 lbs. (3.6 kg)/ in. (2.54 cm)

Weight:	Minimum - 2.0 oz. (56 g)/sq. yd (.836 sq. m)

Hand of fabric:

1)	Machine direction:	Max. avg. - 25.0 g
2)	Cross direction:	Max. avg. - 35.0 g

Electrostatic:	Maximum - 5 x 10 ohms/sq. in. (6.45 sq. cm) at 50 percent RH
Bursting strength:	Minimum average - 18.0 lbs. (8 kgs)
Water repellency:	Minimum - 20 minutes
Color:	Pastel shades of blue, green, or gray, on minimum of one glare-free face.

The drape shall be constructed of fully bleached pulp tissue bonded to nylon or rayon reinforcing yarns and treated for moisture repellency sufficiently to be repellent to low surface tension fluids. Tissue of drape shall be approximately 40 pounds (18 kg) basis weight.

Drape and outerwrap shall be abrasion resistant and free of lint, objectionable noises when used and objectionable odors. Shall have no irritating or toxic elements present in the finished fabric.

Each drape shall be enclosed within a resin bonded, nonwoven fabric outerwrap. Minimum tensile strength of outerwrap fabric to be 4.5 lbs. (2 kgs) per inch (2.54 cm) machine direction and 1.5 lbs. (0.7 kgs) per inch (2.54 cm) cross direction. Color of outerwrap shall be white or pastel shades of blue, green or gray.

Drape and outerwrap shall be functionally folded for optimum aseptic technique. Folded drape shall be clearly marked with the following or similar identifications:

SPLIT SHEET, TAILS (with arrow pointing in direction of tails)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Drape and outerwrap shall be enclosed in a sealed paper or plastic immediate container capable of maintaining sterility of contents until manually opened. If the sealed container is paper, it shall be further enclosed in a plastic bag.

The drape and outerwrap in its immediate container shall be sterile when tested with a biological indicator (or alternate method) as specified in the sterility test.

A certificate of sterility stating the samples from each sterilization lot have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:

a. Item Identification.
b. National Stock Number.
c. Contract or purchase order number.
d. Sterility lot or control number.
e. Date of sterilization.
f. Quantity of product in the sterilizer lot.
g. Type of sterilization (gas, radiation, heat, steam, etc.)
h. Form of biological indicator (i.e., inoculated paper spore strip or carrier, inoculated product, or inoculated simulated product).
i. Test organism(s) and population density used.
j. Test medium, incubation temperature, incubation time, and dates of test.
k. Number of test units.
l. Signature that samples comply with sterility test.
m. Signature of head of laboratory or a designated qualified alternate.

Workmanship. The surgical drape shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing twenty-two individually packaged sterile drapes with an outerwrap for each drape, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contact or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one (1) surgical drape. Unless otherwise specified, sampling for examination shall be conducted at Inspection Level II, AQL 2.5 and sampling for tests shall be conducted at Inspection Level S-1, acceptable number is zero.

Sterility test. The surgical drape shall be tested for sterility in accordance with the method specified in the U.S. Pharmacopeia Convention Publication: Pharmacopeia of the United States, Mack Publishing Co., Easton, PA 18042 (USP), for purified cotton, gauze, surgical dressings and related material using inoculated product, inoculated simulated product, or paper spore strip (carrier). The incubation temperature shall be as listed below. Product containing a biological indicator shall be marked or labeled, so that it clearly differentiates from product intended for product intended for distribution. The incubation medium and minimum time in each case shall be soybean case in digest for 7 days for inoculated product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

ne minimum number of test units, regardless of lot size, shall be 10, except for inoculated carrier which shall be 20. For sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism	Incubation Temperature Degrees C
Steam	B.stearothermophilus	55 - 65
Dry Heat	B.stearothermophilus, or	55 - 65
	B.subtilis var. niger (or globigii)	30 - 35
Gas	B.subtilis var. niger (or globigii)	30 - 35
Radiation	B.pumilus	30 -35

The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

Inoculated carriers (paper sport strips) if used shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method, incubation and number of test units shall be as indicated herein for inoculated product or simulated product. The method of sterilization will control he organism in the spore strip to be tested for as shown above.

NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP.

Alternate method. Suppliers conducting sterility tests in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a Government contract.

Product testing. Testing for sterility of the product after receipt by the Government shall be without a biological indicator and in accordance with the USP.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-01-110-9192

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS.

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES.

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-1948

Location. CID/DRAPE/H18-16

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standard zation documents better, the DoD provides this form for use in submitting comments and suggestions for improvements All users of military standardizat.on documents are invited to provide suggest-ons This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)* and mailed In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was (is) rigid restrictive loose ambiguous or was incompatible, and give proposed wording changes which would alleviate the problems Enter in block 6 any remarks not related to a specific paragraph of the document If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN DPSC-ATTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM CUT ALONG THIS LINE)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-0051837(DM)	Drape, Surgical

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street City State ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a NAME OF SUBMITTER *(Last First MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.