A-A-51792A
31 August 1987
SUPERSEDING
A-A-0051792(DM)
8 January 1987

COMMERCIAL ITEM DESCRIPTION

SIMULATOR, PATIENT

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers a patient simulator designed to assist biomedical electronics specialists in the inspection and repair of multiparameter patient monitors.

Salient characteristics:

The unit shall be a portable hand-held patient simulator capable of producing the following simulated physiological parameters with ± one percent sensitivity, resolution and accuracy full scale:

a. Normal sinus, pulse train, VFib and paced rhythms.
b. Calibrated dynamic arterial and PA blood pressure waveforms of approximately 120/180 mmHg.
c. 13 calibrated static pressures from -200 to 200 mmHg.
d. 4 Respiration rates with 3 different lead selections
e. 3 Respiration impedance variations and APNEA
f. Blood pressure.

The simulator shall provide a full 12-lead ECG simulation with variable and discrete ECG rate control with calibrated output amplitudes.

The electrocardiograph simulator shall be provided with 3 temperature simulation outputs with a range of at least 30 - 40°C (86°F - 104°F).

The unit shall have a vinyl carrying case or equal for protection.

The patient simulator shall be operational utilizing a nominal nine volt standard alkaline battery and have an indication light to indicate low battery conditions.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Workmanship. The patient simulator shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One simulator, as specified, constitutes one unit. Operation and Service Data Manuals shall be furnished as follows:

Operational and Service Data Manuals.

a. Information Content:

The contractor shall furnish manuals, handbooks and/or brochures containing complete operation, installation and service/maintenance instructions (including pictures of illustrations, as necessary) with complete schematics and wiring diagrams. All manuals, handbooks, and/or brochures will be written in the English language and all schematics and wiring diagrams will use American electrical and electronics symbols. The manuals will include electrical data and connection diagrams for all applicable utilities. The instructions shall also contain a complete list of all replaceable parts showing part number, name and quantity required. All service notes, service memos, etc, pertaining to the equipment and issued subsequent to the printing of the original instructions shall also be included. These instructions shall be as a minimum, that furnished to service engineers (or servicemen) that normally install and service the equipment for the company or distributor. When the system being procured includes, or will operate ancillary equipment, the service data shall include complete instructions and drawings which show interfacing of all system components.

b. Quantities, Distribution and Schedules:

Two (2) complete sets of the above information shall be packaged with each unit of the required equipment. Where multiple quantities of the same unit are to be delivered to the same ultimate consignee, then only a total of 2 service manuals need to be supplied to each consignee. However, 2 operation manuals still need to be supplied with each unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one one patient simulator.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-01-207-8236

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PBS
FDA-MPQAS

Project No. 6515-4114

Location: CID/SIMULATOR/HlD-19

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.