A-A-51794B
12 October 1990
Supersedes
A-A-51794A
8 June 1987

COMMERCIAL ITEM DESCRIPTION

PIPET, SEROLOGICAL (Kahn)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers borosilicate glass Kahn serological pipets.

Salient Characteristics:

Shall be borosilicate glass Kahn serological pipets for use in laboratory procedures.

Shall be Style I for the following ASTM Designation E1044 Kahn milliliter capacities 0.1 ml, 0.2 ml, 0.25 ml, and 0.60 ml.

Shall be in accordance with all the applicable requirements of ASTM Designation E1044 for style I as applicable to Kahn pipets (see "Ordering data").

Shall be tested for volumetric accuracy to meet the requirements of the College of American Pathologists (see Quality Assurance Provisions).

Workmanship. The pipets shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing eighteen (18) pipets of the size specified, constitutes one unit.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements.

Volumetric accuracy testing. To meet requirements for certification of the College of American Pathologists, the pipets shall be tested in accordance with ASTM Designation E542 to determine conformance to calibration and capacity tolerance requirements contained in ASTM Designation E1044. A Certificate of Quality may be furnished in lieu of actual performance of individual testing, provided that lot/sample identities can be maintained. The Certificate of Quality shall include the lot number, lot size, sample size and test method.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:01Z
Check the source to verify that this is the currentversion before use.

Other publications. The following documents form a part of this specification to the extent that is specified herein. Unless otherwise indicated, the effect on date of invitation for bids or request for proposals shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM E542 Practice for Calibration of Volumetric Ware

ASTM E1044 Standard Specification for Glass Serological Pipets (General Purpose and Kahn)

(Application for copies should be addressed to American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103).

NOTE: The following National Stock Number is covered by this document.

NSN	CAPACITY	ITEM IDENTIFICATION
6640-01-239-8284	0.25 ml	PIPET, SEROLOGICAL, Kahn 0.25 ml, 18's

Ordering data. (Capacity required, intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.)

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project Number: 6640-1658

Location: ENABLE/PIPSERCD/S31

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.