INTERIM COMMERCIAL ITEM DESCRIPTION

PIPET, SEROLOGICAL, Borosilicate Glass

The General Services Administration has authorized the use of this interim commercial item description in preference to Federal Specification NNN-P-375 as it applies to the serological pipets described herein.

This Interim Commercial Item Description covers one type of serological pipets suitable for use in laboratory procedures.

Salient Characteristics.

Serological pipets shall be fabricated of a borosilicate glass suitable for use in laboratory procedures. The pipets shall have the capacities, markings and graduations as listed in Table I.

TABLE I. Dimensional Characteristics

Capacity	Marked at intervals	Numbered at intervals of	Graduated to at least	Outside diameter	Coding band
0.1 (ml)	0.01 (ml)	0.01 (ml)	0.09 (ml)	5.5 (mm)	white
0.2	0.01	0.02	0.19	5.5	black
1.0	0.01	0.1	0.98	6.0	yellow
5.0	0.1	1.0	4.8	7.5	blue
10.0	0.1	1.0	9.8	9.5	orange

The pipets shall be a one piece construction. Any cross-section taken in a plane perpendicular to the longitudinal axis shall be circular. The shape of the pipet shall permit complete emptying and thorough cleaning. The delivery tip shall be tempered and shall be gradually tapered to the end which is perpendicular to the longitudinal axis of the pipet. The outside edge of the tip shall be beveled at the end and the bevel shall be ground or fire-polished leaving no rough or jagged edges. Pipets shall have a general purpose mouthpiece and shall be calibrated to deliver at 20°C when the last drop is blown out. 5.0 and 10.0 milliliter (ml) capacity pipets shall have a mouthpiece which is reduced in diameter and constricted for cotton plugging.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:01Z
Check the source to verify that this is the current version before use.

Graduation markings on the pipets shall be sharply defined, narrow lines of uniform width in a plane perpendicular to the longitudinal axis and parallel to each other. The width of the markings shall be 0.3 ± 0.1 millimeters (mm). The main graduation markings shall be a continuous line extended at least 3/4 of the way around the pipet. The value of all main graduation markings shall be suitably marked in Arabic numbers directly above the lines referenced. Intermediate graduation markings shall extend at least 1/5 of the way around the pipet and least graduation markings shall extend at least 1/7 of the way around the pipet. All markings shall be etched, fired enamel or fired stain.

The pipet shall be marked with the manufacturer's name or trademark, its capacity, the symbol "TD" and the temperature of calibration (20°C). The pipet shall be identified by a wide band permanently marked on the mouthpiece as specified in Table I. The color coding bands shall be bright and distinct and shall withstand normal laboratory procedures with no damage other than loss of gloss. The distance from the mouthpiece opening to the top graduation of the pipet shall not be less than 10 centimeters (cm) nor more than 15 cm from the 0.0 ml graduation line.

The serological pipets shall meet the requirements for certification of the College of American Pathologists by being tested in accordance with NBS Circular 602 and NBSIR 74-461. A Certificate of Quality may be furnished in lieu of actual performance of individual testing, provided that lot/sample identities can be maintained. The Certificate of Quality shall contain the lot number, lot size, sample size and test method. The pipet shall comply with the bureau of Medical Devices and Diagnostics.

Workmanship. The pipets shall be free from defects which detract from its appearance or impair its serviceability. Pipets shall be transparent, clean, dry, and free of foreign material.

Unit. Package (PG). One package containing eighteen pipets, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractor/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6640-01-239-8283	PIPET, SEROLOGICAL, Borosilicate Glass, 0.1 ml
6640-01-243-3218	PIPET, SEROLOGICAL, Borosilicate Glass, 0.2 ml
6640-01-239-8285	PIPET, SEROLOGICAL, Borosilicate Glass, 1.0 ml
6640-01-239-9799	PIPET, SEROLOGICAL, Borosilicate Glass, 5.0 ml
6640-01-239-6962	PIPET, SEROLOGICAL, Borosilicate Glass, 10.0 ml

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.)

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA - DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITY
VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6640-1425

Location: NCR Disk # 3

5 U.S. GOVERNMENT PRINTING OFFICE: 1987 - 704-036/70622

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:01Z
Check the source to verify that this is the current version before use.