A-A-51802A
17 May 1989
Superseding
A-A-0051802(DM)
13 February 1987

COMMERCIAL ITEM DESCRIPTION

PRESS, HYDRAULIC, OPHTHALMIC LAP

The General Services Administration has authorized the use of this commercial item description.

This Commercial Item Description covers an electric motor operated, bench-mount hydraulic lap press for use in optical fabrication laboratory.

Salient characteristics:

Shall be intended to force a fining pad onto a metallic lap surface to produce a near perfect wrinkle free adhesion between two surfaces. The lap surface to be serviced shall be either convex cylinder, concave cylinder or convex spherical lap. Provide necessary press adapter(s) to service above laps.

Shall consist of two interconnected modules: pump power assembly and ram working platform. The pump power assembly shall contain a hydraulic pressure pump-electric motor module with an oil sump. Control valve and built-in relief valve shall be provided with this module. The ram working platform shall contain ram supports for the hydraulic cylinder and a flat working surface below the cylinder for placement of the lap to be serviced.

The hydraulic pump shall be capable to generate 12,000 psi maximum pressure. Normal operating pressure shall not be less than 8,000 psi.

Provide a complete set of piping, fittings, interconnecting hoses and pressure gage ready for connection and start-up. Also, provide one initial fill of hydraulic oil stored in a separate container.

The motor shall be able to operate on 115/230 volt, 60 hertz power source. An On-Off switch shall be mounted on or near the motor. Electrical cord shall be not less than 6 feet with 3-prong plug.

AMSC N/A FSC 6540

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:01Z
Check the source to verify that this is the currentversion before use.

Service data. The contractor shall enclose, with each unit, two copies each, of one or more service data booklets. The service data shall contain complete instructions for installation, operation, maintenance, and a list of component parts. Wiring diagrams shall be included for electrical items or components. Each part of the unit, other than common fasteners, shall be identified in the service data and parts list by the manufacturer's part number.

Workmanship. The hydraulic press shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One hydraulic press, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:01Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6540-01-210-9491	PRESS, HYDRAULIC, OPHTHALMIC, LAP

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

ARMY - MD
NAVY - MS
AIR FORCE - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

Project No. 6540-1213

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.