A-A-0051823(DM)
24 February 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

TUBE KIT, MICROHEMATOCRIT

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers tube kits suitable for use in microhematocrit procedures.

Salient characteristics:

Capillary Tube Kits covered by this specification shall be of the following types as specified for the appropriate stock number and item identification:

Type I - Venous
Type II - Capillary

Shall be a venous or capillary microhematocrit tube kit to perform 100 venous or capillary hematocrits, total white blood cell counts, platelet counts and partial differential determinations. Shall be compatible with NSN 6630-01-234-6794 (Centrifugal Hematology Analyzer). Shall contain 100 tubes, 100 floats and 100 tube closures.

Type I. Type I tubes shall have a black line at a measured volume of 111.1 microliters.

Type II. Type II tubes shall have two black lines at a measured volume of 55 and 65 microliters.

Tubes. Tubes of both types shall be clean, transparent and fabricated of borosilicate glass, suitable for the purpose intended. The length of the tubes shall be 75 mm and have a inner diameter of 1.677 mm (.06615in. with a tolerance of 0.00045in.). Bore of the tube shall be uniform.

AMSC/NA FSC 6630

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Tubes shall be coated with 9.0 ug acridine orange and 0.81 mg potassium oxalate. Shall be coated to prevent nonseparation of cellular layers. Tubes shall be straight and open at both ends without lip or obstruction. Tubes shall be cut at 90° to the tube axis and shall not have sharp, chipped or jagged ends.

Float. Shall be a solid injection molded plastic cylinder and have a specific gravity of 1.055. Shall occupy 90.66% of the inner diameter of the hematocrit tube. Shall function so as to expand for qualitive purposes, granulocytes, nongranulocytes and platelets.

Tube closure. Shall be a cap suitable for fitting over one end of the tube. Shall have a friction fit and form a leakproof seal to allow safe sampling for hematocrit valves.

Workmanship. The Microhematocrit tube kits shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One Microhematocrit Tube Kit, as specified, constitutes one unit.

Inspection/Test date. The inspection/test date shall be 18 months from the date of manufacture.

Date of Delivery. Not less than 15 months of the inspection/test period shall remain at the time of delivery to the government.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:01Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for test shall be conducted in accordance with commercial practice to determine compliance with specification requirements.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:01Z
Check the source to verify that this is the currentversion before use.

NOTE: The following National Stock Number is covered by this document:

Type I 6630-01-233-7592 Venous Microhematocrit Tube Kit
Type II 6630-01-233-7594 Capillary Microhematocrit Tube Kit

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.)

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

AGENT:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6630-0452

Location: NCR Disk #3

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:01Z
Check the source to verify that this is the current version before use.