A-A-51827
13 February 1987

COMMERCIAL ITEM DESCRIPTION

THERMOMETER, CLINICAL, HUMAN, (Disposable)

The General Services Administration has authorized the use of this commercial item description in preference to MIL-T-37832 (DLA-DM) dated 24 January 1978.

This commercial item description covers the requirements for a disposable thermometer suitable for obtaining oral or axillary temperatures in humans.

Salient characteristics:

Design and construction. The thermometer shall be fabricated from a suitable plastic material that is non-toxic, non-irritating and non-injurious to human tissue. The plastic strip shall be at least 3-1/2 inches in length and shall terminate at the distal end in a sensor matrix.

The sensor matrix shall consist of forty-five temperature indicating dots arranged in five dots to a row. Each dot shall contain different chemical mixtures that melt and recrystallize at specific temperatures. Each dot shall represent 0.2 degree Fahrenheit providing a minimum range from 96 degrees F to 104.8 degrees F. The unfired color of each dot shall be beige and upon being fired shall become blue.

Graduation markings. Each row of sensor dots shall be marked with the whole number temperature from 96°F to 104°F. Each dot in a row shall be marked in 0.2°F increments, from 0.0°F to 0.8°F. The graduation numbers shall be highly legible.

Reversibility. The thermometer shall have the capability of being reversed at low temperature if inadvertently exposed to firing temperatures. Thermometers which have been accidently fired and have not been allowed to remain in fired state for more than 12 hours at a temperature not exceeding 110°F, can be reversed to their unfired state by being placed in low freezer temperature of 18°F or in containers with dry ice.

Sterility. Each thermometer shall be sealed in a peel-open paper wrapper which shall maintain sterility until manually opened or damaged. The sterility shall be tested in accordance with the United States Pharmacopia (USP).

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:01Z
Check the source to verify that this is the current version before use.

Shelf life. Thermometer shall have a minimum shelf life of 24 months. An uncoded expiration date shall be displayed on the labeling of the product. Not more than three months of the expiration dating shall have elapsed at the time of delivery of the product to the Government.

Performance. Thermometer shall conform to all requirements of the current issue of ASTM E825, applicable for the thermometer with Fahrenheit scale.

Order of precedence. In the event of a conflict between the requirements of the ASTM E825 and other requirements described herein, the requirements specified in this document shall take precedence.

Workmanship. The thermometer shall be free from defects which detract from its appearance or impair its serviceability.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Package (PG). One hundred individually packaged thermometers shall be packaged in a box of suitable size and design. Five boxes containing 100 thermometers each (500 thermometers) or twenty boxes containing 100 thermometers each (2000 thermometers) shall be packed in a suitable insulated container capable of maintaining a temperature below 90°F. One package containing 500 or 2000 thermometers, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one thermometer. Inspection level shall be S-2 with an AQL of 1.5.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-042-2486	THERMOMETER, CLINICAL, HUMAN, DISPOSABLE, 500s
6515-01-105-0432	THERMOMETER, CLINICAL, HUMAN, DISPOSABLE, 2000s

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MQPAS

Project No. 6515-3770

U.S. GOVERNMENT PRINTING OFFICE: 1987 - 704-036/70727

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.