A-A-0051836(DM)
7 APRIL 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

FORCEPS, EYE SUTURE, STRAIGHT, TWEEZERS-TYPE (Castroviejo)

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers a straight, tweezers-type eye suture forceps with various size teeth and of the Castroviejo design used for microsurgical eye procedures.

Salient characteristics.

General. Shall be an eye suture forceps of the Castroviejo design having various sized teeth, a tying platform, and wide handles.

Material. The forceps shall be fabricated from Class 4 martensitic corrosion-resistant steel in accordance with ASTM F899.

Hardness. After heat treatment, the forceps shall have a finished Rockwell hardness of C42 to C47 in accordance with ASTM E18 or E92.

Passivation. As a final operation, the instrument shall be passivated.

Corrosion-resistance. When copper sulfate and boil corrosion tested in accordance with Federal Specification GG-I-526, there shall be no evidence of plating or corrosion.

Characteristics. The forceps shall be 4 inches long with either 0.12, 0.3, or 0.5 mm teeth, and shall have a tying platform, wide handles, and interdigitating teeth. When the jaws are closed, there shall be no protruding points. The handles shall have a stop pin, and serrations for a positive grip. The forceps shall be in accordance with DPSC Drawing 23389.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Finish. Shall be mirror, satin, or matte. Should the forceps possess a matte finish, it shall exhibit a uniform, non-directional surface texture equivalent to that produced by lightly blasting the prepared surface with fine abrasives under pressure.

Identification marking. In addition to the requirements specified in ASTM F1026, forceps of foreign manufacture shall be marked "stainless."

Workmanship. The forceps shall be free from cracks, pits, sharp edges, corrosion, foreign matter, or other defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One forceps, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:01Z
Check the source to verify that this is the currentversion before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The inspection level shall be S-2, with dimensional testing at a 2.5% AQL, hardness testing at a 1.5% AQL, and corrosion testing at a 1.0% AQL. The unit of product for test purposes shall be one forceps.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-115-7667	FORCEPS, EYE SUTURE, Castroviejo, Straight, 0.12 mm Teeth

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:01Z
Check the source to verify that this is the current version before use.

NSN	ITEM IDENTIFICATION
6515-01-115-7668	FORCEPS, EYE SUTURE, Castroviejo, Straight, 0.3 mm Teeth
6515-01-115-7816	FORCEPS, EYE SUTURE, Castroviejo, Straight, 0.5 mm Teeth

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4019

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:01Z
Check the source to verify that this is the currentversion before use.