A-A-51838
9 March 1987

COMMERCIAL ITEM DESCRIPTION

STRIP ASSORTMENT, ABRASIVE, DENTAL
(Aluminum Oxide, Metal-Backed)

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification GG-S-00783C (DSA-DM).

This commercial item description covers an assortment of abrasive strips suitable for use in dental polishing procedures.

Salient characteristics:

Shall be an assortment of flat metal, abrasive strips coated on one side with aluminum oxide abrasive and one side safeside. The abrasive strips shall be suitable for finishing restorations in the oral cavity under both wet and dry conditions. The strip assortment shall consist essentially of aluminum oxide abrasive particles bonded to one side of a metal backing and the opposite side a safe side with both faces nickel electroplated. The strip shall be flat and smooth. The coatside surface shall be a closed coat of abrasive particles, uniformly distributed and firmly attached to the metal strip. Style, design and dimensions of the strips shall be as shown in figure 1.

Strip assortment. The strip assortment shall consist of twelve strips, four each of narrow, medium, and wide.


Material. The abrasive shall be 180 grit crystalline aluminum oxide. The strip backing shall be of a low carbon steel, with a carbon content of .08 to .23 percent. The bonding agent shall be electroplated nickel. After electroplating, the steel shall be stress relieved to remove any hydrogen embrittlement.

Performance. The abrasive strips shall show no signs of failure such as breaks or cracks, when subjected to a bend test around a 1/16" round pin or wire.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:01Z
Check the source to verify that this is the current version before use.



NOTES:

1. DIMENSIONS ARE IN INCHES
2. UNCOATED STRIP THICKNESS .0020 TO .0030
3. COATED STRIP THICKNESS .0060 TO .0070

WIDTH OF STRIPS

Size	Width
Narrow	.09 to .10
Medium	.12 to .13
Wide	.15 to .16

Figure 1. Strip Assortment, Abrasive, Dental (Aluminum Oxide, Metal-Backed)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Workmanship. The abrasive strips shall be flat and uniform. The abrasive strips shall be free of defects, such as cracks, missing or incomplete coating, which detract from their appearance or impair their serviceability.

Unit. Box (BX). One box containing 12 strips as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one box of strips. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

Stock Number	Item Identification
6520-00-558-2500	STRIP ASSORTMENT, ABRASIVE, DENTAL, Aluminum Oxide.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:01Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA - DM

CIVIL AGENCY COORDINATING ACTIVITY
VA-OSS
PHS
FDA-MPQAS

Project No. 6520-1972

48180

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.