A-A-0051840(DM)
4 March 1987

# INTERIM COMMERCIAL ITEM DESCRIPTION

## SCISSORS, EAR, SERRATED (Joseph)

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers Joseph style serrated ear scissors used for ear surgery.

Salient characteristics.

Material. The scissors shall be fabricated from Class 4 martensitic corrosion-resistant steel in accordance with ASTM F899.

Hardness. After heat treatment, the finished instrument shall have a Rockwell hardness of C48 to C53 in accordance with ASTM E18 or E92.

Characteristics. Shall be a straight or curved ear scissors with serrated blades used in ear surgery procedures. The scissors shall be medium weight and 5-3/4 inches long. Both tips shall be sharp. The scissors shall be in accordance with DPSC Drawing 23390.

Finish. The finish of the scissors shall be mirror or mirror and satin. The finish shall be uniform and free from cracks, grind or tool marks, burrs, and dirt or polishing materials.

Identification marking. The scissors shall be marked with the name or registered trademark of the contractor and the word "stainless." In addition, instruments of foreign manufacture shall bear the name of the country of manufacture.

Corrosion-resistance. When tested as specified in Federal Specification GG-I-526, the scissors shall show no sign of corrosion or copper plating.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Performance. The scissors shall not bind or snag the test material when tested as specified herein. Each cut shall be made using the most distal seventy-five percent of the blade.

Workmanship. The scissors shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One scissors, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Performance test. The scissors shall be tested in accordance with the scissors performance test in Federal Specification GG-I-526. The test material shall be medium rubber dam in accordance with Federal Specification ZZ-R-690 (See Note 2).



Source: https://assist.dla.mil -- Downloaded: 202610-26T09:01Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The inspection level shall be S-2. Dimensional testing shall be at a 1.5% AQL, hardness testing at a 1.5% AQL, and corrosion and performance testing at a 1.0% AQL. The unit of product for test purposes shall be one scissors.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

NOTE 1: The following National Stock Number are covered by this document:

| NSN | ITEM IDENTIFICATION |
|---|---|
| 6515-01-111-1013 | SCISSORS, EAR, SERRATED, Joseph, Curved Blades |
| 6515-01-112-6408 | SCISSORS, EAR, SERRATED, Joseph, Straight Blades |

NOTE 2: The test material, medium size dental rubber dam, is commercially available from the HYGENIC CORPORATION, AKRON, OHIO.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4022



Source: https://assist.dla.mil -- Downloaded: 202610-26T09:01Z
Check the source to verify that this is the current version before use.