A-A-51842
2 March 1987

COMMERCIAL ITEM DESCRIPTION

BOTTLE, SCREWCAP

The Commercial Item Description shall be used in preference to Federal Specification NNN-B-001618.

This commercial item description covers one type of narrow mouth, plastic, screwcap bottle, suitable for storing laboratory solutions and corrosive liquids and as a shipping container for animal specimens.

Salient characteristics:

Shall be a narrow mouth polyethylene or polypropylene bottle of cylindrical shape. Shall be of natural or amber color. Amber bottles shall meet the requirements for light resistance as currently stated in the United States Pharmacopeia (U.S.P.).

The bottles shall be of the following sizes: 4, 8, 16 and 32 ounces. The normal capacity point of the bottle shall not extend above the bottom level of the neck. The fill of the bottle shall be within +1/2, -0 ounce of the capacity specified. Shall have a narrow mouth with continuous type conventionally used, thread finish. The mouth opening of the bottle shall be large enough to readily permit filling and dispensing of the contents. Shall be symetrical in shape and the walls shall be of uniform thickness. When placed on a plane smooth surface the bottles shall be stable.

The bottle shall be fitted with a one-piece molded polypropylene, screwcap. The screwcap shall be threaded with a continuous type thread. Shall make a close fit with the finish of the bottle forming an effective seal. With the screwcap applied, hand tight, there shall be no leakage. To provide for firm gripping for removing and tightening of the screwcap, the outside vertical surface of the screwcap shall have ribs. The bottle shall be furnished with the screwcap in place and shall be clean, dry and ready for use.

Workmanship. The bottles and cap shall be free from defects which detract from their appearance or impair their serviceability.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

Unit. Package (PG). One package containing twenty-four 32 oz. bottles, forty-eight 16 oz. bottles, seventy-two 4 oz. bottles or seventy-two 8 oz. natural or amber colored bottles, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more tests.

Sampling for tests. Sampling for tests shall be conducted in accordance with commercial practice.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:01Z
Check the source to verify that this is the currentversion before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTES: The following National Stock Numbers are covered by this document:

6640-00-130-3237	8 oz.	(Natural color)
6640-00-935-4064	4 oz	(Amber color)
6640-00-935-4069	8 oz.	(Amber color)
6640-00-935-4070	16 oz.	(Amber color)
6640-00-935-4071	32 oz.	(Amber color)

Ordering data. (Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:01Z
Check the source to verify that this is the current version before use.

A-A-51842

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS
FDA-MPQAS

Project No. 6640-1392

Location: NCR Disk #3

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:01Z
Check the source to verify that this is the current version before use.