FEDERAL SPECIFICATION

DISPENSER, SOAP AND DISPENSER, SURGICAL DETERGENT

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 Scope.—This specification covers wall-mounted, foot-operated, surgical detergent and liquid surgical soap dispensers.

1.2 Classification.

1.2.1 *Types.*—Dispensers shall be of the following types, as specified (see 6.2):

Type I. —Surgical detergent dispenser.
Type II.—Liquid surgical soap dispenser.

2. APPLICABLE SPECIFICATIONS AND STANDARDS

2.1 The following specifications and standards, of the issues in effect on date of invitation for bids, form a part of this specification:

Federal Specifications:

QQ–B–621 — Brass; Castings, Leaded Yellow.
QQ–S–766 — Steel, Corrosion-Resisting, Plates, Sheets, Strips, and Structural Shapes.
PPP–B–621—Boxes, Wood, Nailed and Lock-Corner.
PPP–B–636—Boxes, Fiber.

Federal Standards:

Fed. Std. No. 102—Preservation, Packaging, and Packing Levels.
Fed. Std. No. 123—Marking for Domestic Shipment (Civilian Agencies).
Fed. Test Method Std. No. 151—Metals; Test Methods.

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications, Standards, and Handbooks and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications, Standards, and Handbooks from established distribution points in their agencies.)

Military Specifications:

MIL–L–10547 — Liners, Case, Waterproof.

Military Standards:

MIL–STD–105 — Sampling Procedures and Tables for Inspection by Attributes.
MIL–STD–129—Marking for Shipment and Storage.

(Copies of Military Specifications and Standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material.—Material shall be as listed hereinunder and shall conform to the chemical requirements of the applicable specifications, when tested as specified in 4.4.1.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:46Z
Check the source to verify that this is the current version before use.

Type of dispenser	Part	Material	Specification
I	Nozzle	Corrosion-resisting steel	QQ–S–766, class 5
	Metal components of metering valve	Corrosion-resisting steel	QQ–S–766, class 5
	Mounting bracket, lock spring, nozzle setscrew	Corrosion-resisting steel	QQ–S–766, class 1 or 2
	Bowl and cover, valve body and locknut	White cellulose acetate butyrate	
	Tubing	Transparent, flexible, plasticized polyvinyl chloride	
II	Dispenser assembly	Brass	QQ–B–621, class A, B, or C
	Soap spout delivery assembly	Nickel silver or corrosion-resisting steel	QQ–S–766, class 1, 2, or 4

3.2 Construction.

3.2.1 *Type I.*—Type I surgical detergent dispenser shall consist of a cylindrically shaped, conical-bottomed plastic bowl, fitted with a metal dispensing nozzle which shall be connected internally by means of plastic tubing to a metering valve mounted at the bottom of the bowl. The metering valve shall be connected by plastic tubing to a foot pedal. The design of this combination shall be such that a single application of pressure on the foot pedal to the full extent of its travel will cause 2 cubic centimeters plus or minus 10 percent of surgical detergent (see 6.1) to be ejected from the nozzle. Release of pressure on the pedal shall allow it to resume its original height. Each application of pressure on the foot pedal shall cause another metered ejection of fluid from the nozzle. The bowl shall be fitted with a plastic cover.

3.2.1.1 *Bowl and cover.*—The bowl shall be molded of plastic (see 3.1), with the top portion having an inside diameter of 4¼ inches ± ⅛ inch and thickness of ⅛ inch ± 1/32 inch. The height of the bowl shall be such that it will hold, when filled to within ⅞ inch of the top and with the metering valve in place, not less than 30 ounces of fluid. A projection shall be molded on the back of the bowl, to permit mounting into a tapered wall bracket. On the front of the bowl at the top, there shall be molded a boss, through which a hole shall be drilled, with a center ⅜ inch below the top, to accommodate insertion of the nozzle. The nozzle shall be held in place by a corrosion-resisting steel setscrew. The cover shall be molded of plastic (see 3.1), and shall have a flange not less than ¼ inch deep, which shall seat inside the rim of the bowl with a suitable clearance to permit the cover to be easily removed. An integral plastic projection shall mate with a hole of suitable size to position the cover.

3.2.1.2 *Mounting bracket.*—The mounting bracket shall be designed to receive the tapered projection on the back of the bowl. The bracket shall be fabricated of 0.031-inch (No. 22, U.S. Standard gage) corrosion-resisting steel, and shall be fitted with a corrosion-resisting steel lock spring securely spotwelded into place. The lock spring shall prevent removal of the bowl unless the spring is deliberately depressed. The bracket shall have three holes to accommodate No. 6 roundhead wood screws which shall be furnished with the dispenser. When mounted on the wall, the screws shall be inaccessible when the bowl is in place.

3.2.1.3 *Nozzle.*—The nozzle shall consist of a piece of corrosion-resisting steel tubing, 0.137 inch inside diameter with a 0.025 inch wall. The tubing shall be bent down ⅞ inch

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:46Z
Check the source to verify that this is the current version before use.

± 1/8 inch at one end at a radius of 1/2 inch ± 1/8 inch to give an overall length of 6 1/2 inches ± 1/4 inch. The straight end shall be inserted in the bowl boss and held in place by the setscrew. The nozzle shall be capable of being moved in and out of the hole so that its projection from the front of the bowl can vary from 3 to 5 inches.

3.2.1.4 *Metering valve and connecting tube to nozzle.* — The metering valve shall be mounted in the bottom of the bowl, tight and leakproof, with a threaded projection extending through an opening in the bottom of the bowl and a connection for the plastic tubing leading to the foot pedal. The body of the valve shall be molded of white plastic (see 3.1), and shall be so designed as to contain a diaphragm of neoprene-coated cloth, two ball-type check valves, and a restoring spring. The threaded projection of the valve body shall extend through the bottom of the bowl and shall be held in place by means of a locknut. The resultant connection shall be leakproof. The plastic tubing leading to the foot pedal shall be attached to the valve stem projection and shall be held securely in place by means of a latex rubber compression ring. The tube connection shall be capable of withstanding the test specified in 4.4.3 without breaking or harmfully affecting the connection in any manner. The locknut shall be of white plastic (see 3.1). A central hole in the threaded projection of the valve shall lead to a chamber below the diaphragm, so that when pressure is applied by action of the foot pedal, the diaphragm is displaced, causing fluid to flow from the chamber above the diaphragm through an outlet hub on the valve body. This outlet hub shall be connected to the nozzle tube with a suitable length of transparent flexible plastic tubing (see 3.1). When pressure from the foot pedal is released, the spring shall restore the diaphragm to normal position, drawing fluid detergent into the chamber above the diaphragm through a port in the valve body. A ball check valve in this port shall prevent the reverse flow of fluid, when pressure is increased below the diaphragm. A second ball check valve in the outlet hub shall seal the connection to the nozzle, when fluid is drawn into the upper chamber by the return of the diaphragm to its normal position. The dimensions of the chambers and movement of the diaphragm shall be so limited as to cause the amount of fluid ejected in one operation to be 2 cubic centimeters plus or minus 10 percent. All metal parts of the valve shall be of corrosion-resisting steel (see 3.1). The metering valve shall show no leakage, when tested as specified in 4.4.2.2.

3.2.1.5 *Foot pedal and tubing.*—The pedal shall be molded of pure black rubber, approximately 3 inches wide by 5 1/2 inches long. Molded into the lower part of the pedal shall be a metal plate to provide rigidity for the bottom surface of the pedal. The pedal shall contain a cavity whose volume will be decreased 2 to 3 cubic inches when the top surface is fully depressed by the operator's foot. Into the rear of the pedal there shall be two holes extending into the cavity of the pedal. A hollow aluminum adaptor, one end of which shall be designed to make an interference fit with the foot pedal, shall be pressed into one hole. To this adaptor, the plastic tubing shall be connected by means of a latex rubber compression ring capable of withstanding the test specified in 4.4.3 without breaking or harmfully affecting the connection in any manner. The pump shall contain a rubber check valve designed to permit air to enter the pedal cavity on the return stroke. The tubing connecting the foot pedal to the metering valve shall be of transparent plastic (see 3.1), 6 feet long. The foot pedal, when tested as specified in 4.4.2.1, shall maintain an internal pressure of not less than 5 pounds per square inch for at least 2 seconds.

3.2.1.6 *Plastic tubing.*—The plastic tubing connecting the metering valve and nozzle, and metering valve and foot pedal, shall be transparent, flexible, plasticized polyvinyl chloride tubing, 3/16 inch inside diameter by 1/16 inch wall.

3.2.2 *Type II.* — Type II, liquid surgical soap dispenser, shall consist of a wall bracket

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:46Z
Check the source to verify that this is the current version before use.

dispensing assembly with container, connecting tube, and foot-operated pneumatic pump.

3.2.2.1 *Dispenser assembly.* — The dispenser assembly shall be a heavy, sturdy, brass casting, with integral bracket for wall mounting, and threaded cap for attachment of glass container. The threaded cast nipple for attachment of dispensing tube and container cap shall have machine-cut threads. A rubber washer shall be provided to effect an airtight seal between the glass container and cap. The dispensing tube shall have a knurled sliding nut for the attachment of the tube to cap. A spring locknut assembly shall fasten nut to tubing. The nut shall be machined to form a gasket recess and thread into the cap to act as combination spout swivel device and filler plug. The external side tube shall be bent almost horizontal and end formed to a vertical spout. The container end of the tube shall be provided with a suitable nondrip valve. The bracket casting shall be provided with an air pressure nipple for rubber tubing and an air intake valve. An adjustable control valve shall be provided for controlling the flow of liquid soap. It shall be adjustable to a range of flow of a few drops to a minimum of ½ fluid ounce. All parts of the dispenser in contact with soap shall be constructed of nickel silver or corrosion-resisting steel. A glass one-quart-capacity soap container jar, with threaded neck properly fitting threaded cap, shall be provided.

3.2.2.2 *Foot control pump.*—Foot control pump shall consist of a heavy brass casting with circular flanged base and screwcap forming a cylindrical container for housing the pneumatic pump unit. Base of casting shall be provided with nipple for connection to rubber tubing. Screw thread shall be machine cut. Inside of screwcap shall have a flanged hole machined to receive a plunger for operating the pneumatic pump. Base of casting shall be provided with a rubber ring to prevent slipping and increase stability of unit. Foot pump assembly shall be rugged in design, and weigh not less than 3⅛ pounds.

3.2.2.2.1 *Alternate foot control pump.*—Alternate foot control pump shall be round, and shall consist of a molded, nonmarking, rubber bellows, with a ridged flange on the bottom side to prevent slipping, and an internal cast-iron insert, resting on the rubber flange of the bellows. The cast-iron insert shall be suitably coated on the bottom side with a corrosion-preventing material, and shall be equipped with a hose connection which shall extend through a hole in the side of the rubber bellows. The pump assembly shall be approximately 4¾ inches in diameter and no greater than 2$^{1}/_{16}$ inches in height. The pump shall be capable of being operated by users without the necessity of lifting their foot from the floor.

3.2.2.3 *Rubber tubing.*—The dispensing assembly shall be connected with the foot pump by means of ⅜-inch-outside-diameter flexible rubber tubing, approximately 5 feet long.

3.3 **Finish.**

3.3.1 *Type I.*—All corners and edges shall be free from sharp, ragged, or burred edges. The color of the bowl and cover shall be uniform and free from mottling, bubbles, cracks, and pitting. The cover shall fit snugly in the bowl and give the appearance of one unit to the eye. The nozzle shall have a polished finish.

3.3.2 *Type II.*—All brass castings shall be heavily chromium plated over nickel. All parts in contact with soap shall be nickel silver or corrosion-resisting steel. Underside of dispensing assembly cap shall be heavily nickel silver plated where there is possibility of contact with soap. All exterior exposed metal surfaces shall be polished to a mirror finish.

3.4 **Identification marking.** — The inside surface of each cover of type I and each dispenser assembly of type II shall have the name or registered trade-mark of the manufacturer permanently molded thereon. Mark-

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:46Z
Check the source to verify that this is the current version before use.

ing shall be legible, distinct, clear, and permanent. Decalcomanias shall not be employed.

3.5 **Service data.**—The contractor shall enclose with each unit two copies each of one or more booklets which contain complete instructions for installation, operation, maintenance, and a list of component parts. Each part of the unit, other than common fasteners, etc., shall be identified in the service data and parts list by the manufacturer's part number.

3.5.1 For Military Medical Supply Agency procurement.—The contractor shall forward six additional copies of the service data required in 3.5 to the contracting officer for distribution.

3.6 **Workmanship.**—Workmanship shall be first class throughout. Dispensers shall be free from defects which detract from their appearance or may impair their serviceability.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 The provisions contained in this section are applicable to inspection performed by and for the Government. Sampling, inspection, and test procedures for components or materials described shall be in accordance with this specification and with specifications and standards referenced herein to the extent applicable, except that this specification shall govern in the event of conflict.

4.1.1 *Supplier responsibilities for inspection for military procurement.*—Such examinations and tests as are set forth in this specification or as shall otherwise be appropriate or necessary to insure that supplies conform to specification requirements, shall be performed by and at the expense of the supplier. Suppliers who do not have facilities adequate for such tests shall arrange for the use of test facilities acceptable to the Government. Records of examinations and tests performed by the supplier shall be maintained by the supplier and made available to the Government, upon the Government's request, at any time, or from time to time, for a period of 2 years after delivery of the supplies to which such records relate.

4.1.2 Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and the test results obtained, and indicating that material is as specified, may be furnished.

4.2 **Sampling.** — Sampling shall be conducted in accordance with Military Standard MIL–STD–105 and Appendix thereto.

	Inspection level	AQL (percent defective)	Minimum sample size
For visual examination:			
Major defects	II	1.0	
Total defects	II	4.0	
For dimensional examination	L5	2.5	
For tests	L5	*	3

* Acceptance number shall be zero for all sample sizes.

4.2.1 *Raw materials.*—Four ounces of each lot of each metal used in the construction of the dispensers shall be selected for chemical analysis.

4.2.2 Where possible, the same sample shall be used for the determination of two or more tests.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:46Z
Check the source to verify that this is the current version before use.

4.3 Inspection.—The dispensers shall be inspected to determine compliance with all requirements contained in sections 3 and 5 of this specification. Nonconformance with these requirements will be permitted to the extent indicated in 4.2.

4.3.1 *Classification of defects.*—Inspection shall be conducted in accordance with the following classification of defects:

*Categories and defects****

Major

101. Dispenser not complete with all component parts specified.
102. Mating parts fail to fit or function properly.
103. Component parts not free from cracks, tears, cuts, chipped and deformed areas.
104. Exposed metal component parts not free from burs or sharpness.

Minor

201. Assembled parts not securely fastened.
202. Plated components not free of blisters and bare or unplated areas.
203. Plated exterior metal surfaces not mirror finish.
204. Polished metal surfaces not free of excessively scratched or unpolished areas.
205. Identification marking not present, permanent, and legible.
206. Service data not included in duplicate with each unit.
207. Labels or printed data not legible and free from defacing marks, scratches, and tears.

** Inspection shall not be restricted to the classified possible defects listed above.

4.3.2 *Dimensional examination.*—The dispensers shall be inspected for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.4 Tests.—Tests shall be conducted in accordance with the applicable specifications, as necessary, and as hereinafter specified, to determine compliance with this specification.

4.4.1 *Chemical composition.* — Tests for chemical analysis shall be conducted in accordance with method 111 or 112 of Federal Test Method Standard 151.

4.4.2 *Leakage.*

4.4.2.1 *Foot pedal.*—The foot pedal assembly specified in 3.2.1.5 shall be compressed with a force of 125 pounds.

4.4.2.2 *Metering valve.* — The metering valve assembly specified in 3.2.1.4 shall be immersed in water and the lower chamber shall be subjected to 10 pounds per square inch air pressure.

4.4.3 *Tube connection.*—The tube connections at the metering valve and foot pedal specified in 3.2.1.4 and 3.2.1.5 shall be subjected to a direct pull on the plastic tubing of 3 pounds.

4.4.4 *Delivery volume test.*—Using water as a test medium, the delivered volume shall be collected in a 25-milliliter graduate from ten consecutive strokes of the foot pedal. The total delivered volume shall be divided by 10 to determine compliance with the metering requirement. For type II dispensers, the delivery volume of the adjustable control valve shall be similarly determined.

5. PREPARATION FOR DELIVERY

Non-Military.—The levels of preservation, packaging and packing used are defined in Federal Standard No. 102.

5.1 Packaging.

5.1.1 *Unit of issue.* — One dispenser, as specified, constitutes one unit.

5.1.2 *Packaging quantities.*—The number of units of issue indicated in table I shall be packaged in each unit and exterior container, as applicable, for the required level of protection specified in the procurement document.

5.1.2.1 *Packing variation permitted.* — If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:46Z
Check the source to verify that this is the current version before use.

TABLE I.—*Packaging quantities*

Federal stock number	Unit package	Intermediate package	Exterior container
6530–299–8114	1 unit	Not required	24 units
6530–779–2825	1 unit	Not required	6 units

units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.1.3 *Level A.*

5.1.3.1 *Unit package.*—Each unit shall be disassembled as far as is practicable. Tubing shall be arranged in a coil of such diameter that it will not kink or take a permanent set. Each complete unit shall be packaged in a fiberboard carton of appropriate size and design, constructed in accordance with Federal Specification PPP–B–636. Carton shall be designed with suitable interior fittings or shall contain sufficient cushioning material to insure against shifting of, or damage to, the contents. Each carton shall be adequately secured to prevent accidental opening.

5.1.4 *Level C.*—Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose within the continental limits of the United States of America.

5.2 Packing.

5.2.1 *Level A.*

5.2.1.1 *Exterior container.*—Exterior container shall be a box constructed in accordance with Federal Specification PPP–B–621, table IV.

5.2.1.2 *Case liner.*—Each level A box shall be lined with a waterproof case liner conforming to Military Specification MIL–L–10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto.

5.2.1.3 *Strapping.* — Strapping, when required, shall be in accordance with applicable box specification.

Note.—Strapping will not be required for shipment forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 *Level B.*

5.2.2.1 *Exterior container.*—Exterior container shall be a box constructed in accordance with Federal Specification PPP–B–621, table II.

5.2.3 *Level C.* — The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 *Unit package (Military agencies).*—Each unit package shall be marked in accordance with section 5 of Military Standard MIL–STD–129.

5.3.2 *Shipping containers (Military agencies).*—Exterior shipping containers shall be marked in accordance with Military Standard MIL–STD–129, except that 6.2(c), 6.2(d), and 6.5 shall not be applicable. Service color and service symbol markings are required only on direct shipments from the contractor's plant to overseas activities.

5.3.3 *Shipping containers (Civil agencies).*—In addition to any special marking shown in the contract or order, exterior shipping containers shall be marked in accordance with Federal Standard No. 123.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:46Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Intended use.—Type I, surgical detergent dispenser, shall be suitable to dispense Federal Stock Number 6505–116–1750, Detergent, Surgical, 1 gal. Type II, liquid surgical soap dispenser, shall be suitable to dispense Federal Stock Number 6505–141–1950, Soap, Surgical, Liquid, 1 pt. and 6505–141–1960, Soap, Surgical, Liquid, 1 gal.

6.2 Ordering data.—Purchasers should exercise any desired options offered herein, and procurement documents should specify the following:

(*a*) Title, number, and date of this specification.

(*b*) Type required (see 1.2.1).

(*c*) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification covers the following items appearing in the Armed Services Medical Stock List:

6530–299–8114 — DISPENSER, SURGICAL DETERGENT, Foot Operated.

6530–779–2825 — DISPENSER, SOAP, Foot Operated.

6.4 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of the specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Federal Government requirements.

6.5 Transportation description.—Transportation description applicable to this item is:

Dispenser, Bathroom Fixture, Other than China or Earthenware.
Carload minimum weight 30,000 pounds.

Dispenser, Bathroom Fixture, Not Otherwise Indexed, Other than China, Porcelain, or Plated Steel.
Motor volume minimum weight 30,000 pounds.

Notice. — When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

MILITARY INTERESTS:
Army—M
Navy—Md
Air Force.

☆U.S. GOVERNMENT PRINTING OFFICE: 1974–713-150/3168

Copies of this specification may be purchased for 5 cents each.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:46Z
Check the source to verify that this is the current version before use.