METRIC

A-A-54791
30 January 1993

COMMERCIAL ITEM DESCRIPTION

GLOVES, SURGEONS', Latex Rubber, Pre-Powdered, Disposable, Sterile

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers sterile, disposable, latex rubber surgeons' gloves suitable for use in surgical procedures.

Salient Characteristics:

General. "Gloves", as used herein, shall refer to all surgeons' gloves covered by the scope of this document. Gloves shall be hand-specific: one glove shall fit a user's right hand; the other shall fit a user's left hand. Glove sizes shall be limited to the following: 6, 6-1/2, 7, 7-1/2, 8, and 8-1/2.

Dimensions. Gloves shall comply with the dimensional requirements specified in ASTM D3577. The dimensions of glove samples shall be inspected in accordance with ASTM D3577.

Material. Gloves shall be made of latex rubber. This material shall comply with the requirements for one of the three types of latex rubber defined in ASTM D1076. Material samples shall be cut from finished gloves and subjected to all inspections specified by ASTM D1076.

Performance. Gloves shall be leakproof. Sample glove pairs shall be inspected for leaks in accordance with ASTM D5151-92.

Physical Properties. Gloves shall comply with the Type I physical property requirements specified in ASTM D3577. The physical properties of glove samples shall be inspected in accordance with ASTM D3577.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Finish. Gloves shall have a smooth, uniform finish. Gloves shall be pre-powdered with a USP-grade absorbable dusting powder. Gloves shall not be powdered with talcum powder. Glove samples shall be inspected for finish. Glove powder samples shall be inspected for compliance with USP requirements (See Note 1).

Marking. Glove size shall be legibly and permanently marked on the exterior of the glove cuff. Glove samples shall be inspected for size markings.

Workmanship. Gloves shall be free from blooming, blisters, fish-eyes, frosting, shorts, wrinkles, foreign matter, visible holes, missing fingers, missing thumbs, fingers that are fused together, thumbs that are fused together, and any other defects that detract from their appearance or impair their serviceability. The definitions of the first seven defect terms shall be as specified in ASTM D883. Glove samples shall be inspected for workmanship.

Sterility. Gloves shall be sterile. The sterility assurance level shall be no less than 10^{-6}. Glove samples shall be inspected for sterility in accordance with the latest edition of the USP.

Sterile Packaging. Gloves shall be sterile-packed in pairs. A pair shall consist of one right-handed glove and one left-handed glove.

Each pair of gloves in a unit package shall be sealed in a leak-proof, peel-open type packet designed to preserve the sterility of its contents until opened or damaged (See "Sterility"). Packets shall be assembled in the following order: first, a glove pair shall be enclosed in a wrapper; second, the wrapped glove pair shall be sealed in an overwrap. Wrappers shall be made of paper or other suitable material. Overwraps shall be made of resin-impregnated paper, clear plastic, translucent plastic, or other suitable material. Overwraps shall be sealed by means of heat-sealing, cold-sealing, or any other method that produces a leak-proof seal. Sample packets shall be inspected for leaks in accordance with ASTM D3078. Sample packets shall be examined for compliance with all other requirements listed above.

Packets shall not generate loose fibers or other debris when opened. Sample packets shall be inspected for generation of fibers and debris.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Unit. Package (PG). One package containing 50 sterile-packed glove pairs as specified, constitutes one unit. The unit package shall be a top-opening and dispensing container.

Quality Assurance Provisions.

Responsibility for Inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of inspections performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, shall be defined as the examination and testing of an item. Examination shall be defined as unaided visual, auditory, tactile, or olfactory investigation. Testing shall be defined as the use of laboratory equipment and procedures to determine various properties of an item. Inspections shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. For inspection purposes, the unit of product shall be one pair of gloves.

Sampling for Inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Nine samplings shall be performed - one for each salient characteristic specified herein. For each sampling, inspection level S-2 shall be used; the Acceptable Quality Level, (AQL, shall be 0.65 percent.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor Certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory Requirements.

Federal Food, Drug, and Cosmetic Act. If the products covered by this document have been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered Materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering Data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract or order.

Note 1: USP requirements for glove powder are given in the monograph entitled "Absorbable Dusting Powder".

Note 2: This document covers the following National Stock Numbers and their respective Item Identifications:

NATIONAL STOCK NUMBERS	ITEM IDENTIFICATIONS
6515-01-149-8839	GLOVES, SURGEONS', Latex Rubber, Size 6-1/2, Pre-Powered, Disposable, Sterile, 50 Pairs
6515-01-149-8840	GLOVES,SURGEONS', Latex Rubber, Size 7, Pre-Powdered, Disposable, Sterile, 50 Pairs
6515-01-149-8841	GLOVES, SURGEONS', Latex Rubber, Size 7-1/2, Pre-Powdered, Disposable, Sterile, 50 Pairs
6515-01-149-8842	GLOVES, SURGEONS', Latex Rubber, Size 8, Pre-Powdered, Disposable, Sterile, 50 Pairs
6515-01-149-8843	GLOVES, SURGEONS', Latex Rubber, Size 8-1/2, Pre-Powdered, Disposable, Sterile, 50 Pairs
6515-01-151-1790	GLOVES, SURGEONS', Latex Rubber, Size 6, Pre-Powdered, Disposable, Sterile, 50 Pairs

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

A-A-54791

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6515-5469
Location: c:\word\komada\s1-glove

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given

2 The submitter of this form must complete blocks 4, 5, 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER A-A-54791	2 DOCUMENT DATE (YYMMDD) 30 January 1993

3 DOCUMENT TITLE
GLOVES, SURGEONS", Latex Rubber, Pre-Powered, Disposable, Sterile

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed)

5 REASON FOR RECOMMENDATION

6 SUBMITTER	
a NAME	b ORGANIZATION
c ADDRESS (Include Zip Code)	d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)
	e DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY	
a NAME Defense Personnel Support Center ATTN: DPSC-MSE	b TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (215)7372870 444-2870
c ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.