INCH-POUND

A-A-54794
15 January 1993

COMMERCIAL ITEM DESCRIPTION

HUMIDIFIER, OXYGEN

Diffuser Type

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for MIL-H-36746A, which is cancelled.

This Commercial Item Description covers a sterilizable oxygen humidifier for use in open systems (oxygen mask, tent, etc.).

Salient Characteristics:

The humidifier shall be dual purpose, a bubbler humidifier and a jet type nebulizer, which generates oxygen (or air) into a fog for use in an open system (oxygen face mask, tent, etc.).

Except for the reservoir and diffuser, the complete unit shall be capable of repeated sterilization, 250°F, 20 minutes. The complete unit shall be sterilizable and reusable.

The system shall be designed to remain operational when it is tilted up to 20°.

The reservoir bottle shall have a working capacity of at least 200 cc. It shall be fabricated of suitable transparent plastic permitting visual observation of the liquid level within the bottle. A high and low liquid level shall be clearly marked and identified on the bottle. There shall be approximately 200 cc capacity between the high and low liquid levels. The bottle shall be capable of withstanding an internal pressure of at least 20 psig without bursting.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

The humidifier shall be capable of maintaining a relative humidity at a body temperature as follows:

a. As a bubble type humidifier, with oxygen flowing down through a syphon tube and out a diffuser at the bottom of the reservoir gaining humidity rising through the water to the outlet, from 35 percent relative humidity at 3 Lpm to 20 percent relative humidity at 12 Lpm.

b. As a jet type nebulizer, with water or medicated solution being drawn up the syphone tube through an oxygen mixing orifice to the outlet, from 40 percent relative humidity at 3 Lpm to 67 percent relative humidity at 10 Lpm.

The relative humidities shall be calculated by determining the weight of water expended (with oxygen admitted to the humidifier at 70°F - 73°F and 29 inches of mercury) and the resultant molal humidity of issuant gas. The vapor pressure of the issuant gas divided by its partial pressure of saturated vapor at 98.6°F, shall be defined as the relative humidity.

The humidifier shall be provided with a removable, leakproof top. The top shall have:

* an inlet connector, winged threaded DISS 1240 female swivel, 9/16 inch threaded, for 50 psig oxygen source

* a relief valve

* a selector switch to select bubble or jet operation, and

* an outlet connector, 22 mm male, in accordance with ASTM F 1054.

A leakproof barbed adapter for the outlet shall be provided which allows connection of 3/16 to 5/16 inch tubing.

The pressure relief valve shall prevent internal pressure inside the humidifier from exceeding 2.5 psig. A warning whistle or other audible warning shall be provided which shall operate in conjunction with the pressure relief valve so designed to provide audible sound if the tubing becomes obstructed when the oxygen flow is as low as 3 Lpm, but shall not sound an alarm when a free flow is being used through an unobstructed, 48 inch long delivery tube and catheter.

The humidifier shall be supplied with operating and maintenance instructions.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Workmanship. The oxygen humidifier shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA.) One humidifier, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-1. The unit of product for examination purposes shall be one humidifier.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one humidifier.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

6515-00-469-6573

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA - OSS
USPHS
FDA-MPQAS

Project Number: 6515-5374

Location: HUMIDC12.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-54794	2. DOCUMENT DATE (YYMMDD): 15 January 1993

3. DOCUMENT TITLE: HUMIDIFIER, OXYGEN, Diffuser Type

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 98/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.