INCH-POUND

A-A-54800
16 April 1993

COMMERCIAL ITEM DESCRIPTION

CUTTER, BONE PIN, 18 INCH

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-C-36676(1) which is cancelled.

This Commercial Item Description covers a cutter that is used to cut Steinmann and Kirschner type bone pins.

Salient characteristics:

General. The bone pin cutter shall be 18 inches in length, of the Steinmann type, triple action design for additional leverage and suitable for use in cutting bone pins of the Steinmann and Kirschner type and wire.

Hardness. The bone pin cutter shall have hardness values, as shown below:

Class/Type of Material	Rockwell Hardness
Type 17-4 PH stainless	C35 to C45
Class D2 Tool steel	C59 to C64
Type 440A Martensitic stainless	C50 to C56
Type 440C Martensitic stainless	C59 to C64
Type 302 Austenitic stainless	B75 to B85
Type 303 Austenitic stainless	B75 to B85
Type 304 Austenitic stainless	B75 to B85

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Style, design, material and dimensions. The bone pin cutter shall be in accordance with Defense Personnel Support Center Drawing (DPSC) 24230 for style, design, material and dimensions.

Finish. The face of the cutting jaws (components 5 and 6 on drawing, and both surfaces of the top and bottom aligning plates (components 7 and 9 on drawing) shall have a surface finish equal to, or superior to, an average value of 8 microinches (MU) roughness height in accordance with ANSI Standard B46.1. Other components of the cutter shall be finished as shown on DPSC drawing 24230.

Passivation The bone pin cutter shall be properly passivated.

Corrosion resistance. The bone pin cutter shall show no evidence of corrosion plating or copper when tested for corrosion resistance in accordance with ASTM F1089

Identification marking. The bone pin cutter shall be legibly marked with contractor's name or registered trademark, and the work "stainless". Foreign manufactured bone pin cutters shall bear the name of the country of origin.

Workmanship The bone pin cutter shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One bone pin cutter, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.5 (percent defective) and an inspection level of S-4. The unit of product for test purposes shall be one bone pin cutter.

Tests Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests Sampling for tests (dimensional) shall be conducted in accordance with MIL-STD-105, with an AQL of 1.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one bone pin cutter.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Contractor Certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors'suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

NSN	Item Identification
6515-00-323-5380	CUTTER, BONE-PIN, 18 Inch.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6515-5225

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER [illegible]-54800	2. DOCUMENT DATE (YYMMDD) [illegible] April [illegible]90

3. DOCUMENT TITLE
CUTTER, BONE PIN, 18 INCH

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	c. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON
c. ADDRESS (Include Zip Code)	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 Previous editions are obsolete. 98/90

Source: https://assist.dla.mil - Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.