INCH-POUND

A-A-54811
12 February 1993

COMMERCIAL ITEM DESCRIPTION

SYRINGE AND NEEDLE, HYPODERMIC, Disposable, Allergy, Detachable Needle

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers sterile, disposable, hypodermic syringes with a detachable needle.

Salient characteristics:

General. The syringe shall be a sterile 1 cc syringe with detachable needle suitable for diagnostic and/or therapeutic treatment of allergy patients.

Material. The syringe barrel and plunger shall be fabricated from rigid polypropylene which meets the United States Pharmaceutical (USP) requirements for Class IV plastics. The interior of the barrel shall be lightly coated with an inert silicone lubricant. The plunger tip shall be inert and fabricated from rubber, neoprene, or thermoplastic elastomer, which meets the USP requirements for Elastomeric Closures for Injections. Plastic plunger tips shall meet the USP requirements for Class IV plastics.

The needle cannula shall be Type 304 stainless steel in accordance with American Society for Testing and Materials (ASTM) Designation F899. The needle hub shall be fabricated from rigid polypropylene that meets the USP requirements for Class IV plastics. The sheath shall be fabricated from rigid plastic.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Style, design and dimensions. The syringe shall be a Luer slip type in accordance with ISO 594. The syringe barrel shall be flanged to provide finger grip and anti-roll feature. The inside of the barrel shall have a smooth, uniform surface, free from pits, grind marks, bubbles, flash, or other imperfections. The plunger shall glide through the barrel while maintaining a fluid-tight fit. The plunger tip and barrel shall both be beveled and form a flush fit at zero volume. The barrel shall be transparent and the plunger tip shall have two visible rings touching the sides of the barrel.

The outside surface of the barrel shall be marked with main graduations at 0.05 cc intervals and subgraduations at 0.01 cc. The capacity 1.0 cc shall be marked at the proximal end of the barrel. The markings shall resist smearing or obliteration under normal conditions of use. The length of the scale shall be not less than 2-1/4 inches. Main graduation markings shall be at least 33 percent longer than subgraduations. Each 0.1 cc graduation shall also have the numerical volume indicated. The barrel shall also have the name or registered trademark of the contractor. Actual fluid volume shall be determined by the position of the distal ring of the plunger tip.

Needle. The needle shall be a 27 gauge, 1/2 inch ± 3/64 inch in length. The needle shall be a Luer-lock type in accordance with ISO 594. The needle shall make a mechanically stable, leakproof connection with the syringe. The needle shall have a plastic protective cover which is readily removable and which does not cause the needle to separate from the syringe when the cover is removed.

The cannula shall be coated with a medical grade silicone lubricant. The point shall be sharp to minimize patient trauma. The heel of the needle point shall be sufficiently dulled to prevent coring. There shall be no bent or dull points. The interior and exterior of the cannula shall be free from burs, slivers, or any foreign matter.

The needle style shall be as follows:

Gauge	Point length* (inch)	Outside Diam (inch)	Bevel (inch)
27	0.068 - .084	0.0155 - 0.0165	0.021 - 0.039

*Measured from heel to tip and parallel to longitudinal axis of cannula.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Needle bending resistance test. The needle cannula shall not crack or separate from the needle hub when bent 25 degrees from its rest position and subsequently bent 25 degrees into the opposite direction.

Accuracy test. The syringe shall deliver ± 5 percent of the capacity indicated at each main graduation mark greater than 0.2 cc and ± 0.007 cc at graduations of less than or equal to 0.2 cc when tested against a calibration burette. As an alternate, the determination of capacity test listed in ISO 7886 is acceptable.

Leakage test. The water pressure test system shall have a water pressure gauge and a rubber plug. The needle shall be secured to the Luer-tip of the syringe by applying it with a 5 pound axial force and while maintaining this compressive stress twisting the needle hub 45 degrees relative to the Luer-tip. The needle shall be inserted into the plug. The syringe shall be tested for liquid leakage at the syringe piston and syringe nozzle under compression and during aspiration in accordance with ISO 7886.

Sterility, pyrogenicity, toxicity. The syringe and needle shall be nonpyrogenic and nontoxic in accordance with USP. The syringe shall be sterile in accordance with USP or AAMI requirements. Sterility assurance level shall be 0.0001.

Needle assembly test. The needle cannula shall not crack or separate from the needle hub when subjected to a 5 pound tensile force for one minute (See Note 2).

Instructions. Each unit package shall be provided with instructions describing the proper techniques for dispensing the sterile syringes and needles, as applicable, from the containers and preparing them for use.

Workmanship. The syringe shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing one hundred individually packaged sterile syringes, as specified, constitutes one unit. The individual package shall be a suitable container capable of maintaining sterility of contents unless package is damaged or opened.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspection are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one syringe. Sampling shall be inspection level S-2 with an AQL of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:38Z
Check the source to verify that this is the current version before use.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Note 1: The following National Stock Number and its respective Item Identification is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-01-152-3612	SYRINGE AND NEEDLE, HYPODERMIC, Disposable, Allergy, 27 gauge, 1/2 inch needle, 1 cc syringe

Note 2: An alternate test, such as a pull-to-destruction test, is acceptable in lieu of the 5 pound separation test provided product found acceptable of the pull-to-destruction test also passes the 5 pound separation test.

Note 3: ANSI and ISO standard are available from the American National Standards Institute, Inc., 11 West 42nd Street, New York, NY 10036.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5503
Location: c:\word\doc\levin\syringe

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revisio letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE. This form may not be used to request copies of documents, nor to request waivers, or clarification o requirements on current contracts. Comments submitted on this form do not constitute or imply authorization tc waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1 DOCUMENT NUMBER	2 DOCUMENT DATE (YYMMDD)
	A-A-54811	12 February 1993

3 DOCUMENT TITLE

SYRINGE AND NEEDLE, HYPODERMIC, Disposable, Allergy, Detachable Needle

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a NAME (Last, First, Middle Initial)	b ORGANIZATION
c ADDRESS (Include Zip Code)	d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable) 7 DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME	b TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center ATTN: DPSC-MSE	(215)7372870	444-2870
ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403 Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete — 198/290

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:38Z
Check the source to verify that this is the current version before use.