INCH-POUND

A-A-54817
12 February 1993

# COMMERCIAL ITEM DESCRIPTION

## SCISSORS, ORAL SURGICAL

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-S-36456B, which is cancelled.

This commercial item description covers scissors suitable for use in oral surgical procedures.

Salient characteristics:

General. Scissors shall be of the following types:

Type 1 - Curved, Blunt-Blunt Tip, 4-3/4 inches
Type 2 - Angular, Dean, Sharp-Sharp Tip, 6-3/4 inches

Material. The material shall be Class 4 corrosion resistant steel in accordance with ASTM F899.

Hardness. Scissor blades shall have a Rockwell Hardness C scale (HRC) of 50 to 55 when tested in accordance with ASTM Designations E18 or E92 whichever is applicable. Conversion of hardness values shall be in accordance with ASTM Designation E140.

Style, design and dimensions. Scissors types shall be in accordance with Defense Personnel Support Center (DPSC) Drawing Number 20458 for style, design and dimensions.

Performance. Scissors shall be tested in accordance with scissors performance requirements of ASTM Designation F1079. Testing material shall be latex rubber sheet (0.007 to 0.009 inch thick) for lightweight scissors.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Construction. The scissors shall be of three piece construction consisting of two scissor sections locked together by a screw, which is permanently set in a fixed position. The screw lock and scissor sections shall be properly assembled to provide a cutting portion which shall be free of both excessive stiffness and looseness. The inside of the cutting portion of each scissors shall be ground to properly mate with the opposing scissor during cutting procedures. The cutting edges of the blades shall be properly sharpened.

Finish. Finish shall be entirely mirror, satin or matte, except the cutting edges and inside surfaces of the blade shall have a crocus finish. Should the instrument be of a mirror finish, the locks shall be polished smooth, but need not possess a mirror finish. The matte finish shall exhibit a uniform nondirectional surface texture equivalent to that produced by lightly blasting the prepared surface with a fine abrasive under pressure.

Corrosion resistance. All scissors shall meet the boil and copper sulfate tests in accordance with ASTM Designation F1089.

Identification markings. The scissors shall be marked with the word "stainless", the manufacturer, and the country of origin. As an alternate to the word "stainless", the use of a registered trade name identifying the steel as stainless shall be acceptable, provided prior approval is received from the cognizant officer. All markings shall be permanent and legible.

Workmanship. Scissors shall meet the workmanship, finish, and appearance requirements in ASTM Designation F1079 as applicable.

Unit. Each (EA). One pair of scissors of the type specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Government reserves the right to perform any of the inspections set forth in the specification where such inspection are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one scissor as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 2.5 except for dimensions which shall be conducted at inspection level S-1 with an AQL of 2.5.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one scissor. Sampling shall be inspection level S-1 with an AQL of 1.0, except for material and cutting performance, the acceptance number shall be zero for all sample sizes.

Chemical analysis. In the absence of contractor certification for the raw material, a chemical analysis shall be conducted. The chemical analysis shall be performed by reagent solutions, combustion methods, nonemission methods, or by spectrochemical analysis. Any ASTM method shall be acceptable.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers and their respective Item Identifications are covered by this document:

| NATIONAL STOCK NUMBER | ITEM IDENTIFICATION |
|---|---|
| 6520-00-554-9150 | SCISSORS, ORAL SURGICAL, Curved, 4-3/4 inch (12.06 cm) |
| 6520-00-555-1150 | SCISSORS, ORAL SURGICAL, Dean Angular, 6-3/4 inch (17.14 cm) |

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6520-2507
Location: c:\word\doc\blann\scisscid



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

## INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given

The submitter of this form must complete blocks 4, 5, 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE. This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

| I RECOMMEND A CHANGE: | 1 DOCUMENT NUMBER | 2 DOCUMENT DATE (YYMMDD) |
|---|---|---|
| | A-A-54817 | 12 February 1993 |

3 DOCUMENT TITLE

SCISSORS, ORAL SURGICAL

4 NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed )*

5 REASON FOR RECOMMENDATION

6. SUBMITTER

| a. NAME (Last, First, Middle Initial) | b. ORGANIZATION | |
|---|---|---|
| c. ADDRESS (Include Zip Code) | d. TELEPHONE (Include Area Code)<br>(1) Commercial<br>(2) AUTOVON (If applicable) | 7. DATE SUBMITTED (YYMMDD) |

8 PREPARING ACTIVITY

| a NAME | b TELEPHONE (Include Area Code) | |
|---|---|---|
| | (1) Commercial | (2) AUTOVON |
| Defense Personnel Support Center<br>ATTN: DPSC-MSE | (215)7372870 | 444-2870 |
| ADDRESS (Include Zip Code)<br>2800 South 20th Street<br>Philadelphia, PA 19145 | IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT<br>Defense Quality and Standardization Office<br>5203 Leesburg Pike Suite 1403 Falls Church VA 22041-3466<br>Telephone (703) 756-2340 AUTOVON 289-2340 | |

DD Form 1426, OCT 89 *Previous editions are obsolete* 198/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.