INCH-POUND

A-A-54825
5 March 1993

COMMERCIAL ITEM DESCRIPTION

BLOOD COLLECTING-DISPENSING BAG, Indirect Transfusion, Plastic, Disposable Type

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers a sterile, pyrogen-free, disposable, empty, plastic blood transfer unit consisting of an integrally attached coupling device, connecting tube, plastic bag with outlet tubes, and end protectors all sterilized as a unit.

Salient characteristics:

General. Shall be suitable for such procedures as aseptic subdivision of whole blood into small units, separation of plasma from cellular elements and preparation of platelet rich plasma or platelet concentrate. Shall be suitable for transferring blood from a blood bottle, transferring blood or components from the standard blood collecting bags, and for administering blood with standard blood recipient sets.

Material. The unit shall be fabricated of suitable, transparent, non-toxic, flexible, non-pyrogenic plastic. Material shall be compatible with the liquid contents. This flexibility shall be maintained at temperatures of 0°C to 50°C. The plastic shall not become tacky following sterilization or during storage for a minimum period of four years. The unit shall not deteriorate when held at -20°C, nor break down at +50°C.

The bag shall have a capacity of 300 ml. The shape of the bag shall permit centrifugation in a standard trunnion cup. It shall be provided with two outlet ports, each containing a diaphragm which can be readily punctured in order to couple and administer blood and/or its components through a standard blood recipient set. The entire exterior surface of the connecting tubes shall be sterile and have its sterility maintained by a tamper-proof closure. The outlet

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

port shall be so designed that the piercing device of the blood recipient set makes a leak-proof fluid path connection with the outlet port prior to puncture of the membrane.

There shall be an integral means of suspending the bag during administration of blood or blood components.

The coupling device shall have a formed ring at one end that functions as an automatic stop. The tip shall be designed with either a side opening or distal opening, and shall be of sufficient length to rupture the inlet seal on the existing blood bags; however, not of such length as to reach far enough into the bag as to perforate the side walls during insertion.

The connecting tube shall consist of a length of plastic tubing not less than 27 inches with an external diameter not more than 3/16 inch. Shall bear identical numbers and "X"s alternately and indelibly imprinted along the tubing for the purpose of dividing the tubing into segments for samples. The center-to-center distance shall be not less than 2-1/2 inches nor more than 3 inches. The connecting tube of each unit shall bear a different number. The final number shall be immediately adjacent to the bag so that at the time the tube is severed the identifying number remains with the bag. One end of the tubing shall be connected to the coupler device and the other to the bag.

Instructions for use. Detailed instructions shall be provided with or imprinted on the immediate container of each bag.

Pyrogenicity and toxicity. The bag and attachments shall be non-pyrogenic and non-toxic when tested in accordance with the United States Pharmacopiea (USP).

Biological tests. The system shall meet requirements of Class IV biological tests when tested in accordance with the USP.

Sterility. The bag and attachments shall be sterile in accordance with the USP or AAMI guidelines. Sterility assurance level shall be 0.000001.

Workmanship. The blood bags shall be free from defects which detract from their appearance or impair their serviceability. The item shall be free from tests, cuts, flash, holes, weak spots, rough areas, and contaminants.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Unit. Package (PG). One package, containing ninety-six individually packaged wrapped blood bags together with instructions for use as specified, constitutes one unit. The individual package shall be a sealed, peel-open type container capable of maintaining sterility of contents unless package is damaged or opened.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspections requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one blood collecting-dispensing bag as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one blood collecting-dispensing bag. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number and its respective Item Identification is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-01-364-8549	BLOOD COLLECTING DISPENSING-BAG, Indirect Transfusion, Plastic, Disposable Type, 300 ml, 96's

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5548
Location: c:\word\doc\nish\blood3cd

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54825	5 March 1993

3. DOCUMENT TITLE

BLOOD COLLECTING-DISPENSING, BAG, Indirect Transfusion, Plastic, Disposable Type

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center ATTN: DPSC-MSE	(215)7372870	444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403 Falls Church, VA 22041-3466 Telephone (703) 756 2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — *Previous editions are obsolete* — *96/29

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.