METRIC

A-A-54829
19 March 1993

COMMERCIAL ITEM DESCRIPTION

CATHETER, CARDIOVASCULAR, Flow Directed, Thermodilution

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Federal Specification GG-C-2736, which is cancelled.

This Commercial Item Description covers a sterile, cardiovascular flow directed, thermodilution four-lumen catheter suitable for use in determining hemodynamic pressures, cardiac output and thermal change and conducting blood sampling and solution infusion.

Salient characteristics:

Material. The rubber components shall be fabricated from natural rubber. The plastic tubing shall be fabricated from polyvinyl chloride which meets the requirements of the United State Pharmacopiea (USP) for Class IV plastics.

Style and design. The catheter shall be of a four-lumen design with one each for the distal lumen, the balloon, the thermistor and the proximal injectate post. At 1.5 cc inflation the tip of the catheter shall not protrude beyond the balloon. The tubing shall be radiopaque and shall have depth markings every 10 cm.

The hubs for the proximal injectate distal lumen and balloon shall be Luer in accordance with ISO 594.

The thermistor connector shall be compatible with the cardiac computer(s) designated by the catheter manufacturer on the individual catheter package.

Each catheter shall be provided with a 3 cc syringe with an internal stop which will limit syringe capacity to 1.5 cc.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

The catheter shall be a size 7FR with an effective length of 110 cm (minimum). The catheter shall meet the following requirements:

Thermistor resistance, unadjusted at 37°C	14,000 ± 2100 ohms
Diameter of inflated balloon	12 to 13 mm
Diameter of deflated balloon	8 Fr
Balloon volume	1.5 cc
Distance of injectate port from tip	29 to 30 cm
Distance of thermistor from tip	3.5 to 4 cm
Injectate lumen volume	0.70 to 0.95 cc
Deflation time from 1.5 cc air	No more than 5 secs.

With the balloon inflated the tip of the catheter shall not protrude beyond the balloon. The balloon lumen shall have a shut-off valve located next to the hub designed to maintain balloon inflation when the valve is closed.

Heavy metals. Plastic and rubber components coming in contact with intravenous solutions or patient's blood shall have lead and cadmium not greater than 5 parts per million (ppm) and barium shall not exceed 10 ppm when tested in accordance with USP.

Sterility, freedom from pyrogens and non-toxicity. The catheter shall be non-pyrogenic and non-toxic in accordance with USP. The catheter shall be sterile in accordance with USP or AAMI requirements. Sterility assurance level shall be 0.000001.

Pressure test. The balloon shall be inflated with 4.5 cc of air, immersed in water and examined for leaks for 10 seconds.

Thermistor response time. The thermistor shall exhibit a 60 percent response when the catheter is taken from room temperature (20 to 25°C) to immersion in 37°C ± 1°C USP isotonic saline. The measuring current shall not exceed 10 microamperes.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Saline soak. The balloon shall be placed in USP isotonic saline solution at 37°C ± 1°C for 16 hours. The balloon shall then be inflated with 1.5 cc air for 2 to 3 seconds every 20 minutes for 8 hours. After soaking for a total of 72 hours the inflation test shall be repeated every 20 minutes for 8 hours. The balloon shall not fail during this test.

Instructions. Each catheter shall be provided with instructions describing use of device.

Workmanship. The catheter shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One individually packaged sterile catheter as specified, constitutes one unit. The individual package shall be a sealed, peel-open type container capable of maintaining sterility of contents unless package is damaged or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one catheter as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one catheter. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commerical Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this documert are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, theofferor/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier s rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

NOTE: The following National Stock Number and its respective Item Identification is covered by this document:

National Stock Number	Item Identification
6515-01-030-2055	CATHETER, CARDIVASCULAR, Flow Directed, Thermodilution

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6515-5553
Location: C:MARK\CATHETER\D

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER	2 DOCUMENT DATE (YYMMDD)
	A-A-54829	19 March 1993

3 DOCUMENT TITLE

CATHETER, CARDIOVASCULAR, Flow Directed, Thermodilution

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center ATTN: DPSC-NSE	(215)737-2870	444-2870
ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete — 198-290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.