INCH-POUND

A-A-54830
19 March 1993

COMMERCIAL ITEM DESCRIPTION

TABLE, AUTOPSY

The General Services Administration has authorized the use of this Commercial Item Description

This Commercial Item Description covers an autopsy table that shall provide a stable working surface for postmortem examinations.

Salient Characteristics:

The autopsy table shall be constructed of stainless steel or corrosion-resistant material that has smooth edges and surfaces that are free of screws or rivets, as well as finished welds that are smooth and free of crevices, to facilitate cleaning and prevent the accumulation of micro-organisms.

The autopsy table shall consist of the following:

a. One removable slat for support of the cadaver's head.
b. A pedestal type table with a trough-shaped surface that slopes toward the sink for rapid and positive drainage.
c. A one-gallon aspirator bottle.
d. Waterproof electrical outlets.
e. A perimeter flushing system.
f. A disposal unit with a 1/2 or 3/4 hp motor.
g. One aspirator that includes a vacuum breaker with a variable control valve to adjust the amount of suction. A minimum 6 feet long suction hose and nozzle shall be provided.
h. One irrigator unit that includes a vacuum breaker with [illegible] knee-operated valves for hot and cold water and a minimum 8 feet long spray hose and nozzle.
i. One knee-operated waste control valve.
j. One rubber headrest.

AMSC N/A

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

k. One dissecting board 12 inches plus or minus 1/2 inches square and 3/4 inch plus or minus 1/8 inch thick with the cutting surface covered in neoprene cork.

l. One organ rinse container of at least 4 gallons capacity, approximately 10 inches high and 12 inches in diameter. Included shall be a grid plate, supported by the bottom of the container, that facilitates thorough rinsing. Overflow holes shall be provided around the entire top periphery of the container. A minimum of 8 feet of transparent tubing and a locking cover shall also be included.

m. Five removable slats for support of the body.

n. One dissecting tray, 25 inches plus or minus 1 inch long, 21 inches plus or minus 1 inch wide and 9 inches plus or minus 1/2 inch in depth. The tray shall have an elevated dissecting area, a removable perforated grid plate, a permanently attached instrument tray, and shall provide positive drainage.

o. One large, single compartment, integral sink at the end of the table, approximately 11 inches deep, 12 inches wide, and 20 inches long.

The autopsy table shall have overall dimensions of 34 inches plus or minus 1 inch high by 31-1/2 inches plus or minus 1 inch wide by 101 inches plus or minus 3 inches long and shall weigh 385 pounds plus or minus 15 pounds.

The electrical components of the autopsy table shall operate on single phase, alternating current, at 60 Hertz and 120 volts.

The autopsy table leakage current shall comply with Underwriters Laboratories, UL 544, Standard for Medical and Dental equipment, Type C non-patient equipment.

Identification marking. Each table shall be legibly and permanently marked with the following:

National Stock Number
Manufacturer's name
Contract number

Operation and service manuals. The contractor shall furnish manuals, handbooks, and/or brochures containing complete operation, installation and service/maintenance instructions with each table.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Workmanship. The autopsy table shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One autopsy table, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-3. The unit of product for examination purposes shall be one autopsy table.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one autopsy table.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Tests shall include, but shall not be limited to, the following:

The purpose table shall have operational testing by the [illegible] for compliance with the requirements of this document. Results of all tests shall be made available to the Government to assure that the apparatus will function as intended.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables specified using conversion to revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such performance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order

NOTE: The following National Stock Number is assigned by this document

6530-01-349-2748

MILITARY INTERESTS

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MP

Agent:

DLA-DH

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA - OSS

Project Number: 6530-2354

Location: PACIFC11.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

1. RECOMMEND A CHANGE.	1. DOCUMENT NUMBER A-A-54830	2. DOCUMENT DATE (YYMMDD) 29 MARCH 1990

3. DOCUMENT TITLE
TABLE, AUTOPSY

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870	(2) AUTOVON: 444-2870
c. ADDRESS (Include Zip Code) 2800 SOUTH 20TH STREET PHILADELPHIA, PA 19145	d. IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.