INCH-POUND

A-A-54835/1
8 October 1993

COMMERCIAL ITEM DESCRIPTION SHEET

RONGEUR, KERRISON,
90 Degree Upbite

The General Services Administration has authorized the use of this Commercial Item Description Sheet.

The complete requirements for procuring Kerrison rongeurs shall consist of this document and the issue in effect of Commercial Item Description A-A-54835.

Salient characteristics:

Style, design and dimensions. This specification covers the following style and classes of Kerrison rongeurs:

Style I - 90 degree upbite

Class 1 - 8 inch shaft

Class 2 - 4 inch shaft

The rongeurs shall be in accordance with Figures 1 and 2 as applicable.

Notes:

1. Intended use: Rongeurs of the Kerrison design covered in this Commercial Item Description Sheet are intended to remove bony or cartilaginous segments during surgical operations.

2. Copies of Commercial Item Description A-A-54835 may be obtained by writing the Defense Printing Service Detachment Office; Bldg. 4D; DoD-SSP; 700 Robbins Avenue; Philadelhpia, PA; 19222-5094.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

3. National Stock Numbers, Item Identifications and Classes.

National Stock Number	Item Identification	Class
6515-01-139-8240	RONGEUR, KERRISON, 5 mm, 90 Degree Upbite, 8 inch Straight Shaft	1
6515-00-331-5170	RONGEUR, KERRISON, 3 mm, 90 Degree Upbite, 4 inch Straight Shaft	2
6515-01-139-8980	RONGEUR, KERRISON, 3.9 mm, 90 Degree Upbite, 4 inch Straight Shaft	2
6515-01-139-8981	RONGEUR, KERRISON, 6.3 mm, 90 Degree Upbite, 4 inch Straight Shaft	2

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:


DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5614
Location: c:\word\aoc\david\rongeur1\sg

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.




NSN 6515 01-139-8240

UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES		
TOLERANCES ON:		
FRACTIONS	DECIMALS	ANGLES
±1/32	±	±2°

FIGURE 1. RONGEUR, KERRISON, 8 INCH

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:39Z
Check the source to verify that this is the current version before use.



NSN	A ±0.25MM	B ±0.5MM
6515-00-33-5170	3.00	2.0
6515-01-139-8980	3.90	4.0
6515-01-139-8981	6.30	4.0

UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES.
TOLERANCES ON
FRACTIONS ± 1/32 DECIMALS ± ANGLES ±

FIGURE 2 RONGEUR, KERRISON, 4 INCH

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:39Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54835/1	931008

3. DOCUMENT TITLE
RONGEUR, KERRISON, 90 Degree Upbite

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)
	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center ATTN: DPSC-MSE	(215)7372870	444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.