A-A-0053610(DM)
29 August 1988

INTERIM COMMERCIAL ITEM DESCRIPTION

DRESSING, HYDROCOLLOID, OCCLUSIVE

The General Services Administration has authorized the use of this Interim Commercial Item Description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This Commercial Item Description covers sterile, occlusive, hydrocolloid dressings.

Salient characteristics:

Shall be a sterile, occlusive, hydrocolloid dressing suitable for the treatment of ulcers, burns and skin graft donor sites. The dressings shall be sterile when tested in accordance with the USP. The dressings shall be of the following sizes:

SIZE	INCHES +/- 1/8 INCH
1	4 by 4
2	8 by 8

The dressing shall consist of a flexible water resistant outer layer and a self-adherent adhesive layer with moisture-reactive particles. The adhesive shall cover the entire area of one side of the dressing and the adhesive shall be covered by an easily peel-off paper facing to protect the adhesiveness of the dressing until used. The dressing shall create a soft gel by the interaction of the dressing with wound exudates. The dressing shall promote rapid epithelialization by acting as a barrier against secondary infection, shall adhere to dry or moist skin surfaces, shall be hypoallergenic and nonirritating, shall be easy to apply and remove, shall conform to bony prominences, shall be capable of being cut to size, shall not adhere to new granulating tissue upon removal of the dressing and the dressing shall have radiused corners.

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:22Z
Check the source to verify that this is the current version before use.

The soft gel shall be created by the interaction of wound exudates and the moisture reactive particles. The particles shall be composed of gelatin; pectin and coboxymethylcellulose. The particles shall be a matrix in a polymer, the combination of which shall be a homogenous mass. The adhesive shall become part of the gel.

The water resistant outer layer shall be a closed cell, thin; polyethylene foam layer.

The thickness of the dressing without the adhesive paper facing shall be a nominal 0.080 inches thick.

Workmanship. The dressings shall be free from defects which detract from their appearance or impair their serviceability.

Unit.

For Size 1 - Package (PG). Each dressing shall be individually sealed in a commercially available peel-open container capable of maintaining sterility unless package is damaged or opened. One package containing 5 individually sealed dressings, as specified, constitutes one unit.

For Size 2 - Package (PG). Each dressing shall be individually sealed in a commercially available peel-open container capable of maintaining sterility unless package is damaged or opened. One package containing 3 individually sealed dressings, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:22Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Testing and examination shall be conducted to determine compliances with all specification requirements. Test methods shall be suitable, accurate and reproducible.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for examination shall be one dressing. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5 maximum percent defective.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	SIZE
6510-01-181-7133	DRESSING, HYDROCOLLOID, OCCLUSIVE, 4 by 4 inch, 5s	1
6510-01-181-7134	DRESSING, HYDROCOLLOID, OCCLUSIVE, 8 by 8 inch, 3s	2

Ordering data. Expiration dating required, intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:22Z
Check the source to verify that this is the currentversion before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 71089

Project No. 6510-0760

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.