COMMERCIAL ITEM DESCRIPTION

TRIMMERS, GINGIVAL MARGIN, (Carbon Steel)

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification GG-T-1368B.

Salient characteristics:

This commercial item description covers carbon steel dental gingival margin trimmers for use in cavity preparation during dental operative procedures.

Classification. Trimmers covered by this commercial item description shall be of the following styles and numbers as specified in "Identification marking" and "NOTE" section.

Style 1 - No. 26
Style 2 - No. 27
Style 3 - No. 28
Style 4 - No. 29

Requirements. The trimmers shall be in accordance with the applicable requirements as specified herein.

Material. The trimmers shall be fabricated from water hardening carbon steel. The steel shall meet the requirements of ASTM A686 for Grade A (regular hardening), Class W1 1.00 to 1.20 percent carbon, spheroidized annealed condition, steel for chemical composition, raw material hardness, macrostructure, surface decarburization, carbide form and distribution, hardness penetration and fracture grain size.

Hardness. Operative ends and shanks of the finished instrument shall have a Rockwell hardness of C56 to C62. Handles shall have a minimum Rockwell hardness of C20.

Style, design and dimensions. Trimmers shall be one piece, double-ended trimmers of Dr. Black's design as modified by Dr. Ferrier. The style, design and dimensions shall be in accordance with Defense Personnel Support Center, (DPSC) Drawing 24406.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

The trimmers shall conform to the applicable Black's formula as specified in the drawing and as specified in table I herein for the applicable instrument number. Trimmers shall be suitable for planing enamel and cutting dentin in cavity preparation during dental operative procedures.

TABLE I. Instrument Number and applicable Black's Formula.

Style	Instrument number	Black's Formula
1	26	13-95-8-14
2	27	13-80-8-14
3	28	10-95-7-14
4	29	10-80-7-14

Drawings. The drawings forming a part of this commercial item description are not production drawings. Drawings establish limitations and control design and configuration. The supplier shall be responsible for preparing his own shop drawings. The supplier shall select dimensions and tolerances within the limitations prescribed on the drawings to provide a product with proper design, configuration, and function. No deviation from the dimensional limitations prescribed on the drawings is permissible without prior approval of the contracting officer.

Finish.

Operative ends. Operative ends (blades) of trimmers, including faces of beveled portion forming sharpened edges, shall have a fine, honed finish. The trimmer blades shall have a sharp cutting (trimming) edge and shall be free of defects such as rounded edges, coarse grind marks, cracks, burrs, feathers, and deep pits when examined under magnification, as specified herein.

Handles and shanks. Handles shall be straight and flat. Handles and shanks shall be plated as specified herein and shall have a bright finish. Handles shall have 40 $\pm$ 10 fillet or notched-type serrations per inch on seven faces and one smooth face for identification. Depth of serrations shall be 0.002 to 0.004 inch.

Plating. Handles and shanks shall be chromium (nickel over copper undercoat) or nickel over copper plated.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Corrosion resistance. The plated portions (handles and shanks) of the trimmers shall meet the requirements of the boil test as specified herein.

Identification marking. The trimmers shall be permanently and legibly marked on the smooth face of the handle. The trimmers shall be marked with the applicable instrument number and Black's formula. Alternate design numbers shall be as specified on the drawing and in table II below, for the respective NSN. The use of a score line on the handle of the instrument to distinguish the left operative end from the right operative end is optional.

TABLE II. Instrument alternate design numbers.

NSN	Instrument number	Alternate design number
6520-01-053-6361	26	30
6520-01-053-6362	27	31
6520-01-053-6363	28	28
6520-01-053-6364	29	29

Workmanship. The operative end, shank, and handle of the trimmers shall be free from defects such as coarse grind marks, cracks, or burrs which may detract from their appearance or impair their serviceability, proper functioning or intended use.

Unit. Each (EA). One trimmer, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification, where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one trimmer. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5 (percent defective), except for dimensions of operative end the inspection level shall be S-1 and sampling for tests (when applicable) shall be conducted at inspection level S-2. For tests, acceptance number shall be zero (0) for all sample sizes.

Examination. Examination shall be as specified herein. Examination shall include, but shall not be limited to the following defects listed below:

Defects

Style and design for the applicable type, and shape trimmers not as specified.
Cutting edge of operative end rounded or not sharp.
Cutting edge not free from nicks, burrs, feathers, indentations, or other irregularities that affect cutting performance.
Operative end finish not as specified.
Operative end not free of corrosion.
Operative end dimensions not as specified.
Plating of handle and shank not free from flaking, cracks, or blisters.
Plating of handle and shank incomplete or not continuous.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Handle and shank finish not as specified.
Handle not straight and flat.
Identification marking not present, not complete, not properly located, or illegible.
Trimmers not free from dirt, polishing compound, or foreign matter.
Handle serrations not uniform or not complete.
Handle serration dimensions not as specified.

Examination under magnification. The instruments shall be examined under four power (4X) magnification for the following defects:

Defects

Cutting edge of operative end rounded or not sharp.
Cutting edge not free from nicks, burrs, feathers, indentations, or other irregularities that affect cutting performance.
Operative end finish not as specified.

Tests.

Hardness. Finished instrument hardness shall be determined in accordance with ASTM Designation E18 (Rockwell) or E92 (Vickers or Diamond Pyramid).

Boil test. The trimmers shall be subjected to the boil test as specified in ASTM F1089.

Magnification inspection. The operative ends (blade portion) shall be inspected under four power (4X) magnification.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document.

National Stock Number	Item Identification
6520-01-053-6361	TRIMMER, GINGIVAL MARGIN, Black's Formula 13, 95, 8, 14; No. 26, Carbon Steel, Plated.
6520-01-053-6362	TRIMMER, GINGIVAL MARGIN, Black's Formula 13, 80, 8, 14; No. 27, Carbon Steel, Plated.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

6520-01-053-6363	TRIMMER, GINGIVAL MARGIN, Black's Formula 10, 95, 7, 14; No. 28, Carbon Steel, Plated.
6520-01-053-6364	TRIMMER, GINGIVAL MARGIN, Black's Formula 10, 80, 7, 14; No. 29, Carbon Steel, Plated.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2126

05903

Location: NCR Disk#6

☆U.S. GOVERNMENT PRINTING OFFICE: 1988 - 604-033/92799

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C.

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Staff Director
Defense Medical Material Board
Fort Detrick
Frederick Md. 21701

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.