INTERIM COMMERCIAL ITEM DESCRIPTION

RESUSCITATOR, HAND OPERATED

Adult Single Use

The General Services Administration has authorized the use of this Interim Commercial Item Description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This Interim Commercial Item Description covers a package of 6, single use, hand powered resuscitators, each with face mask, 90° airway, patient valve hand compression bag, oxygen accumulator tube, and oxygen supply tubing, for use in emergency medical care requiring artificial resuscitation.

Salient characteristics:

The resuscitator shall be a light weight single use type. It shall include a face mask, airway connector, patient valve, hand compression bag, oxygen accumulator tube, oxygen supply tubing, and operator instructions and/or markings.

The face mask shall be single use, adult size, transparent, and have an air cushioned cuff. The connector port shall accept a standard 22 mm/15 mm airway connector.

The airway connector shall be transparent with 22 mm/15 mm ends. It shall connect the face mask at 90 degrees to the patient valve.

The patient valve shall have an antilock feature which shall prevent lock-up during high flow conditions and be transparent.

The hand compression bag shall be single use, transparent, and have a minimum capacity of 1800 ml.

The oxygen accumulator tube shall be flexible, approximately 36 inches long by 22 mm ID, and shall connect to the back end of the compression bag.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

The oxygen supply tubing shall be transparent. It shall be approximately 7 feet long by approximately 3/16 or 1/4 inch ID. One end shall connect to the back end of the compression bag and the other end shall have a universal connector for oxygen supply.

The operator instructions and/or markings shall be consistent with ASTM F 920.

The resuscitator shall not have a pressure limiting system or be required to be disassembled by user. It shall be in compliance with and meet all examination and test requirements of ASTM F 920 except delete all references to demand valve and substitute 4.4° C for specified low end operating temperature.

Workmanship. The resuscitator shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 6 resuscitators individually sealed, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-4. The unit of product for examination purposes shall be one resuscitator as specified in this description.

Defects.

Component other than accumulator tube not transparent.

Missing component.

Resuscitator not individually sealed.

One unit is not 6 resuscitators.

Face mask not adult size.

Face mask does not have an air cushioned cuff.

Not meeting any one examination requirement, for the specified resuscitator, in ASTM F 920.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-2. The unit of product for test purposes shall be one resuscitator.

Tests

All tests in ASTM F 920 relating to the specified resuscitator shall be performed.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:22Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:22Z
Check the source to verify that this is the currentversion before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-01-279-8276

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4468

Location: ICID/RESUSCITATOR/H12-32

☆U.S. GOVERNMENT PRINTING OFFICE: 1989 - 604-033/10071

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.