A-A-50628
10 August 1988

COMMERCIAL ITEM DESCRIPTION

RONGEUR, STRAIGHT (Adson)

The General Services Administration has authorized the use of this commercial item description as a replacement for Military Specification MIL-R-42018(DM), which is cancelled.

This commercial item description covers an Adson straight rongeur suitable for removing cranial bone in neurosurgical procedures.

Salient characteristics:

General. The rongeur shall have an overall length of 8 inches. The rongeur shall have a straight shaft and jaws. The jaw faces shall be oval and smooth. The return action of the handles shall be provided by a double spring or other suitable mechanism.

Material. The material of the rongeur shall be Class 4 corrosion-resistant steel in accordance with ASTM Designation F899. The screws shall be fabricated from Class 3 or Class 4 corrosion-resisting steel in accordance with ASTM Designation F899.

Hardness. The rongeur jaws shall have a Rockwell Hardness C scale (HRC) of 50, minimum. The rongeur spring and spring screws shall have a Rockwell Hardness C scale (HRC) of 38 to 47 and the pivot screw shall have a Rockwell Hardness B scale (HRB) of 75, minimum. The hardness shall be tested in accordance with ASTM Designation E18 or ASTM Designations E92 and E140.

Style, design and dimensions. The rongeur shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 2178 for style, design and dimensions.

Finish. The finish of the rongeur shall be mirror, satin or matte. The matte finish shall exhibit a uniform nondirectional surface texture equivalent to that produced by lightly blasting the prepared surface with fine abrasives under pressure.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Designation. The rongeur shall be properly designated.

Corrosion resistance. The rongeur shall show no sign of corrosion when tested in accordance with ASTM Designation F1089.

Performance. The box lock mechanism and rongeur action shall be in accordance with ASTM Designation F1026.

Identification markings. The rongeurs shall bear the contractor's name or registered trademark and the word "stainless". Rongeurs of foreign manufacture shall bear the name of the country of origin.

Workmanship. The rongeur shall be free from defects which detract from its appearance or impair its serviceability.

Product Quality Level

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements

Unit. Each (EA). One rongeur, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one rongeur. Defects shall be those listed below.

Defects

Rongeur not free from cracks, pits, burs, sharp edges, scale, rough areas or corrosion.

Finish of the rongeur not as specified.

Style, design and dimensions not as specified.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one rongeur. AQL shall be 1.0 (percent defective) with an inspection level of S-2.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-299-8053	RONGEUR, Straight, Adson, 8 Inch

Ordering data. Intermediate exterior package quantities and labeling and marking must be specified in the contract and or order

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

A-A-53612

CIVIL AGENCY COOPERATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 44882

Project No. 6515-4265

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

See Instructions – Reverse Side

1. DOCUMENT NUMBER	2. DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION (*Mark one*)
	☐ VENDOR
	☐ USER
b. ADDRESS (*Street, City, State, ZIP Code*)	☐ MANUFACTURER
	☐ OTHER (*Specify*) _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER (*Last, First, MI – Optional*)	7b. TELEPHONE NUMBER (*Include Area Code*) – (*Optional*)
7c. MAILING ADDRESS (*Street, City, State, ZIP Code*) – (*Optional*)	8. DATE OF SUBMISSION (*YYMMDD*)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In [illegible] the qualification documents [illegible] Do[illegible] [illegible] suggestions This form may be detached folded along the lines indicated taped and mailed In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems Enter in block 6 any remarks not related to a specific paragraph of the document If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents nor to request waivers, deviations or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO [illegible] Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN DPSC-RST
2800 South 20th Street
Philadelphia, Pennsylvania 19101-0419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.