A-A-0053633(DM)
12 August 1988

INTERIM COMMERCIAL ITEM DESCRIPTION

RESUSCITATOR, MOUTH TO MASK,

Adult, Single Use, 10s

The General Services Administration has authorized the use of this Interim Commercial Item Description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This Interim Commercial Item Description covers a package of 10 single use, filtered, mouth to mask resuscitators for use in medical emergency care requiring artifical resuscitation.

Salient characteristics:

The resuscitator shall be single use type for mouth to mask resuscitation. It shall consist of a mouth piece, extension hose, bacterial/viral filter, patient valve, face mask, and operator instructions and markings.

The mouth piece shall have an elliptic shaped opening with the long axis approximately 1 inch long and the short axis approximately 5/16 inch long. It shall have a flared lip to help secure a leak proof fit to a rescuer's mouth. The other end shall be fitted to the extension hose.

The extension hose shall be flexible, capable of being bent into a "U" shape with the centerline at the top of the "U" being 1-1/2 inches apart without kinking or collapsing the hose. It shall be at least 5-3/4 inches long but not more than 10-1/4 inches long with an ID of approximately 22 mm.

The filter shall provide a bacterial/viral barrier to protect the rescuer. It shall be transparent.

The patient valve shall have an antilock feature and be transparent. It may be single use type. It shall direct patient exhalation away from the rescuer.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

The face mask shall be single use type, adult size, transparent, and have air cushioned cuff. The connector port shall accept a standard 22 mm/15 mm airway connector and shall connect to the patient valve.

The operator instructions and markings shall be consistent with ASTM F 920.

The materials, face mask, patient valve function after contamination with vomitus, patient valve leakage, operating and storage capability, and operator instructions and markings shall be in accordance with ASTM F 920 except substitute 4.4°C for low end operating temperature. These features, components, and materials shall be examined and tested in accordance with ASTM F 920.

Workmanship. The mouth to mask resuscitator shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 10 resuscitators, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-4. The unit of product for examination purposes shall be one resuscitator as specified in this description.

Defects.

Face mask, patient valve, or filter not transparent.

Missing component.

Resuscitator not individually sealed.

One unit not 10 resuscitators.

Face mask not adult size.

Face mask does not have an air cushioned cuff.

Patient valve does not direct patient exalation away from rescuer.

A specified feature/component/material does not pass ASTM F 920 examination.

Sampling for tests. Sampling for tests (dimensional) shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-2. The unit of product for test purposes shall be one resuscitator.

Tests

All features/components/materials specified herein to be tested in accordance with ASTM F 920 shall be so tested.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:22Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:22Z
Check the source to verify that this is the currentversion before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-01-280-1841

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4471

Location:
ICID/RESUSCITATOR/MOUTH/H12-33

☆U.S. GOVERNMENT PRINTING OFFICE: 1989 - 604-033/10070

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:22Z
Check the source to verify that this is the current version before use.