COMMERCIAL ITEM DESCRIPTION

CATHETER, URETHRAL (Foley, Carson Type)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers one size of Olive Coude tipped Foley urethral catheter.

Salient characteristics.

Material. The catheter shall be manufactured from latex rubber. The natural rubber catheter shall be coated throughout its working length with polytetrafluoroethylene. The coating shall be uniform and shall completely cover the natural rubber.

Style, design, and performance. The catheter shall be 6 millimeters (18 French) in diameter, and shall have a working length of 340 millimeters. The catheter shall be toleranced in accordance with, and shall meet all requirements of American Society for Testing and Materials (ASTM) standard F623.

The catheter shall have one drainage eye located 1.50 $\pm$ 0.25 inches from the proximal end (see note 2) of the balloon, measured from the distal end of the eye. The catheter shall have a 5.0 mm diameter olive tip. The catheter shall taper from 6 mm diameter to a nominal 4.5 mm diameter at the beginning of the olive tip. The coude tip shall form a $45^0 \pm 6^0$ angle. The angle shall be formed with a 1 inch radius of curvature centered 7/8 inch from the distal tip. The drainage eye shall be positioned in the center of the bend on a side of the catheter (i.e. not on inner or outer circumference of bend).

Markings. The catheter shall be permanently and legibly marked with the manufacturer's name or registered trademark, the French size, and the balloon capacity.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Physical Properties. Catheters shall have the following physical properties:

Breaking Strength	60 Pounds(Force)
Elongation (Ultimate)	750 Percent

Catheters shall be resistant to oxidation and aging when stored properly up to 4 years.

Sterility. Catheters shall be sterile.

Workmanship. The catheter shall be processed in such a manner as to be uniform and free from defects which affect life, utility, or appearance.

Product Quality Level

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Package (PG). One package containing twelve individually packaged sterile catheters, as specified, constitutes one unit. The individually package shall be a sealed, peel-open type container capable of maintaining sterility of contents unless package is damaged or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:22Z
Check the source to verify that this is the currenversion before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Catheters shall be examined to determine compliance with specification requirements.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one catheter.

Tests.

Performance Tests. Catheters shall be tested in accordance with the test methods of ASTM F623. The AQL shall be 1.5 with sampling level S-2.

Certificate of biocompatibility. A certificate of compliance, stipulating offered items meet the requirements of ASTM F623, shall be acceptable, provided that the contractor has supplied acceptable items to the Government within the past three years, and that there has been no changes to material, formulation and method of processing. The certificate shall include the current contract number and applicable NSN, the date of actual testing, the test method and results, the name and title of the person evaluating test results, and a statement to the effect that material formulation and processing have not changed since the testing was performed.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:22Z
Check the source to verify that this is the current version before use.

Physical properties. Catheters shall be tested for physical properties in accordance with ASTM standard D412, except as stated herein. Catheters selected for testing need not be coated with polytetrafluoroethylene, but shall otherwise be processed identically to the remainder of the lot. Sampling shall be level S-1 with an AQL of 1.0.

Two test specimens shall be cut from each catheter in the sample. One specimen shall be tested as cut. The second specimen shall be aged in accordance with ASTM D573 for 7 days at 70°C.

In either the aged or unaged testing, any test failure which is determined to have been caused by slippage of the specimen in the jaws of the testing machine shall be disregarded. In this event, an entire new catheter shall be tested.

Breaking strength shall be the observed force at break. Acceptance shall be as follows:

Physical properties (unaged). Calculate the arithmetic mean for both the ultimate elongation and breaking force of all specimens. Acceptance shall be based upon the arithmetic mean, unless any ultimate elongation is less than 625 percent. An elongation value less than 625 percent which is not attributable to grip slippage shall be cause for rejection.

Physical properties (aged). Each value obtained from an aged specimen shall be compared to the value obtained from the unaged specimen cut from the same catheter. The elongation and breaking strength of the aged specimen shall not be less than 75% of the value obtained from the unaged specimen. Elongation and breaking strength shall be considered separate tests. Acceptance shall be based upon an AQL of 1.0.

Sterility. Catheters shall be tested for sterility in accordance with the USP. Sampling and acceptance shall be in accordance with the USP.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standard, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:22Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTES:

1. The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-205-2313	CATHETER, URETHRAL, CARSON, Polytetraflouroethylene Coated Rubber, Olive Coude Tip, 5 ml Balloon, 18 French, 12s.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

2. ASTM Standard F623 defines "proximal" as the end of the catheter to be inserted into the patient. This document uses the common definition of "proximal", as that end closest to the physician. In both cases, "distal" is defined as the end opposite to the proximal end. Caution should be exercised in the origin of the word "proximal" and its definition when using this document and ASTM F623.

Ordering data . Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4342

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in ubmitting comments and suggestions for improvements All users of military standardization documents are invited to provide iggestions This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems Enter in block 6 any remarks not related to a specific paragraph of the document If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4968 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
PHILADELPHIA,PENNSYLVANIA 19101-8419

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:22Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53634	CATHETER, URETHRAL (Foley, Carson Type)

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil — Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.