A-A-53635
12 August 1988

COMMERCIAL ITEM DESCRIPTION

CONNECTOR, TRACHEAL TUBE (ACUTE ANGLED)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers three sizes of no-port acute angled tracheal tube connectors.

Salient characteristics.

Sizes. Connectors shall be supplied in the following sizes, as specified in contract or purchase order (see page 4 for NSN/size cross-reference).

Size	Size tracheal tube with which connector is designed to be used
6	6 millimeter inside diameter
7	7 millimeter inside diameter
8	8 millimeter inside diameter

Material. Connectors shall be manufactured from brass or aluminum. Connectors manufactured from brass shall be chrome plated.

Finish. Connectors manufactured from aluminum shall be anodized. The plating on brass connectors shall be uniform and free from peeling, flaking, blistering, pitting, scratches, or other defects. The plating shall extend into the lumen of the connector at both ends.

Style and design. Connectors shall be in accordance with American National Standards Institute (ANSI) Standard Z79.2. In addition, the 15 mm taper shall be in accordance with American Society for Testing and Materials (ASTM) Standard F1054. In addition to the requirements of ANSI Z79.2, the connector may have 360° circumferential ridges on the tube end. The connector shall be designed for low flow resistance, shall have a bend radius (R) of 3/4 inch minimum (centerline) and an observed angle (O) of 60 ± 5^{0}.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Corrosion resistance. The connector shall be capable of withstanding repeated sterilization by steam, ethylene oxide, or chemical disinfectant. Connectors shall be tested for corrosion resistance as specified under Quality Assurance Provisions.

Workmanship. The connectors shall be processed in such a manner as to be uniform and free from defects which affect life, utility, or appearance.

Product Quality Level

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Each (EA). One connector, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Connectors shall be examined to determine compliance with specification requirements.

Sampling for plating examination. Sampling of plated connectors shall be conducted in accordance with MIL-STD-105, level S-2.

Plating examination. Plated connectors shall be examined for signs of porosity, flaking, non-adhesion, blisters, and other plating defects. If any unit shows signs of defects the entire sample shall be tested as specified under Referee Plating Test. The AQL shall be 0.25.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one connector. Sampling shall be level S-1. The same sample shall be used for the taper test and the corrosion test.

Tests.

Corrosion tests. Connectors shall be tested in accordance with Resistance to Boil Test of ASTM Standard F1089.

Taper Test. The 15 mm taper shall be tested in accordance with ASTM Standard F1054.

Referee Plating Test. Connectors requiring the referee plating test (see plating examination) shall be tested in accordance with the Adhesion Test (paragraph 4.5.2) of Federal Specification QQ-C-320.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standard, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	SIZE
6515-01-139-7841	CONNECTOR, ACUTE ANGLE CURVED, Tracheal Tube, 6 mm I.D.	6
6515-01-139-7842	CONNECTOR, ACUTE ANGLE CURVED, Tracheal Tube, 7 mm I.D.	7
6515-01-139-7843	CONNECTOR, ACUTE ANGLE CURVED, Tracheal Tube, 8 mm I.D.	8

Ordering data . Intermediate exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity.

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 06639

Project No. 6515-4330

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53635	CONNECTOR, TRACHEAL TUBE (ACUTE ANGLED)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI) – Optional*	b. WORK TELEPHONE NUMBER *(Include Area Code) – Optional*
c. MAILING ADDRESS *(Street, City, State, ZIP Code) – Optional*	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge (*DO NOT STAPLE*), and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

Fold along this line

Fold along this line

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4586 WASHINGTON D.C.

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
PHILADELPHIA, PENNSYLVANIA 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.