A-A-53636
10 August 1988

COMMERCIAL ITEM DESCRIPTION

CLIP, HEMOSTATIC, Large

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a large hemostatic suitable for use in blood vessel occlusion in surgical procedures.

Salient characteristics:

General. The clip shall be a large size stainless steel or tantalum hemostatic clips mounted in color-coded cartridges.

Material. Stainless steel clips shall be fabricated of 316L stainless steel in accordance with American Society for Testing and Material (ASTM) Designation F56 or ASTM Designation F138, Grade 2. The tantalum clips shall be fabricated of material in accordance with ASTM Designation F560. Cartridges shall be fabricated of a suitable rigid plastic.

Style and design. The clips shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 23384 for style and design.

Clip. The shape of the clips shall resemble that of a staple with its inner surface textured to prevent slippage of the vessel being ligated. The center of the stainless steel clip shall have reduced thickness to aid closure. The ends of the clip shall be serrated to enable pick-up and firm retention in the forceps jaws. The clip shall be 8 mm when open and 12 mm when closed, as measured from the center of the clip to either end.

Cartridge. The cartridge shall be uniquely color coded to identify and distinguish the particular size of the clips from other sizes of clips. The particular color of the cartridge shall match the color of those forceps compatible with large clips.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Compatibility and closure. When subjected to the closure test specified herein, the clips shall exhibit no splitting or fracture.

Sterility. The clips shall be tested for sterility in accordance with procedures contained in the USP.

Workmanship. The hemostatic clips shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 18 cartridges, each cartridge containing 6 individually packaged sterile clips in a peel-open type container, capable of maintaining sterility of the contents unless package is damaged or opened, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. At inspection level S-2, dimensional testing shall be at a 1.5% AQL and material and closure testing shall be at a 2.5% AQL. The unit of product for test purposes shall be one hemostatic clip.

Compatibility and closure test. Remove a clip with forceps having the same handle color coding as that for the cartridge. If the forceps fails to engage the clip properly, verify that the forceps is the proper size. If so, the lot is rejected. Close the clip completely. Examine with the naked eye for splits or fractures. A minimum of 6 clips shall be tested.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics fo this description and concorms to the producer's own drawings, specifications, standards and quality assurance practices. The Government reserves the right to require proof of such conformance prior to firest delivery and thereafter as may be otherwise provide under hte provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE 1: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-139-9104	CLIP, HEMOSTATIC, Stainless Steel, Large, 108 Clips in 18 Cartridges
6515-01-164-8881	CLIP, HEMOSTATIC, Tantalum, Large, 108 Clips in 18 Cartridges

NOTE 2. The items covered by the following National Stock Numbers are compatible with and color coded to match this specification item.

6515-01-139-7885	FORCEPS, HEMOSTATIC CLIP, Applying, Large, Angled 15 Degrees, 10-1/2 Inch
6515-01-167-6675	FORCEPS, HEMOSTATIC CLIP, Applying, Large, Angled 45 Degrees, 10-1/2 Inch

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4244

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53636	CLIP, HEMOSTATIC, LARGE

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI) – Optional*	b. WORK TELEPHONE NUMBER *(Include Area Code) – Optional*
c. MAILING ADDRESS *(Street, City, State, ZIP Code) – Optional*	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements All users of military standardization documents are invited to provide suggestions This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements

(Fold along this line)

(Fold along this line)

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4961 Alexandria VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RST
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.