A-A-53637
15 August 1988

COMMERCIAL ITEM DESCRIPTION

TEST PAPER, HYDROGEN ION

The General Services Administration has authorized the use of this commercial item description in preference to Interim Federal Specification NNN-T-001529.

This commercial item description covers rolls and strips of indicator test papers suitable for determining the hydrogen ion concentration of solutions.

Salient characteristics.

Indicator test papers shall be of the following ranges, classes and styles:

Ranges:

Indicator test papers shall show distinct and rapid color change corresponding to the furnished comparison color chart for the specified pH unit as indicated for the appropriate pH range.

Wide range. Comparison chart for each full pH unit.

Range 1 - pH 2.0 to 10.0 (even pH units)
Range 2 - pH 1.0 to 11.0 (odd pH units)
Range 3 - pH 1.0 to 11.0 (consecutive pH units)
Range 4 - pH 1.0 to 12.0
Range 5 - pH 2.0 to 9.0
Range 6 - pH 0.0 to 14.0

Medium range. Comparison chart for each full pH unit.

Range 1 - pH 1.0 to 5.0
Range 2 - pH 4.0 to 8.0
Range 3 - pH 6.0 to 11.0
Range 4 - pH 10.0 to 14.0

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Medium range. Comparison chart for each 0.5 pH unit.

Range 5 - pH 0.0 to 6.0
Range 6 - pH 5.0 to 10.0
Range 7 - pH 7.5 to 14.0

Short range. Comparison chart for each 0.5 pH unit.

Range 1 - pH 6.0 to 8.0
Range 2 - pH 8.0 to 9.5

Short range. Comparison chart for each 0.3 pH unit.

Range 3 - pH 0.0 to 2.5
Range 4 - pH 2.5 to 4.5
Range 5 - pH 4.0 to 7.0
Range 6 - pH 6.5 to 10.0
Range 7 - pH 11.0 to 13.0

Classes:

Class A. Class A indicator test paper shall be furnished in the form of a roll, 1/4 ± 1/16 inch wide by 15 ± 1, -0 feet long. The diameter of the roll shall be not more that 1-3/4 inches.

Class B. Class B indicator test paper shall be furnished in strips, 100 strips in a vial, each strip 2 ± 1/8 inches long and 7/32 ± 1/32 incnes wide.

Styles:

Style 1. Style 1 indicator test papers shall be furnished in a dispenser.

Style 2. Style 2 indicator test papers shall be furnished without a dispenser.

The indicator test paper shall be prepared from white, unsized filter paper which is composed of 100 percent fully bleached fibers. The unimpregnated filter paper shall have a basis weight of 20 ± 2 pounds and shall have a thickness of 0.006 ± 0.001 inch.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:22Z
Check the source to verify that this is the currentversion before use.

A color comparison chart shall be affixed to or accompany each dispenser, or if no dispenser is furnished, it shall accompany each package of paper or vial. The chart shall show the colors produced by the paper, and shall state the corresponding pH value required to produce each color.

Unless otherwise specified, a dispenser shall be supplied with each roll, set of rolls, or pack of papers. Dispensers for roll paper shall be fabricated from transparent polystyrene or acrylic plastic material and shall be designed to hold two rolls approximately 1-3/4 inches outside diameter, and 1/4 inch wide. Roll dispensers shall be so designed so that a convenient length can be torn for the test without opening the dispenser. The dispenser shall have a removable tightly fitting cover so that replacement roll of test paper and color charts may be placed therein. Each replacement roll of paper shall be wrapped in vinyl coated, heat-sealed aluminum foil, and then enclosed in an amber colored polystyrene or acrylic plastic container. The cover shall fit tightly and securely over the container. The container shall be approximately 2 inches long by 2 inches high by 1-3/8 inches deep, and designed to hold five rolls.

To determine color reaction, sample test papers shall be tested using buffer solutions corresponding to pH values within the specified range.

Workmanship. The indicator test papers shall be free from defects which detract from their appearance or impair their serviceability.

Unit. For NSN 6640-00-442-8000/6640-00-442-9005
Each (EA). Two rolls of test paper, two color charts, and one dispenser, as specified, constitutes one unit.

For NSN 6640-00-442-8010/6640-00-442-8020
6640-00-442-9015/6640-00-442-9025
Package (PG). One package containing five individually wrapped rolls of test paper and one color chart enclosed in a suitable amber-colored plastic container, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test.

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one unit of test papers.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: This document covers, but is not limited to, the following National Stock Numbers:

NSN	Range	Class	Style
6640-00-442-8000	Short, pH range 1 and 2	A	1
6640-00-442-8010	Short, pH range 1	A	2
6640-00-442-8020	Short, pH range 2	A	2
6640-00-442-9005	Wide, pH range 1 and 2	A	1
6640-00-442-9015	Wide, pH range 1	A	2
6640-00-442-9025	Wide, pH range 2	A	2

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:22Z
Check the source to verify that this is the current version before use.

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6640-1508

Location: Disk No. 13

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:22Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER
A-A-53637

2 DOCUMENT TITLE
TEST PAPER, HYDROGEN ION

3a NAME OF SUBMITTING ORGANIZATION

b ADDRESS (*Street, City, State, ZIP Code*)

4 TYPE OF ORGANIZATION (*Mark one*)

- ☐ VENDOR
- ☐ USER
- ☐ MANUFACTURER
- ☐ OTHER (*Specify*)

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER (*Last, First, MI – Optional*)

b. MAILING ADDRESS (*Street, City, State, ZIP Code*) – Optional

b WORK TELEPHONE NUMBER (*Include Area Code*) – Optional

8 DATE OF SUBMISSION (*YYMMDD*)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements All users of military standardization documents are invited to provide suggestions This form may be detached, folded along the lines indicated, taped along the loose edge (*DO NOT STAPLE*), and mailed In block 5 be as specific as possible about particular problem areas such as wording which required interpretation was too rigid restrictive loose ambiguous or was incompatible and give proposed wording changes which would alleviate the problems Enter in block 6 any remarks not related to a specific paragraph of the document If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN DPSC-ASTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.