INCH-POUND

A-A-53644A
31 May 1994
SUPERSEDING
A-A-53644
12 August 1988

COMMERCIAL ITEM DESCRIPTION

SYRINGE AND NEEDLE, HYPODERMIC, DISPOSABLE

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers sterile, disposable, hypodermic syringe and needle units suitable for a single injection.

Salient characteristics:

General. The syringe and needle units shall have a syringe component with a volume of 3 milliliter (ml) and a barrel length of 2-1/2 to 3-1/4 inches.

Material. Needle cannula shall be fabricated from Class 3 Type 304 stainless steel in accordance with American Society for Testing and Materials (ASTM) Designation F899. The needle hub shall be fabricated from rigid polypropylene that meets the United States Pharmacopeia (USP) requirements for Class IV plastic or aluminum. The sheath shall be fabricated from rigid plastic. Syringe barrel and plunger shall be rigid polypropylene which meets the USP requirements for Class IV plastics. The interior of the barrel and the exterior of the needle shall be lightly coated with an inert pharmaceutical grade silicone lubricant. The plunger tip shall be inert natural rubber or neoprene. As an alternate the plunger tip shall be natural or synthetic isoprene based. The plunger tip material shall meet the USP requirements for Elastomeric closures for injections.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Style and design. Syringe and needle shall be Luer type in accordance with International Standards Organization (ISO) 594. Syringe barrel shall be flanged to provide finger grip and anti-roll feature. Inside of the barrel shall have a smooth, uniform surface, free from pits, grind marks, bubbles, flash, or other imperfections. Plunger shall glide through the barrel while maintaining fluid-tight fit. The plunger tip and barrel shall both be beveled and form a flush fit at zero volume. Barrel shall be transparent and the plunger tip shall have two visible rings touching the sides of the barrel.

The outside surface of the barrel shall have each 0.5 ml marked with a main graduation marking and with the numerical volume. Each 0.1 ml shall be subgraduated. Main graduation markings shall be at least 33 percent longer than subgraduation markings. Barrel shall also have the name or registered trademark of the contractor. Actual fluid volume shall be determined by position of the distal ring of the plunger tip.

Needle shall be furnished with a removable, protective cap. The tip of the needle point shall be ground to maximum keenness. The heel of the needle point shall be dull to prevent coring.

Needle Cannula length (inch)	Tolerance (inch)
5/8	+3/64, -1/6
3/4 - 1 1/2	±1/16

Gauge	Point length (inch)
20	0.148 - 0.180
22	0.120 - 0.146
23	0.105 - 0.129
25	0.078 - 0.103

Aluminum hub. Needles with aluminum hubs shall show no evidence of corrosion when covered with a 3 percent sodium chloride solution at 20 - 25° C for one hour. There shall be no discoloration visible or any sign of reaction.

Except as specified herein the needle shall be in accordance with ISO 7864. Sterile Hypodermic needles for single use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Accuracy test. The syringe shall deliver ± 5 percent of the capacity indicated at each major graduation mark when tested against a calibrated burette. As an alternate, the determination of capacity test listed in ISO 7886 is acceptable.

Leakage test The water pressure test system shall have a water pressure gauge and a rubber plug. The needle shall be secured to the Luer tip of the syringe by applying it with a 5 pound axial force and while maintaining this compressive stress twisting the needle hub 45 degrees relative to the Luer tip. The needle shall be inserted into the plug. The syringe shall be tested for liquid leakage at the syringe piston and syringe nozzle under compression and during aspiration in accordance with ISO 7886.

Sterility, pyrogenicity, toxicity. The syringe needle, shall be nonpyrogenic and nontoxic in accordance with the USP. The syringe and needle shall be sterile in accordance with USP or the Association for the Advancement of Medical Instrumentation (AAMI) requirements, the sterility assurance level shall be 0.0001.

Color coding. The hub of the needle shall be color coded. The color designations specified in the International Standards Organization (ISO) 6009 are optional. If colored the needle sheath is not required to meet ISO 6009.

Instructions. Each unit package shall be provided with instructions describing the proper techniques for dispensing the sterile syringes and needles, as applicable, from the containers and preparing them for use. In addition, information as to methods for destroying syringes and needles, as applicable, to prevent their re-use shall also be shown.

Workmanship. The syringe and needle shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 100 individually packaged, sterile and needle units of one style only, as specified, constitutes one unit. The individual package shall be either a sealed, peel-open container or a rigid plastic tamper-proof container capable of maintaining sterility of contents unless package is damaged or opened.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one syringe and needle unit. Sampling shall be inspection level S-1 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

NOTE 1: The following National Stock Number and its respective Item Identification is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-00-149-1206	SYRINGE AND NEEDLE, Hypodermic, Disposable, 23 gauge, 1 inch needle, 3 cc, 100s
6515-00-719-7217	SYRINGE AND NEEDLE, Hypodermic, Disposable, 25 gauge, 5/8 inch, needle, 3 cc, 100s
6515-00-059-2791	SYRINGE AND NEEDLE, Hypodermic, Disposable, 25 gauge, 5/8 inch needle, 3 cc, 100s
6515-00-059-2790	SYRINGE AND NEEDLE, Hypodermic, Disposable, 20 gauge, 1-1/2 inch needle, 3 cc, 100s
6515-00-053-0945	SYRINGE AND NEEDLE, Hypodermic, Disposable, 22 gauge, 1 inch needle, 3 cc, 100s
6515-00-059-2789	SYRINGE AND NEEDLE, Hypodermic, Disposable, 22 gauge, 1-1/2 inch needle, 3 cc, 100s

NOTE 2: ISO Standards are available from the American National Standards Institute Inc., 11 West 42nd Street, New York, NY 10036

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5697
Location: levin/syrneed/sg

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.