COMMERCIAL ITEM DESCRIPTION

CLIP, HEMOSTATIC (Medium)

This commercial item description covers medium size hemostatic ligating clips suitable for use in blood vessel occlusion in surgical procedures.

Salient characteristics:

General. The clips shall be medium size stainless steel or tantalum hemostatic clips mounted in color-coded cartridges.

Material. Stainless steel clips shall be fabricated of 316L stainless steel in accordance with ASTM Designation F56 or Designation F138, Grade 2. The tantalum clips shall be fabricated from material in accordance with ASTM Designation F560. Cartridges shall be fabricated of hardened nylon.

Style and design. The clips shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 24584 for style and design.

Clip. The shape of the clips shall resemble that of a staple with its inner surface textured to prevent slippage of the vessel being ligated. The center of the stainless steel clip shall have reduced thickness to aid closure. The ends of the clip shall be serrated to enable pick-up and firm retention in the forceps jaws. The clip shall be 3 mm when open and 5 mm when closed, as measured from the center of the clip to either end.

Cartridge. The cartridge shall be uniquely color coded to identify and distinguish the particular size of the clips from other sizes of clips. The particular color of the cartridge shall match the color of those forceps compatible with large clips.

Compatibility and closure. When subjected to the closure test specified herein, the clips shall exhibit no splitting or fracture.

Sterility. The clips shall be tested for sterility in accordance with USP or AAMI guidelines.

AMSC/NA FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Workmanship. The hemostatic clips shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 36 cartridges, each cartridge containing 6 individually packaged sterile clips in a peel-open type container, capable of maintaining sterility of the contents unless package is damaged or opened, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for Inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. At inspection level S-2, dimensional testing shall be at a 1.5 percent AQL and material and closure testing shall be at a 2.5 percent AQL. The unit of product for test purposes shall be one hemostatic clip.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Compatibility and closure test. Remove a clip with forceps having the same handle color coding as that for the cartridge. If the forceps fails to engage the clip properly, verify that the forceps is the proper size. If so, the lot is rejected. Close the clip completely. Examine with the naked eye for splits or fractures. A minimum of 6 clips shall be tested.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE 1: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-139-9084	CLIP, HEMOSTATIC, Medium, Corrosion-Resisting Steel, 216 Clips in 36 Cartridges
6515-01-173-8924	CLIP, HEMOSTATIC, Medium, Tantalum, 216 Clips in 36 Cartridges

NOTE 2. The items covered by the following National Stock Numbers are compatible with and color coded to match this specification item.

NSN	ITEM IDENTIFICATION
6515-01-139-7883	FORCEPS, HEMOSTATIC CLIP, Angled, 15 Degrees, 7.5 Inch
6515-01-167-6671	FORCEPS, HEMOSTATIC CLIP, Straight, 5 Inch
6515-01-167-6672	FORCEPS, HEMOSTATIC CLIP, Angled, 15 Degrees, 5.75 Inch
6515-01-167-6673	FORCEPS, HEMOSTATIC CLIP, Angled, 45 Degrees, 10.5 Inch
6515-01-167-6674	FORCEPS, HEMOSTATIC CLIP, Angled, 15 Degrees, 10.5 Inch

Ordering data (Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4364

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53645	CLIP, HEMOSTATIC (Medium)

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street City State ZIP Code*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a Paragraph Number and Wording

b Recommended Wording

c Reason Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last First MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

(FOLD DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

DD FORM 1426

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge (*DO NOT STAPLE*), and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(*s*) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.