A-A-53656
19 August 1988

COMMERCIAL ITEM DESCRIPTION

THERMOMETER, CLINICAL, HUMAN (Electronic, Battery Powered)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers an electronic clinical thermometer suitable for body temperature measurement in humans either orally or rectally.

Salient characteristics:

General. The thermometer shall be a portable, battery operated thermometer capable of pulse measurement and temperature measurement in both Fahrenheit and Celsius scales. Temperature shall be measured in either predictive mode or monitor mode as selected by the operator.

The thermometer system shall consist of a thermometer, oral probe, rectal probe, battery charger/storage base, disposable probe covers, and a dispenser probe cover bracket.

The thermometer shall meet the following physical and performance characteristics.

Temperature range

Predictive mode	90.0°F to 107.9°F 32.2°C to 42.1°C
Monitor mode	88.0°F to 107.9°F 31.1°C to 42.1°C
Accuracy	± 0.2°F, ± 0.1°C when tested as specified
Pulse mode	20 to 200 beats per minute

Battery

Type	Rechargeable Nickel-Cadmium
Recharging time	Maximum 10 hours for a fully discharged battery

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Battery life	5 hours continuous operation when fully charged
Probe	
Oral	Blue color
Rectal	Red color

Construction. Thermometer proper shall consist of internal rechargeable nickel-cadmium battery, LED (light emitting diode) five digit display panel, probe storage well, probe connector socket, Fahrenheit/Celsius selector switch, predictive mode/monitor mode selector switch, pulse button, and carrying strap.

The thermometer case shall be made of durable lightweight plastic. The volume of the thermometer shall be approximately 50 cubic inches and the weight shall be approximately 12 ounces. When properly placed on the charger unit, the battery in the thermometer will be automatically charged. A small charge light shall appear on the display panel to indicate that the battery is charging. The switches and the pulse button shall be located on the thermometer for convenient operation. The probe socket shall accept either an oral or rectal probe and provide proper fit and good electrical contact.

The storage well shall be suitable for storing the probe when the probe is not in use. The case of the thermometer shall have a compartment for holding a box of disposable probe covers.

The cord for each probe shall be in a coil form. The length of the cord shall be approximately 2 feet when coiled and approximately 4 feet when stretched.

Each thermometer shall be supplied with 1000 clean, disposable, rigid plastic probe covers. The probe covers shall be of appropriate shape and size and capable of fitting onto the probe firmly without having to touch the probe cover. Each probe shall have an ejection button at the base of the probe which shall allow ejecting the probe cover into a waste container without human contact.

The battery charger/storage base shall have an approximate volume of 85 cubic inches. The weight of the charger/base shall not exceed 1.5 pounds. The charger unit shall be supplied with a cord of at least 6 feet long with a 3 prong hospital grade plug. The electrical characteristics shall be consistent with the requirements contained in the latest edition of MIL-D-42048.

The thermometer shall include a carrying strap capable of being hung around the neck for convenience during use. The length of the carrying strap shall be adjustable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Each thermometer shall be supplied with a probe cover dispenser and wall bracket.

Material. Unless specified elsewhere in this document, all components of the thermometer system shall be fabricated from materials which are inherently resistant to corrosion or which are treated to render them corrosion-resistant. Where painting or plating is used, appropriate cleaning and priming shall be performed prior to finishing.

Unless otherwise specified, the finish shall be equivalent or superior to that of comparable commercial equipment, and shall be free of gaps, scratches, chips, or other imperfections.

Operation.

Predictive mode. When this mode is selected by the switch, the thermometer shall be capable of measuring the patient temperature either orally or rectally. An audible tone will sound and the patient's temperature will be displayed on the display panel in degrees and tenth of a degree. No more than 30 seconds will be required to attain this temperature reading. Once the temperature reading is displayed, it shall remain displayed for about 20 seconds. Then the display will clear and the thermometer unit shall shut off automatically. The displayed temperature shall be either in Fahrenheit or Celsius as selected by the operator.

Monitor mode. In monitor mode, the thermometer shall continuously measure the patient's temperature as it rises or falls. The thermometer shall take no more than five minutes to attain the patient's current temperature. Then, the thermometer display shall change as patient's temperature rises or falls. The temperature reading shall be either in Fahrenheit or Celsius scale selected by the operator.

Pulse mode. By depressing the pulse mode button on the first pulse beat and again pressing it on the twentieth pulse beat, the thermometer shall calculate and display patient's pulse rate in beats per minute. The pulse rate reading shall remain displayed for 20 seconds before automatically shutting off.

Error messages. The thermometer shall display the error messages under the following conditions:

1. When the thermometer unit fails to obtain a temperature reading within one minute after probe insertion.

2. When the unit senses a loss of proper tissue contact within the mouth for more than 30 seconds.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

3. When the probe temperature exceeds 42.1°C, 107.9°F.

4. Low battery charge condition.

Workmanship. The electronic thermometer system shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One electronic thermometer, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, inspection level II and AQL (percent defective) 1.5 for major defects and AQL 2.5 for minor defects. Unit of product for tests purposes shall be one thermometer.

Sampling for tests. Sampling for tests shall be indicated for inspection level S-1, acceptance number zero, in accordance with MIL-STD-105. Unit of product for tests purposes shall be one thermometer.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Examination. The thermometer shall be examined to determine compliance with the requirements of this specification in accordance with the defects listed below:

Defects

Major:

All components not supplied as required.
Unit not operable on the power supply as required.
Probe not color coded as required.
Probe cover fit poorly on the probe.
Probe cable not coiled and in the length specified.
Power cord not in the length specified.
Power cord plug not of the type specified.

Minor:

Instructions for use not provided on the unit.
Unit not free of scratches and dents.

Tests. Tests shall be conducted to determine compliance with specification.

Accuracy test. Set the thermometer in the Monitor mode and set the °F - °C switch to °F(Fahrenheit) or °C(Celsius) as desired. Install probe cover on the probe. Insert probe tip into a circulating water bath that has a known temperature between 31.1°C (88°F) and 42.1°C (107.9°F) and is accurate to within ± 0.03°F. After 3 to 5 minutes, the displayed temperature should completely stop changing and shall equal to known bath temperature to within ± 0.1°C (± 0.2°F).

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-187-6606	THERMOMETER, CLINICAL, HUMAN, Electronic, Battery Powered, with Battery Charger

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MQPAS

Project No. 6515-4350

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In [illegible] submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53656	THERMOMETER, CLINICAL, HUMAN (Electronic, Battery Powered)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.