A-A-53657
26 August 1988

COMMERCIAL ITEM DESCRIPTION

TEST TUBE(Disposable)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers the requirements for disposable test tubes suitable for use in clinical laboratory procedures in military medical facilities.

Salient characteristics.

Shall be disposable glass culture test tubes that are lipless, for one item use in laboratory procedures. Shall be of the following sizes:

Size 1	10 by 75 millimeters (mm)
Size 2	12 by 75 mm
Size 3	13 by 100 mm
Size 4	16 by 125 mm

Tubes shall conform with ASTM Standard E890 (Standard Specification for Disposable Glass Culture Tubes).

Shall be fabricated of borosilicate glass. Glass shall conform to the requirements of ASTM Standard E438, Type I (Specification for Glasses in Laboratory Apparatus).

The 12 by 75 mm size shall be capable of withstanding autoclaving at 250°F for at least five minutes.(See Quality Assurance Provisions)

Workmanship. The test tubes shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing one thousand test tubes, as specified, constitutes one unit.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:22Z
Check the source to verify that this is the current version before use.

Reference Standard. Where standard is referenced herein, the use of that standard in effect on the date of solicitation shall be used to determine compliance with stated requirements.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for tests shall be conducted in accordance with the applicable sections of ASTM E1157 (Standard Specification for Sampling and Testing of Reusable Laboratory Glassware).

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for two or more test characteristics.

Autoclaving test. The bottle shall be tested for autoclavability in accordance with ASTM E1157.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers are covered by this document:

Size 1	6640-01-120-3552	TEST TUBE, 10 by 75 mm
Size 2	6640-01-119-0013	TEST TUBE, 12 by 75 mm
Size 3	6640-01-049-2928	TEST TUBE, 13 by 100 mm
Size 4	6640-01-119-7690	TEST TUBE, 16 by 125 mm

ASTM Standard E890, E438 and E1157 can be obtained from American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103.

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

PGC

Project No. 6640-1503

Location: Disk No. 13
CID-TEST-TUBE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER
-A-53657

2 DOCUMENT TITLE
TEST TUBE (Disposable)

3a NAME OF SUBMITTING ORGANIZATION

b ADDRESS (Street, City, State, ZIP Code)

4 TYPE OF ORGANIZATION (Mark one)
- ☐ VENDOR
- ☐ USER
- ☐ MANUFACTURER
- ☐ OTHER (Specify)

5 PROBLEM AREAS

a Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER (Last, First, MI) – Optional

b WORK TELEPHONE NUMBER (Include Area Code) – Optional

c. MAILING ADDRESS (Street, City, State, ZIP Code) – Optional

8 DATE OF SUBMISSION (YYMMDD)

DD FORM 82 MAR 1426

PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:22Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
PHILADELPHIA, PENNSYLVANIA 19101-8419

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:22Z
Check the source to verify that this is the current version before use.